

LB125422


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58842



A= 0.0000e+000

B= 1.6688e-004

C= -1.6539e-002

Rho= 0.9991241

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.014	-0.014	17	0.000	17					-
0.2	0.200	0.170	-0.030	1120	0.0 %	1120					-15
2.5	2.500	2.547	0.047	15361	0.0 %	15361					2
5.0	5.000	4.894	-0.106	29426	0.0 %	29426					-2
7.5	7.500	7.800	0.300	46837	0.0 %	46837					4
10.0	10.000	9.803	-0.197	58842	0.0 %	58842					-2

LB125422 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 May 2023 15:55:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0000	7471B	PPB	11 May 2023 15:57:41	S	Std
0.2 - 1	50.2	1120	-	0.2000	7471B	PPB	11 May 2023 15:59:58	S	Std
2.5 - 1	52.5	15361	-	2.5000	7471B	PPB	11 May 2023 16:02:15	S	Std
5.0 - 1	55	29426	-	5.0000	7471B	PPB	11 May 2023 16:04:33	S	Std
7.5 - 1	57.5	46837	-	7.5000	7471B	PPB	11 May 2023 16:06:51	S	Std
10.0 - 1	50	58842	-	10.0000	7471B	PPB	11 May 2023 16:09:09	S	Std
ICV74 - 1	ICV74	23202	3.8554	-	7471B	PPB	11 May 2023 16:11:54	U	SMPL
ICB74 - 1	ICB74	8	-0.0152	-	7471B	PPB	11 May 2023 16:17:59	U	SMPL
CCV51 - 1	CCV51	29334	4.8787	-	7471B	PPB	11 May 2023 16:20:14	U	SMPL
CCB51 - 1	CCB51	-154	-0.0422	-	7471B	PPB	11 May 2023 16:22:29	U	SMPL
CRA - 1	CRA	1096	0.1664	-	7471B	PPB	11 May 2023 16:24:46	U	SMPL
HighStd - 1	HighStd	58563	9.7563	-	7471B	PPB	11 May 2023 16:27:01	U	SMPL
ChkStd - 1	ChkStd	41737	6.9485	-	7471B	PPB	11 May 2023 16:29:17	U	SMPL
PB152742BL - 1	PBS	-155	-0.0424	-	7471B	PPB	11 May 2023 16:31:36	U	SMPL
PB152742BS - 1	LCSS	20610	3.4228	-	7471B	PPB	11 May 2023 16:33:54	U	SMPL
O2719-01 - 1	SB-04_0-2_20230509	27050	4.4975	-	7471B	PPB	11 May 2023 16:36:11	U	SMPL
O2719-01DUP - 1	SB-04_0-2_20230509DUP	23817	3.9580	-	7471B	PPB	11 May 2023 16:38:29	U	SMPL
O2719-04 - 1	SB-04_10-12_20230509	422	0.0539	-	7471B	PPB	11 May 2023 16:45:24	U	SMPL
CCV52 - 1	CCV52	29446	4.8974	-	7471B	PPB	11 May 2023 16:47:42	U	SMPL
CCB52 - 1	CCB52	-122	-0.0369	-	7471B	PPB	11 May 2023 16:49:58	U	SMPL
O2719-05 - 1	SB-X01_10-12_20230509	876	0.1296	-	7471B	PPB	11 May 2023 16:52:15	U	SMPL
O2719-06 - 1	SB-02_0-2_20230509	4909	0.8027	-	7471B	PPB	11 May 2023 16:54:30	U	SMPL
O2719-07 - 1	SB-02_10-12_20230509	709	0.1018	-	7471B	PPB	11 May 2023 16:58:50	U	SMPL
O2719-08 - 1	SB-02_5-7_20230509	755	0.1095	-	7471B	PPB	11 May 2023 17:01:05	U	SMPL
O2719-09 - 1	SB-01_0-2_20230509	20341	3.3779	-	7471B	PPB	11 May 2023 17:03:21	U	SMPL
O2719-10 - 1	SB-01_10-12_20230509	860	0.1270	-	7471B	PPB	11 May 2023 17:05:38	U	SMPL
O2719-11 - 1	SB-01_5-7_20230509	830	0.1220	-	7471B	PPB	11 May 2023 17:07:56	U	SMPL
O2721-01 - 1	WC-1	12359	2.0459	-	7471B	PPB	11 May 2023 17:10:13	U	SMPL
O2724-01 - 1	TOPSOIL-OU3-GRILLO-1	6973	1.1471	-	7471B	PPB	11 May 2023 17:12:30	U	SMPL
O2743-01 - 1	DRUM-A-SLUDGE	533	0.0724	-	7471B	PPB	11 May 2023 17:14:46	U	SMPL
CCV53 - 1	CCV53	29927	4.9776	-	7471B	PPB	11 May 2023 17:17:02	U	SMPL
CCB53 - 1	CCB53	-80	-0.0299	-	7471B	PPB	11 May 2023 17:19:17	U	SMPL
O2743-03 - 1	DRUM-B-SLUDGE	291	0.0320	-	7471B	PPB	11 May 2023 17:21:35	U	SMPL
O2743-05 - 1	RO-COMP	7387	1.2162	-	7471B	PPB	11 May 2023 17:23:51	U	SMPL
O2745-01 - 1	NB-08-051023	1052	0.1590	-	7471B	PPB	11 May 2023 17:26:06	U	SMPL
O2719-02 - 1	O2719-01MS	48441	8.0672	-	7471B	PPB	11 May 2023 17:28:22	U	SMPL
O2719-03 - 1	O2719-01MSD	48333	8.0492	-	7471B	PPB	11 May 2023 17:30:39	U	SMPL
O2719-01LX5 - 1		4786	0.7821	-	7471B	PPB	11 May 2023 17:32:58	U	SMPL
O2719-01A - 1		52038	8.6675	-	7471B	PPB	11 May 2023 17:35:17	U	SMPL
CCV54 - 1	CCV54	29519	4.9095	-	7471B	PPB	11 May 2023 17:37:34	U	SMPL
CCB54 - 1	CCB54	-142	-0.0402	-	7471B	PPB	11 May 2023 17:39:54	U	SMPL