



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB125434

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/12/2023	09:15
2	CAL2	1	Water	NA	NA	20.6	7.01	05/12/2023	09:16
3	CAL3	1	Water	NA	NA	20.6	10.03	05/12/2023	09:17
4	ICV	1	Water	NA	NA	20.8	6.99	05/11/2023	09:20
5	CCV1	1	Water	NA	NA	20.6	2.01	05/11/2023	09:25
6	O2693-02	1	Water	NA	NA	20.5	6.00	05/12/2023	09:30
7	O2693-02DUP	1	Water	NA	NA	20.6	6.01	05/12/2023	09:31
8	CCV2	1	Water	NA	NA	20.6	12.02	05/12/2023	09:33

JB 125434

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02963 w WorkList ID : 169928 Department : Wet-Chemistry Date : 05-12-2023 09:04:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/09/2023	Water	O2693-02	Corrosivity	Cool 4 deg C	PSEG03	O11	BEL-23-001	05/09/2023	9040C

Date/Time 05.12.23 09:10
Raw Sample Received by: JDCDC1
Raw Sample Relinquished by: JDCDC1

Date/Time 05.12.23 10:20
Raw Sample Received by: JDCDC1
Raw Sample Relinquished by: JDCDC1