



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 13:50
In Date: 05/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:52
Out Date: 05/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125435

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2708-01	EYKY2	1	1.14	8.43	9.57	8.95	92.6	
O2708-02	EYKY3	2	1.18	8.79	9.97	9.18	91.0	
O2708-03	EYKY4	3	1.18	8.48	9.66	9.16	94.1	
O2708-04	EYKY6	4	1.19	8.49	9.68	8.93	91.2	
O2708-05	EYKY7	5	1.16	8.50	9.66	8.94	91.5	
O2708-06	EYKY8	6	1.15	8.81	9.96	9.14	90.7	
O2708-07	EYKY9	7	1.19	8.62	9.81	9.34	94.5	
O2708-08	EYKZ0	8	1.19	8.51	9.7	9.22	94.4	
O2708-09	EYKZ1	9	1.18	8.38	9.56	8.84	91.4	
O2708-10	EYKZ2	10	1.19	8.40	9.59	8.76	90.1	
O2708-11	EYKZ3	11	1.15	8.61	9.76	9.22	93.7	
O2708-12	EYKZ4	12	1.16	8.48	9.64	9.48	98.1	
O2708-13	EYKZ5	13	1.19	8.51	9.7	9.5	97.6	
O2708-14	EYKZ6	14	1.18	8.49	9.67	9.55	98.6	
O2708-15	EYKZ7	15	1.15	8.80	9.95	9.57	95.7	
O2708-16	EYKZ8	16	1.19	8.60	9.79	9.61	97.9	
O2708-17	EYKZ9	17	1.16	8.81	9.97	9.57	95.5	
O2708-18	EYL00	18	1.13	8.63	9.76	9.2	93.5	
O2708-19	EYL01	19	1.12	8.40	9.52	8.54	88.3	
O2708-20	EYL08	20	1.13	8.71	9.84	9.14	92.0	
O2708-21	EYL08D	21	1.13	8.71	9.84	9.14	92.0	
O2708-22	EYL08S	22	1.13	8.71	9.84	9.14	92.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$