


LB125440

7470A

INSTRUMENT ID: CV1

μ Abs.:

42438



Linear

A= 0.0000e+000

B= 2.3541e-004

C= 3.8534e-002

Rho= 0.9999103

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.056	0.056	76	0.000	76					
0.2	0.200	0.226	0.026	797	0.0 %	797	0				
2.5	2.500	2.401	-0.099	10036	0.0 %	10036					13
5.0	5.000	4.987	-0.013	21020	0.0 %	21020					-4
7.5	7.500	7.500	0.000	31697	0.0 %	31697					0
10.0	10.000	10.029	0.029	42438	0.0 %	42438					0

LB125440 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 May 2023 11:12:15

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	76	-	0.0000	7470A	PPB	12 May 2023 11:28:18	S	Std
0.2 - 1	50.2	797	-	0.2000	7470A	PPB	12 May 2023 11:30:35	S	Std
2.5 - 1	52.5	10036	-	2.5000	7470A	PPB	12 May 2023 11:32:53	S	Std
5.0 - 1	55	21020	-	5.0000	7470A	PPB	12 May 2023 11:35:10	S	Std
7.5 - 1	57.5	31697	-	7.5000	7470A	PPB	12 May 2023 11:37:26	S	Std
10.0 - 1	510	42438	-	10.0000	7470A	PPB	12 May 2023 11:39:44	S	Std
ICV75 - 1	ICV75	18273	4.3402	-	7470A	PPB	12 May 2023 11:42:29	U	SMPL
ICB75 - 1	ICB75	-116	0.0112	-	7470A	PPB	12 May 2023 11:44:44	U	SMPL
CCV55 - 1	CCV55	20566	4.8800	-	7470A	PPB	12 May 2023 11:47:01	U	SMPL
CCB55 - 1	CCB55	-117	0.0110	-	7470A	PPB	12 May 2023 11:49:16	U	SMPL
CRA - 1	CRA	741	0.2130	-	7470A	PPB	12 May 2023 11:51:33	U	SMPL
HighStd - 1	HighStd	42679	10.0857	-	7470A	PPB	12 May 2023 11:53:48	U	SMPL
ChkStd - 1	ChkStd	28827	6.8248	-	7470A	PPB	12 May 2023 11:56:04	U	SMPL
PB152759BL - 1	PBW	-211	-0.0111	-	7470A	PPB	12 May 2023 11:58:23	U	SMPL
PB152759BS - 1	LCSW	14737	3.5078	-	7470A	PPB	12 May 2023 12:00:41	U	SMPL
O2719-12 - 1	FB-01_20230509	-89	0.0176	-	7470A	PPB	12 May 2023 12:02:57	U	SMPL
O2719-13 - 1	FB-01_20230509	55	0.0515	-	7470A	PPB	12 May 2023 12:05:15	U	SMPL
O2742-01 - 1	14-051023	1762	0.4533	-	7470A	PPB	12 May 2023 12:07:32	U	SMPL
O2743-07 - 1	DRUM-A-WATER	354	0.1219	-	7470A	PPB	12 May 2023 12:09:49	U	SMPL
O2744-01 - 1	PL-OILY WATER	340	0.1186	-	7470A	PPB	12 May 2023 12:12:05	U	SMPL
CCV56 - 1	CCV56	20456	4.8541	-	7470A	PPB	12 May 2023 12:14:20	U	SMPL
CCB56 - 1	CCB56	-46	0.0277	-	7470A	PPB	12 May 2023 12:16:36	U	SMPL
O2755-01 - 1	A5560	153	0.0746	-	7470A	PPB	12 May 2023 12:18:52	U	SMPL
O2755-01DUP - 1	A5560DUP	231	0.0929	-	7470A	PPB	12 May 2023 12:21:07	U	SMPL
O2755-01MS - 1	A5560MS	14952	3.5584	-	7470A	PPB	12 May 2023 12:23:23	U	SMPL
O2755-01MSD - 1	A5560MSD	13817	3.2912	-	7470A	PPB	12 May 2023 12:25:39	U	SMPL
O2755-01LX5 - 1		-60	0.0244	-	7470A	PPB	12 May 2023 12:27:55	U	SMPL
O2755-01A - 1		17324	4.1168	-	7470A	PPB	12 May 2023 12:30:11	U	SMPL
CCV57 - 1	CCV57	22648	5.3702	-	7470A	PPB	12 May 2023 12:32:28	U	SMPL
CCB57 - 1	CCB57	-109	0.0129	-	7470A	PPB	12 May 2023 12:34:45	U	SMPL