



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125441

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/12/2023	08:40
2	O2742-01		1	12	100.00	>212.0	>212.0	05/12/2023	09:10
3	O2742-01DUP		1	14	100.00	>212.0	>212.0	05/12/2023	09:40
4	O2743-07		1	14	100.00	>212.0	>212.0	05/12/2023	10:10
5	O2744-01		1	13	100.00	>212.0	>212.0	05/12/2023	10:40
6	O2755-01		1	14	100.00	>212.0	>212.0	05/12/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB125441

WorkList Name : fp-5-11

WorkList ID : 169878

Department : Wet-Chemistry

Date : 05-11-2023 08:24:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/15/2023	Water	O2742-01	Flash Point	Cool 4 deg C	PSEG03	K31	14-051023	05/10/2023	1010B
05/17/2023	Water	O2743-07	Flash Point	Cool 4 deg C	PSEG03	O11	DRUM-A-WATER	05/10/2023	1010B
05/17/2023	Water	O2744-01	Flash Point	Cool 4 deg C	PSEG03	M11	PL-OILYWATER	05/10/2023	1010B
05/15/2023	Water	O2755-01	Flash Point	Cool 4 deg C	PSEG03	O11	A5560	05/11/2023	1010B

Date/Time 05/12/2023 08:30
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JAC

Date/Time 05/12/2023 08:50
 Raw Sample Received by: JAC
 Raw Sample Relinquished by: RM