



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 14:50
In Date: 05/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 05/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125443

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2713-01	EYL02	1	1.14	8.56	9.7	8.41	84.9	
O2713-02	EYL03	2	1.19	8.42	9.61	8.00	80.9	
O2713-03	EYL04	3	1.15	8.42	9.57	8.04	81.8	
O2713-04	EYL05	4	1.19	8.41	9.6	8.05	81.6	
O2713-05	EYL06	5	1.19	8.62	9.81	8.22	81.6	
O2713-06	EYL07	6	1.17	8.50	9.67	8.19	82.6	
O2713-07	EYL09	7	1.15	8.35	9.5	8.00	82.0	
O2713-08	EYL10	8	1.19	8.51	9.7	8.27	83.2	
O2713-09	EYL11	9	1.15	8.40	9.55	8.26	84.6	
O2713-10	EYL12	10	1.16	8.37	9.53	8.04	82.2	
O2713-11	EYL93	11	1.18	8.50	9.68	8.13	81.8	
O2713-12	EYL94	12	1.15	8.81	9.96	8.45	82.9	
O2713-13	EYL95	13	1.1	8.51	9.61	8.47	86.6	
O2713-14	EYL96	14	1.19	8.58	9.77	8.38	83.8	
O2713-15	EYL97	15	1.18	8.72	9.9	8.47	83.6	
O2713-16	EYL99	16	1.19	8.79	9.98	8.45	82.6	
O2713-17	EYLA0	17	1.15	8.53	9.68	8.53	86.5	
O2713-18	EYLA1	18	1.14	8.60	9.74	8.1	80.9	
O2713-19	EYLA2	19	1.19	8.58	9.77	8.04	79.8	
O2713-20	EYLA2D	20	1.19	8.58	9.77	8.04	79.8	
O2713-21	EYLA2S	21	1.19	8.58	9.77	8.04	79.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$