


LB125456

7471B

INSTRUMENT 1): CV1

 μ Abs.:

56999



Linear

A= 0.0000e+000

B= 1.7296e-004

C= 1.8220e-002

Rho= 0.9996264

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.024	0.024	31	0.000	31					9.11
0.2	0.200	0.206	0.006	1083	0.0 %	1083					1
2.5	2.500	2.414	-0.086	13853	0.0 %	13853					3
5.0	5.000	4.986	-0.014	28721	0.0 %	28721					-3
7.5	7.500	7.694	0.194	44380	0.0 %	44380					30
10.0	10.000	9.877	-0.123	56999	0.0 %	56999					1

LB125456 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 May 2023 15:10:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7471B	PPB	12 May 2023 15:56:24	S	Std
0.2 - 1	512	1083	-	0.2000	7471B	PPB	12 May 2023 15:58:41	S	Std
2.5 - 1	52.5	13853	-	2.5000	7471B	PPB	12 May 2023 16:00:59	S	Std
5.0 - 1	55	28721	-	5.0000	7471B	PPB	12 May 2023 16:03:16	S	Std
7.5 - 1	57.5	44380	-	7.5000	7471B	PPB	12 May 2023 16:05:34	S	Std
10.0 - 1	511	56999	-	10.0000	7471B	PPB	12 May 2023 16:10:26	S	Std
ICV76 - 1	ICV76	22553	3.9190	-7471B	PPB	12 May 2023 16:15:41	U	SMPL	
ICB76 - 1	ICB76	-125	-0.0034	-7471B	PPB	12 May 2023 16:17:56	U	SMPL	
CCV58 - 1	CCV58	28828	5.0043	-7471B	PPB	12 May 2023 16:20:13	U	SMPL	
CCB58 - 1	CCB58	-139	-0.0058	-7471B	PPB	12 May 2023 16:22:29	U	SMPL	
CRA - 1	CRA	1060	0.2016	-7471B	PPB	12 May 2023 16:24:45	U	SMPL	
HighStd - 1	HighStd	57049	9.8854	-7471B	PPB	12 May 2023 16:27:00	U	SMPL	
ChkStd - 1	ChkStd	41001	7.1097	-7471B	PPB	12 May 2023 16:29:16	U	SMPL	
PB152771BL - 1	PBS	-240	-0.0233	-7471B	PPB	12 May 2023 16:31:35	U	SMPL	
PB152771BS - 1	LCSS	19742	3.4328	-7471B	PPB	12 May 2023 16:33:54	U	SMPL	
O2747-01 - 1	SB-03-0-2-20230510	21486	3.7344	-7471B	PPB	12 May 2023 16:36:11	U	SMPL	
O2747-02 - 1	SB-03-10-12-20230510	728	0.1441	-7471B	PPB	12 May 2023 16:38:29	U	SMPL	
O2747-03 - 1	SB-03-5-7-20230510	1067	0.2028	-7471B	PPB	12 May 2023 16:40:48	U	SMPL	
O2750-01 - 1	PAT-1-051123	8950	1.5662	-7471B	PPB	12 May 2023 16:43:05	U	SMPL	
O2750-03 - 1	PAT-2-051123	452	0.0964	-7471B	PPB	12 May 2023 16:45:24	U	SMPL	
CCV59 - 1	CCV59	28475	4.9432	-7471B	PPB	12 May 2023 16:47:39	U	SMPL	
CCB59 - 1	CCB59	-112	-0.0012	-7471B	PPB	12 May 2023 16:49:55	U	SMPL	
O2753-01 - 1	OE-3-051123	3956	0.7024	-7471B	PPB	12 May 2023 16:52:13	U	SMPL	
O2757-01 - 1	WC-1	17119	2.9791	-7471B	PPB	12 May 2023 16:54:29	U	SMPL	
O2758-01 - 1	CONC-WC-1	948	0.1822	-7471B	PPB	12 May 2023 16:56:45	U	SMPL	
O2760-01 - 1	HR-2-051123	3247	0.5798	-7471B	PPB	12 May 2023 16:59:01	U	SMPL	
O2760-03 - 1	HR-4-051123	2181	0.3954	-7471B	PPB	12 May 2023 17:07:28	U	SMPL	
O2760-03DUP - 1	HR-4-051123DUP	2148	0.3897	-7471B	PPB	12 May 2023 17:09:44	U	SMPL	
O2760-03MS - 1	HR-4-051123MS	22521	3.9134	-7471B	PPB	12 May 2023 17:12:00	U	SMPL	
O2760-03MSD - 1	HR-4-051123MSD	24830	4.3128	-7471B	PPB	12 May 2023 17:14:17	U	SMPL	
O2760-03LX5 - 1		241	0.0599	-7471B	PPB	12 May 2023 17:16:36	U	SMPL	
O2760-03A - 1		26641	4.6260	-7471B	PPB	12 May 2023 17:18:53	U	SMPL	
CCV60 - 1	CCV60	28952	5.0257	-7471B	PPB	12 May 2023 17:21:09	U	SMPL	
CCB60 - 1	CCB60	-131	-0.0044	-7471B	PPB	12 May 2023 17:23:25	U	SMPL	