



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125460

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	05/12/2023	15:30
2	CAL2	1	Water	NA	NA	20.9	7.00	05/12/2023	15:31
3	CAL3	1	Water	NA	NA	20.7	10.03	05/12/2023	15:33
4	ICV	1	Water	NA	NA	20.8	6.99	05/12/2023	15:35
5	CCV1	1	Water	NA	NA	20.7	2.01	05/12/2023	15:40
6	O2781-01	1	Water	NA	NA	26.2	6.96	05/12/2023	15:45
7	O2781-01DUP	1	Water	NA	NA	26.3	6.97	05/12/2023	15:46
8	CCV2	1	Water	NA	NA	20.8	12.03	05/12/2023	15:50

VB 125460

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2781

WorkList ID : 169966

Department : Wet-Chemistry

Date : 05-12-2023 15:22:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2023	Water	O2781-01	pH		Cool 4 deg C	PSEG03 O11	MGP-WASTE-WASTER-5/12/23	05/12/2023	9040C

Date/Time 05.12.23 15.25

Raw Sample Received by: JD (WOC)

Raw Sample Relinquished by: JD (CSM)

Date/Time

05.12.23

Raw Sample Received by: JD (CSM)

Raw Sample Relinquished by: JD (WOC)