



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125462

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	05/12/2023	15:30
2	CAL2	1	Water	NA	NA	20.9	7.00	05/12/2023	15:31
3	CAL3	1	Water	NA	NA	20.7	10.03	05/12/2023	15:33
4	ICV	1	Water	NA	NA	20.8	6.99	05/12/2023	15:35
5	CCV1	1	Water	NA	NA	20.7	2.01	05/12/2023	15:40
6	O2784-01	1	Solid	20.02	20	25.5	7.83	05/12/2023	16:25
7	O2784-01DUP	1	Solid	20.03	20	25.6	7.84	05/12/2023	16:26
8	CCV2	1	Water	NA	NA	20.7	12.02	05/12/2023	16:30

WB 125462

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02784

WorkList ID : 169967

Department : Wet-Chemistry

Date : 05-12-2023 16:06:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2023	Solid	O2784-01	pH	Cool 4 deg C	PSEG05	O11	BU-03-051223	05/12/2023	9045D

05.12.23

16:15

05.12.23

10 (WC1)

05.12.23

17:00

05.12.23

Raw Sample Received by:

RM CWS

Raw Sample Received by:

RM CWS

Raw Sample Relinquished by:

Raw Sample Relinquished by: