



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 05/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125479

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2728-01	EYLP5	1	1.14	8.58	9.72	7.73	76.8	
O2728-02	EYLP6	2	1.13	8.43	9.56	7.92	80.5	
O2728-03	EYLP7	3	1.19	8.53	9.72	7.93	79.0	
O2728-04	EYLP8	4	1.18	8.74	9.92	8.08	78.9	
O2728-05	EYLP9	5	1.15	8.39	9.54	7.52	75.9	
O2728-06	EYLP0	6	1.13	8.81	9.94	8.2	80.2	
O2728-07	EYLP1	7	1.15	8.50	9.65	7.86	78.9	
O2728-08	EYLP2	8	1.15	8.83	9.98	8.17	79.5	
O2728-09	EYLP4	9	1.1	8.76	9.86	8.12	80.1	
O2728-10	EYLP5	10	1.12	8.80	9.92	7.48	72.3	
O2728-11	EYLP6	11	1.18	8.78	9.96	7.43	71.2	
O2728-12	EYLP7	12	1.15	8.82	9.97	8.14	79.3	
O2728-13	EYLP8	13	1.15	8.82	9.97	7.53	72.3	
O2728-14	EYLP9	14	1.18	8.64	9.82	7.47	72.8	
O2728-15	EYLR0	15	1.16	8.81	9.97	7.81	75.5	
O2728-16	EYLR1	16	1.19	8.70	9.89	8.1	79.4	
O2728-17	EYLR2	17	1.12	8.59	9.71	7.25	71.4	
O2728-18	EYLR3	18	1.18	8.42	9.6	7.54	75.5	
O2728-19	EYLR4	19	1.15	8.62	9.77	7.58	74.6	
O2728-20	EYLR5	20	1.19	8.57	9.76	8.08	80.4	
O2728-21	EYLR5D	21	1.19	8.57	9.76	8.08	80.4	
O2728-22	EYLR5S	22	1.19	8.57	9.76	8.08	80.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$