



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:40
In Date: 05/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 05/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125480

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2729-01	EYLR6	1	1.15	8.83	9.98	7.93	76.8	
O2729-02	EYLR7	2	1.12	8.45	9.57	7.47	75.1	
O2729-03	EYLR8	3	1.19	8.75	9.94	7.77	75.2	
O2729-04	EYLR9	4	1.19	8.78	9.97	7.79	75.2	
O2729-05	EYLS0	5	1.13	8.75	9.88	7.8	76.2	
O2729-06	EYLS1	6	1.16	8.43	9.59	7.93	80.3	
O2729-07	EYLS2	7	1.15	8.81	9.96	8.43	82.6	
O2729-08	EYLS3	8	1.17	8.44	9.61	7.57	75.8	
O2729-09	EYLS4	9	1.18	8.79	9.97	7.88	76.2	
O2729-10	EYLS5	10	1.19	8.55	9.74	7.64	75.4	
O2729-11	EYLS6	11	1.19	8.67	9.86	7.99	78.4	
O2729-12	EYLS7	12	1.18	8.59	9.77	7.63	75.1	
O2729-13	EYLS7D	13	1.18	8.59	9.77	7.63	75.1	
O2729-14	EYLS7S	14	1.18	8.59	9.77	7.63	75.1	
O2729-15	EYLS8	15	1.13	8.58	9.71	8.00	80.1	
O2729-16	EYLS9	16	1.12	8.40	9.52	7.63	77.5	
O2729-17	EYLT0	17	1.18	8.74	9.92	8.39	82.5	
O2729-18	EYLT1	18	1.18	8.43	9.61	8.08	81.9	
O2729-19	EYLT2	19	1.19	8.49	9.68	8.16	82.1	
O2729-20	EYLT3	20	1.19	8.69	9.88	8.09	79.4	
O2729-21	EYLT4	21	1.18	8.48	9.66	8.29	83.8	
O2729-22	EYLT5	22	1.19	8.80	9.99	8.28	80.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$