



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:15
In Date: 05/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:50
Out Date: 05/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125481

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2730-01	EYLT6	1	1.18	8.56	9.74	8.06	80.4	
O2730-02	EYLT7	2	1.19	8.69	9.88	8.01	78.5	
O2730-03	EYLT8	3	1.19	8.51	9.7	7.94	79.3	
O2730-04	EYLT9	4	1.19	8.39	9.58	7.02	69.5	
O2730-05	EYLOW	5	1.19	8.79	9.98	7.9	76.3	
O2730-06	EYLOW1	6	1.16	8.35	9.51	7.38	74.5	
O2730-07	EYLOW2	7	1.19	8.53	9.72	7.78	77.3	
O2730-08	EYLOW3	8	1.19	8.63	9.82	8.55	85.3	
O2730-09	EYLOW4	9	1.18	8.47	9.65	7.79	78.0	
O2730-10	EYLOW5	10	1.12	8.66	9.78	7.99	79.3	
O2730-11	EYLOW6	11	1.15	8.45	9.6	7.84	79.2	
O2730-12	EYLOW7	12	1.17	8.60	9.77	7.6	74.8	
O2730-13	EYLOW8	13	1.14	8.82	9.96	7.77	75.2	
O2730-14	EYLOW9	14	1.19	8.45	9.64	7.46	74.2	
O2730-15	EYLOW	15	1.18	8.53	9.71	7.55	74.7	
O2730-16	EYLOW0	16	1.18	8.53	9.71	7.55	74.7	
O2730-17	EYLOW0S	17	1.18	8.53	9.71	7.55	74.7	
O2730-18	EYLOW1	18	1.19	8.59	9.78	8.23	82.0	
O2730-19	EYLOW2	19	1.18	8.74	9.92	8.26	81.0	
O2730-20	EYLOW3	20	1.15	8.81	9.96	8.34	81.6	
O2730-21	EYLOW4	21	1.17	8.49	9.66	7.97	80.1	
O2730-22	EYLOW5	22	1.18	8.52	9.7	7.76	77.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$