



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125482

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/15/2023	08:15
2	CAL2	1	Water	NA	NA	20.7	7.00	05/15/2023	08:16
3	CAL3	1	Water	NA	NA	20.7	10.03	05/15/2023	08:20
4	ICV	1	Water	NA	NA	20.7	6.99	05/15/2023	08:22
5	CCV1	1	Water	NA	NA	20.7	2.01	05/15/2023	08:25
6	O2774-04	1	Solid	20.02	20	24.5	8.14	05/15/2023	08:30
7	O2774-04DUP	1	Solid	20.03	20	24.6	8.15	05/15/2023	08:31
8	O2782-01	1	Solid	20.04	20	25.2	6.35	05/15/2023	08:35
9	CCV2	1	Water	NA	NA	20.7	12.02	05/15/2023	08:40

VB125482

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY 02782 WorkList ID : 169988 Department : Wet-Chemistry Date : 05-15-2023 08:02:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/11/2023	Solid	O2774-04	Corrosivity	Cool 4 deg C	PORT06	O11	WC-N-1	05/11/2023	9045D
05/12/2023	Solid	O2782-01	Corrosivity	Cool 4 deg C	PSEG03	O11	D3034	05/12/2023	9045D

Date/Time 05-15-23 08:10
 Raw Sample Received by: JDL (w/c)
 Raw Sample Relinquished by: JDL SM

Date/Time 05-15-23 12:00
 Raw Sample Received by: JDL SM
 Raw Sample Relinquished by: JDL (w/c)