

LB125493


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4476E



A= 0.0000e+000

B= 2.1926e-004

C= 1.1226e-002

Rho= 0.9989569

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o)
0.0	0.000	0.016	0.016	21	0.000	21	0				1
0.2	0.200	0.206	0.006	890	0.0 %	890					3
2.5	2.500	2.505	0.005	11373	0.0 %	11373					0
5.0	5.000	4.819	-0.181	21927	0.0 %	21927					-4
7.5	7.500	7.827	0.327	35648	0.0 %	35648					4
10.0	10.000	9.827	-0.173	44766	0.0 %	44766					-2

LB125493 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2023 13:53:51

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.000	7470A	PPB	15 May 2023 13:54:03	S	Std
0.2 - 1	50.2	890	-	0.200	7470A	PPB	15 May 2023 13:56:21	S	Std
2.5 - 1	52.5	11373	-	2.500	7470A	PPB	15 May 2023 13:58:39	S	Std
5.0 - 1	55	21927	-	5.000	7470A	PPB	15 May 2023 14:00:56	S	Std
7.5 - 1	57.5	35648	-	7.500	7470A	PPB	15 May 2023 14:06:01	S	Std
10.0 - 1	51.8	44766	-	10.000	7470A	PPB	15 May 2023 14:10:56	S	Std
ICV77 - 1	ICV77	18429	4.0520	-	7470A	PPB	15 May 2023 14:13:49	U	SMPL
ICB77 - 1	ICB77	-163	-0.0245	-	7470A	PPB	15 May 2023 14:16:03	U	SMPL
CCV61 - 1	CCV61	22940	5.0410	-	7470A	PPB	15 May 2023 14:18:20	U	SMPL
CCB61 - 1	CCB61	-164	-0.0247	-	7470A	PPB	15 May 2023 14:20:36	U	SMPL
CRA - 1	CRA	853	0.1983	-	7470A	PPB	15 May 2023 14:22:53	U	SMPL
HighStd - 1	HighStd	44073	9.6747	-	7470A	PPB	15 May 2023 14:25:08	U	SMPL
ChkStd - 1	ChkStd	33034	7.2543	-	7470A	PPB	15 May 2023 14:27:23	U	SMPL
PB152789BL - 1	PBW	394	0.0976	-	7470A	PPB	15 May 2023 14:29:41	U	SMPL
PB152789BS - 1	LCSW	16337	3.5933	-	7470A	PPB	15 May 2023 14:31:59	U	SMPL
O2693-02 - 1	BEL-23-001	-125	-0.0162	-	7470A	PPB	15 May 2023 14:34:15	U	SMPL
O2693-02DUP - 1	BEL-23-001DUP	149	0.0439	-	7470A	PPB	15 May 2023 14:39:31	U	SMPL
O2693-02MS - 1	BEL-23-001MS	15761	3.4670	-	7470A	PPB	15 May 2023 14:41:48	U	SMPL
O2693-02MSD - 1	BEL-23-001MSD	15574	3.4280	-	7470A	PPB	15 May 2023 14:44:05	U	SMPL
O2775-01 - 1	360	-104	-0.0116	-	7470A	PPB	15 May 2023 14:46:21	U	SMPL
CCV62 - 1	CCV62	23093	5.0746	-	7470A	PPB	15 May 2023 14:48:37	U	SMPL
CCB62 - 1	CCB62	-173	-0.0267	-	7470A	PPB	15 May 2023 14:50:53	U	SMPL
O2775-02 - 1	361	2	0.0117	-	7470A	PPB	15 May 2023 14:53:10	U	SMPL
O2782-04 - 1	D3032	96	0.0323	-	7470A	PPB	15 May 2023 14:55:26	U	SMPL
O2787-02 - 1	2791	73	0.0272	-	7470A	PPB	15 May 2023 14:57:42	U	SMPL
O2787-03 - 1	2834	97	0.0325	-	7470A	PPB	15 May 2023 14:59:58	U	SMPL
PB152790BL - 1	PBW	20	0.0156	-	7470A	PPB	15 May 2023 15:02:14	U	SMPL
PB152790BS - 1	LCSW	15291	3.3639	-	7470A	PPB	15 May 2023 15:04:30	U	SMPL
O2758-02 - 1	CONC-WC-1	-84	-0.0072	-	7470A	PPB	15 May 2023 15:06:47	U	SMPL
O2774-04 - 1	WC-N-1	11	0.0136	-	7470A	PPB	15 May 2023 15:09:05	U	SMPL
O2774-04DUP - 1	WC-N-1DUP	5	0.0123	-	7470A	PPB	15 May 2023 15:11:22	U	SMPL
O2774-04MS - 1	WC-N-1MS	15682	3.4497	-	7470A	PPB	15 May 2023 15:13:38	U	SMPL
CCV63 - 1	CCV63	22916	5.0358	-	7470A	PPB	15 May 2023 15:15:54	U	SMPL
CCB63 - 1	CCB63	-157	-0.0232	-	7470A	PPB	15 May 2023 15:18:10	U	SMPL
O2774-04MSD - 1	WC-N-1MSD	15618	3.4356	-	7470A	PPB	15 May 2023 15:20:28	U	SMPL
O2782-01 - 1	D3034	-72	-0.0046	-	7470A	PPB	15 May 2023 15:22:43	U	SMPL
O2787-01 - 1	2833	23	0.0163	-	7470A	PPB	15 May 2023 15:25:00	U	SMPL
PB152791BL - 1	PBW	43	0.0207	-	7470A	PPB	15 May 2023 15:27:15	U	SMPL
PB152791BS - 1	LCSW	14895	3.2771	-	7470A	PPB	15 May 2023 15:29:32	U	SMPL
O2724-02 - 1	TOPSOIL-OU3-GRILLO-1	669	0.1579	-	7470A	PPB	15 May 2023 15:31:48	U	SMPL
O2724-02DUP - 1	TOPSOIL-OU3-GRILLO-1DUP	557	0.1334	-	7470A	PPB	15 May 2023 15:34:05	U	SMPL
O2724-02MS - 1	TOPSOIL-OU3-GRILLO-1MS	15918	3.5014	-	7470A	PPB	15 May 2023 15:36:23	U	SMPL
O2724-02MSD - 1	TOPSOIL-OU3-GRILLO-1MSD	16183	3.5595	-	7470A	PPB	15 May 2023 15:38:40	U	SMPL
PB152792BL - 1	PBW	-80	-0.0063	-	7470A	PPB	15 May 2023 15:40:56	U	SMPL
CCV64 - 1	CCV64	22116	4.8604	-	7470A	PPB	15 May 2023 15:43:12	U	SMPL
CCB64 - 1	CCB64	-129	-0.0171	-	7470A	PPB	15 May 2023 15:45:28	U	SMPL
PB152792BS - 1	LCSW	18269	4.0189	-	7470A	PPB	15 May 2023 15:50:27	U	SMPL
O2781-01 - 1	MGP-WASTE-WASTER-5/12/23	-67	-0.0035	-	7470A	PPB	15 May 2023 15:52:43	U	SMPL
O2781-01DUP - 1	MGP-WASTE-WASTER-5/12/23DUP	135	0.0408	-	7470A	PPB	15 May 2023 15:54:59	U	SMPL
O2781-01MS - 1	MGP-WASTE-WASTER-5/12/23MS	15294	3.3646	-	7470A	PPB	15 May 2023 15:57:14	U	SMPL
O2781-01MSD - 1	MGP-WASTE-WASTER-5/12/23MSD	15472	3.4036	-	7470A	PPB	15 May 2023 15:59:30	U	SMPL
PB152756TB - 1	PB152756TB	-117	-0.0144	-	7470A	PPB	15 May 2023 16:01:48	U	SMPL
PB152770TB - 1	PB152770TB	3	0.0119	-	7470A	PPB	15 May 2023 16:04:05	U	SMPL
PB152791TB - 1	PB152782TB	-3	0.0106	-	7470A	PPB	15 May 2023 16:06:22	U	SMPL
CCV65 - 1	CCV65	20741	4.5589	-	7470A	PPB	15 May 2023 16:08:40	U	SMPL
CCB65 - 1	CCB65	-145	-0.0206	-	7470A	PPB	15 May 2023 16:10:56	U	SMPL
O2693-02LX5 - 1		14	0.0143	-	7470A	PPB	15 May 2023 16:13:13	U	SMPL
O2693-02A - 1		17051	3.7498	-	7470A	PPB	15 May 2023 16:15:28	U	SMPL
O2774-04LX5 - 1		-77	-0.0057	-	7470A	PPB	15 May 2023 16:17:44	U	SMPL
O2774-04A - 1		16669	3.6661	-	7470A	PPB	15 May 2023 16:20:00	U	SMPL
O2724-02LX5 - 1		-71	-0.0043	-	7470A	PPB	15 May 2023 16:22:16	U	SMPL
O2724-02A - 1		18364	4.0377	-	7470A	PPB	15 May 2023 16:24:32	U	SMPL
O2781-01LX5 - 1		-56	-0.0011	-	7470A	PPB	15 May 2023 16:26:48	U	SMPL
O2781-01A - 1		19173	4.2151	-	7470A	PPB	15 May 2023 16:29:06	U	SMPL
CCV66 - 1	CCV66	21640	4.7560	-	7470A	PPB	15 May 2023 16:31:23	U	SMPL
CCB66 - 1	CCB66	-190	-0.0304	-	7470A	PPB	15 May 2023 16:33:41	U	SMPL