



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125498

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASFPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	05/15/2023	11:40
2	O2787-01		1	2	8.00	46.4	46.4	05/15/2023	12:10
3	O2787-01DUP		1	2	8.00	46.4	46.4	05/15/2023	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB125498

WorkList Name : f p s 5-15

WorkList ID : 170011

Department : Wet-Chemistry

Date : 05-15-2023 11:29:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2023	Solid	O2787-01	Flash Point	Cool 4 deg C	PSEG03	M11	2833	05/12/2023	1010B

Date/Time 05/15/2023 11:35
 Raw Sample Received by: R121 CWL
 Raw Sample Relinquished by: HCD

Date/Time 05/15/2023 13:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC