



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125499

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	05/15/2023	13:10
2	O2775-01		1	12	100.00	>212.0	>212.0	05/15/2023	13:40
3	O2775-02		1	14	100.00	>212.0	>212.0	05/15/2023	14:10
4	O2781-01		1	12	100.00	>212.0	>212.0	05/15/2023	14:40
5	O2787-02		1	13	100.00	>212.0	>212.0	05/15/2023	15:10
6	O2787-03		1	14	100.00	>212.0	>212.0	05/15/2023	15:40
7	O2787-03DUP		1	14	100.00	>212.0	>212.0	05/15/2023	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-5-15

WorkList ID : 170010

Department : Wet-Chemistry

Date : 05-15-2023 11:28:16

LB125499

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2023	Water	O2775-01	Flash Point	Cool 4 deg C	PSEG03	M11	360	05/12/2023	1010B
05/19/2023	Water	O2775-02	Flash Point	Cool 4 deg C	PSEG03	M11	361	05/12/2023	1010B
05/15/2023	Water	O2781-01	Flash Point	Cool 4 deg C	PSEG03	O11	MGP-WASTE-WASTER-5/12/23	05/12/2023	1010B
05/19/2023	Water	O2787-02	Flash Point	Cool 4 deg C	PSEG03	M11	2791	05/12/2023	1010B
05/19/2023	Water	O2787-03	Flash Point	Cool 4 deg C	PSEG03	M11	2834	05/12/2023	1010B

Date/Time 05/15/2023 13:05
 Raw Sample Received by: Rm cwe
 Raw Sample Relinquished by: JH(coc)

Date/Time 05/15/2023 16:45
 Raw Sample Received by: JH(coc)
 Raw Sample Relinquished by: Rm cwe