



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125502

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	05/16/2023	10:10
2	CAL2	1	Water	NA	NA	20.5	7.00	05/16/2023	10:11
3	CAL3	1	Water	NA	NA	20.6	10.03	05/16/2023	10:15
4	ICV	1	Water	NA	NA	20.7	6.98	05/16/2023	10:20
5	CCV1	1	Water	NA	NA	20.8	2.01	05/16/2023	10:25
6	O2794-02	1	Water	NA	NA	22.6	8.80	05/16/2023	10:30
7	O2794-02DUP	1	Water	NA	NA	22.8	8.81	05/16/2023	10:31
8	O2794-03	1	Water	NA	NA	23.3	8.18	05/16/2023	10:37
9	CCV2	1	Water	NA	NA	20.6	12.02	05/16/2023	10:40

WORKLIST(Hardcopy Internal Chain)

VB 12 5302

WorkList Name : PH O2794

WorkList ID : 170038

Department : Wet-Chemistry

Date : 05-16-2023 09:56:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/15/2023	Water	O2794-02	pH	A	Cool 4 deg C	PSEG03 O11	AUD-1997	05/15/2023	9040C
05/15/2023	Water	O2794-03	pH	A	Cool 4 deg C	PSEG03 O11	A287	05/15/2023	9040C

Date/Time 05-16-23 10:00
 Raw Sample Received by: JPLCC
 Raw Sample Relinquished by: JPLCC

Date/Time 05-16-23 11:30
 Raw Sample Received by: JPLCC
 Raw Sample Relinquished by: JPLCC