

LB125503


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

60436



A= 0.0000e+000

B= 1.6461e-004

C= -4.5903e-002

Rho= 0.9994654

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	-0.014	-0.014	192	0.000	192					
0.2	0.200	0.214	0.014	1581	0.0 %	1581					
2.5	2.500	2.554	0.054	15796	0.0 %	15796					
5.0	5.000	4.830	-0.170	29620	0.0 %	29620					
7.5	7.500	7.713	0.213	47137	0.0 %	47137					
10.0	10.000	9.902	-0.098	60436	0.0 %	60436					

1
3
2
7
1

LB125503 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 May 2023 10:37:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	192	-	0.0000	7471B	PPB	16 May 2023 10:47:15	S	Std
0.2 - 1	50.2	1581	-	0.2000	7471B	PPB	16 May 2023 10:49:33	S	Std
2.5 - 1	52.1	15796	-	2.5000	7471B	PPB	16 May 2023 10:51:50	S	Std
5.0 - 1	55	29620	-	5.0000	7471B	PPB	16 May 2023 10:54:07	S	Std
7.5 - 1	57.5	47137	-	7.5000	7471B	PPB	16 May 2023 10:56:26	S	Std
10.0 - 1	50	60436	-	10.0000	7471B	PPB	16 May 2023 10:58:44	S	Std
ICV78 - 1	ICV78	23295	3.7887	-7471B	PPB	16 May 2023 11:01:34	U	SMPL	
ICB78 - 1	ICB78	-143	-0.0694	-7471B	PPB	16 May 2023 11:03:49	U	SMPL	
CCV67 - 1	CCV67	29772	4.8549	-7471B	PPB	16 May 2023 11:06:06	U	SMPL	
CCB67 - 1	CCB67	-162	-0.0726	-7471B	PPB	16 May 2023 11:08:22	U	SMPL	
CRA - 1	CRA	1571	0.2127	-7471B	PPB	16 May 2023 11:10:39	U	SMPL	
HighStd - 1	HighStd	60145	9.8546	-7471B	PPB	16 May 2023 11:12:54	U	SMPL	
ChkStd - 1	ChkStd	42016	6.8704	-7471B	PPB	16 May 2023 11:15:09	U	SMPL	
PB152798BL - 1	PBS	388	0.0180	-7471B	PPB	16 May 2023 11:17:28	U	SMPL	
PB152798BS - 1	LCSS	22881	3.7205	-7471B	PPB	16 May 2023 11:22:43	U	SMPL	
O2770-01 - 1	CHRT-25990	1121	0.1386	-7471B	PPB	16 May 2023 11:24:59	U	SMPL	
O2774-03 - 1	WC-N-1	23974	3.9005	-7471B	PPB	16 May 2023 11:27:18	U	SMPL	
O2782-02 - 1	ETGI-363	6639	1.0469	-7471B	PPB	16 May 2023 11:29:35	U	SMPL	
O2784-01 - 1	BU-03-051223	5630	0.8809	-7471B	PPB	16 May 2023 11:31:54	U	SMPL	
O2784-01DUP - 1	BU-03-051223DUP	5187	0.8079	-7471B	PPB	16 May 2023 11:34:10	U	SMPL	
CCV68 - 1	CCV68	29413	4.7958	-7471B	PPB	16 May 2023 11:36:26	U	SMPL	
CCB68 - 1	CCB68	-106	-0.0634	-7471B	PPB	16 May 2023 11:38:42	U	SMPL	
O2784-01MS - 1	BU-03-051223MS	28646	4.6695	-7471B	PPB	16 May 2023 11:40:59	U	SMPL	
O2784-01MSD - 1	BU-03-051223MSD	29076	4.7403	-7471B	PPB	16 May 2023 11:43:15	U	SMPL	
O2784-01LX5 - 1		901	0.1024	-7471B	PPB	16 May 2023 11:45:33	U	SMPL	
O2784-01A - 1		28677	4.6746	-7471B	PPB	16 May 2023 11:47:50	U	SMPL	
CCV69 - 1	CCV69	29387	4.7915	-7471B	PPB	16 May 2023 11:50:06	U	SMPL	
CCB69 - 1	CCB69	57	-0.0365	-7471B	PPB	16 May 2023 11:52:25	U	SMPL	