


LB125506

245.1

INSTRUMENT ID: CV1

 μ Abs.:

44735



Linear

A= 0.0000e+000

B= 2.2651e-004

C= -7.7942e-003

Rho= 0.9997637

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/0/1
0.0	0.000	0.014	0.014	98	0.000	98					
0.05	0.050	0.044	-0.006	227	0.0 %	227					-12
0.2	0.200	0.233	0.033	1061	0.0 %	1061	0				17
2.5	2.500	2.525	0.025	11180	0.0 %	11180					1
5.0	5.000	4.975	-0.025	21998	0.0 %	21998					-1
7.5	7.500	7.335	-0.165	32415	0.0 %	32415					-2
10.0	10.000	10.125	0.125	44735	0.0 %	44735					1

LB125506 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 16 May 2023 14:13:06

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	98	-	0.0000	245.1	PPB	16 May 2023 15:47:32	S	Std
0.05 - 1	50.05	227	-	0.0500	245.1	PPB	16 May 2023 15:52:38	S	Std
0.2 - 1	50.2	1061	-	0.2000	245.1	PPB	16 May 2023 15:54:53	S	Std
2.5 - 1	52.5	11180	-	2.5000	245.1	PPB	16 May 2023 15:57:11	S	Std
5.0 - 1	55	21998	-	5.0000	245.1	PPB	16 May 2023 15:59:28	S	Std
7.5 - 1	59.5	32415	-	7.5000	245.1	PPB	16 May 2023 16:01:44	S	Std
10.0 - 1	511	44735	-	10.0000	245.1	PPB	16 May 2023 16:06:46	S	Std
ICV79 - 1	ICV79	17330	3.9177	-245.1		PPB	16 May 2023 16:09:56	U	SMPL
ICB79 - 1	ICB79	-93	-0.0289	-245.1		PPB	16 May 2023 16:12:11	U	SMPL
CCV70 - 1	CCV70	21620	4.8894	-245.1		PPB	16 May 2023 16:14:28	U	SMPL
CCB70 - 1	CCB70	-153	-0.0425	-245.1		PPB	16 May 2023 16:16:44	U	SMPL
CRA - 1	CRA	279	0.0554	-245.1		PPB	16 May 2023 16:21:43	U	SMPL
HighStd - 1	HighStd	44750	10.1286	-245.1		PPB	16 May 2023 16:23:59	U	SMPL
ChkStd - 1	ChkStd	31409	7.1067	-245.1		PPB	16 May 2023 16:26:14	U	SMPL
PB152840BL - 1	PBW	-325	-0.0814	-245.1		PPB	16 May 2023 16:28:34	U	SMPL
PB152840BS - 1	LCSW	16808	3.7994	-245.1		PPB	16 May 2023 16:30:52	U	SMPL
O2213-10 - 1	LOD-MDL-WATER-04-QT2-2023	209	0.0395	-245.1		PPB	16 May 2023 16:36:03	U	SMPL
O2213-10RE - 1	LOD-MDL-WATER-04-QT2-2023RE	726	0.1567	-245.1		PPB	16 May 2023 16:38:19	U	SMPL
O2213-12 - 1	MDL-WATER-06-QT2-2023	287	0.0572	-245.1		PPB	16 May 2023 16:43:16	U	SMPL
CCV71 - 1	CCV71	21426	4.8455	-245.1		PPB	16 May 2023 16:45:33	U	SMPL
CCB71 - 1	CCB71	-184	-0.0495	-245.1		PPB	16 May 2023 16:47:48	U	SMPL