



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125515

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	05/17/2023	08:45
2	O2782-04		1	14	100.00	>212.0	>212.0	05/17/2023	09:15
3	O2782-04DUP		1	13	100.00	>212.0	>212.0	05/17/2023	09:45
4	O2794-02		1	14	100.00	>212.0	>212.0	05/17/2023	10:15
5	O2794-03		1	12	100.00	>212.0	>212.0	05/17/2023	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB125515

WorkList Name : f p s 5-16

WorkList ID : 170021

Department : Wet-Chemistry

Date : 05-16-2023 08:20:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2023	Water	O2782-04	Flash Point	Cool 4 deg C	PSEG03	O11	D3032	05/12/2023	1010B
05/18/2023	Water	O2794-02	Flash Point	Cool 4 deg C	PSEG03	O11	AUD-1997	05/15/2023	1010B
05/18/2023	Water	O2794-03	Flash Point	Cool 4 deg C	PSEG03	O11	A287	05/15/2023	1010B

Date/Time 05/17/2023 08:30
 Raw Sample Received by: RM (CWC)
 Raw Sample Relinquished by: JBC (CWC)

Date/Time 05/17/2023 11:20
 Raw Sample Received by: JBC (CWC)
 Raw Sample Relinquished by: RM (CWC)