

LB125518


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

62203



A= 0.0000e+000

B= 1.5925e-004

C= -4.1314e-003

Rho= 0.9998101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	11	0.000	11					-1
0.2	0.200	0.198	-0.002	1267	0.0 %	1267					-1
2.5	2.500	2.467	-0.033	15518	0.0 %	15518					-1
5.0	5.000	4.994	-0.006	31385	0.0 %	31385					2
7.5	7.500	7.642	0.142	48015	0.0 %	48015					2
10.0	10.000	9.902	-0.098	62203	0.0 %	62203					-1

LB125518 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 May 2023 11:33:58

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7471B	PPB	17 May 2023 11:45:04	S	Std
0.2 - 1	50.2	1267	-	0.2000	7471B	PPB	17 May 2023 12:04:30	S	Std
2.5 - 1	52.5	15518	-	2.5000	7471B	PPB	17 May 2023 12:06:47	S	Std
5.0 - 1	55	31385	-	5.0000	7471B	PPB	17 May 2023 12:09:04	S	Std
7.5 - 1	57.5	48015	-	7.5000	7471B	PPB	17 May 2023 12:11:22	S	Std
10.0 - 1	51.0	62203	-	10.0000	7471B	PPB	17 May 2023 12:13:40	S	Std
ICV80 - 1	ICV80	25682	4.0857	-	7471B	PPB	17 May 2023 12:16:30	U	SMPL
ICB80 - 1	ICB80	12	-0.0022	-	7471B	PPB	17 May 2023 12:18:46	U	SMPL
CCV72 - 1	CCV72	31846	5.0673	-	7471B	PPB	17 May 2023 12:21:03	U	SMPL
CCB72 - 1	CCB72	41	0.0024	-	7471B	PPB	17 May 2023 12:23:18	U	SMPL
CRA - 1	CRA	1259	0.1964	-	7471B	PPB	17 May 2023 12:25:35	U	SMPL
HighStd - 1	HighStd	61739	9.8277	-	7471B	PPB	17 May 2023 12:27:51	U	SMPL
ChkStd - 1	ChkStd	44158	7.0280	-	7471B	PPB	17 May 2023 12:30:06	U	SMPL
PB152855BL - 1	PB152855BL	-178	-0.0325	-	7471B	PPB	17 May 2023 12:32:25	U	SMPL
PB152855BS - 1	PB152855BS	25800	4.1045	-	7471B	PPB	17 May 2023 12:34:43	U	SMPL
O2500-11 - 1	PT-MET-SOIL	685494	HIGH	-	7471B	PPB	17 May 2023 12:37:00	X	
O2794-01 - 1	AUD-1927	42	0.0026	-	7471B	PPB	17 May 2023 12:41:49	U	SMPL
O2798-01 - 1	WATERWATCH-AVE-SAMPLE	2183	0.3435	-	7471B	PPB	17 May 2023 12:44:06	U	SMPL
O2801-01 - 1	CF-19	3402	0.5376	-	7471B	PPB	17 May 2023 12:46:23	U	SMPL
O2801-02 - 1	CF-20	3644	0.5762	-	7471B	PPB	17 May 2023 12:48:39	U	SMPL
CCV73 - 1	CCV73	30527	4.8572	-	7471B	PPB	17 May 2023 12:50:54	U	SMPL
CCB73 - 1	CCB73	79	0.0084	-	7471B	PPB	17 May 2023 12:53:10	U	SMPL
O2802-01 - 1	ORA-1639	1736	0.2723	-	7471B	PPB	17 May 2023 12:55:27	U	SMPL
O2802-03 - 1	ORA-1640	628	0.0958	-	7471B	PPB	17 May 2023 12:57:42	U	SMPL
O2807-01 - 1	360	1399	0.2187	-	7471B	PPB	17 May 2023 12:59:58	U	SMPL
O2811-01 - 1	72-11995	6635	1.0525	-	7471B	PPB	17 May 2023 13:02:15	U	SMPL
O2813-01 - 1	WC-1	9171	1.4563	-	7471B	PPB	17 May 2023 13:04:32	U	SMPL
O2813-01DUP - 1	WC-1DUP	7780	1.2348	-	7471B	PPB	17 May 2023 13:06:48	U	SMPL
O2813-01MS - 1	WC-1MS	31398	4.9960	-	7471B	PPB	17 May 2023 13:09:05	U	SMPL
O2813-01MSD - 1	WC-1MSD	38218	5.7635	-	7471B	PPB	17 May 2023 13:11:22	U	SMPL
O2813-01LX5 - 1		816	0.1258	-	7471B	PPB	17 May 2023 13:13:41	U	SMPL
O2813-01A - 1		34235	5.4477	-	7471B	PPB	17 May 2023 13:15:59	U	SMPL
CCV74 - 1	CCV74	31055	4.9413	-	7471B	PPB	17 May 2023 13:18:15	U	SMPL
CCB74 - 1	CCB74	-5	-0.0049	-	7471B	PPB	17 May 2023 13:20:33	U	SMPL
O2500-11DLX10 - 1	PT-MET-SOIL	128966	20.5335	-	7471B	PPB	17 May 2023 13:38:28	U	SMPL
O2500-11DL2X50 - 1	PT-MET-SOIL	25330	4.0296	-	7471B	PPB	17 May 2023 13:43:21	U	SMPL
CCV75 - 1	CCV75	30294	4.8201	-	7471B	PPB	17 May 2023 13:48:00	U	SMPL
CCB75 - 1	CCB75	47	0.0034	-	7471B	PPB	17 May 2023 13:50:18	U	SMPL