



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125519

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/17/2023	13:25
2	CAL2	1	Water	NA	NA	20.7	7.00	05/17/2023	13:26
3	CAL3	1	Water	NA	NA	20.7	10.03	05/17/2023	13:29
4	ICV	1	Water	NA	NA	20.7	6.98	05/17/2023	13:30
5	CCV1	1	Water	NA	NA	20.7	2.01	05/17/2023	13:34
6	O2806-01	1	Water	NA	NA	22.1	5.23	05/17/2023	13:38
7	O2806-01DUP	1	Water	NA	NA	22.3	5.24	05/17/2023	13:39
8	CCV2	1	Water	NA	NA	20.8	12.02	05/17/2023	13:42

VB 125519

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02806 WorkList ID : 170086 Department : Wet-Chemistry Date : 05-17-2023 13:11:55

Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
Water	O2806-01	pH	Cool 4 deg C	PSEG03	O11	33922	05/16/2023	9040C

Date/Time 05/17/23 13:20
Raw Sample Received by: JPLWCJ
Raw Sample Relinquished by: 12/20/

Date/Time 05/17/23 14:25
Raw Sample Received by: JPLWCJ
Raw Sample Relinquished by: JPLWCJ