

LB125518


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

62203



A= 0.0000e+000

B= 1.5925e-004

C= -4.1314e-003

Rho= 0.9998101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	11	0.000	11					1
0.2	0.200	0.198	-0.002	1267	0.0 %	1267					1
2.5	2.500	2.467	-0.033	15518	0.0 %	15518					1
5.0	5.000	4.994	-0.006	31385	0.0 %	31385					1
7.5	7.500	7.642	0.142	48015	0.0 %	48015					2
10.0	10.000	9.902	-0.098	62203	0.0 %	62203					1

LB125522 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 May 2023 15:50:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000	7470A	PPB	17 May 2023 16:00:48	S	Std
0.2 - 1	51.2	915	-	0.2000	7470A	PPB	17 May 2023 16:03:05	S	Std
2.5 - 1	52.5	10536	-	2.5000	7470A	PPB	17 May 2023 16:05:23	S	Std
5.0 - 1	55	21247	-	5.0000	7470A	PPB	17 May 2023 16:07:40	S	Std
7.5 - 1	57.1	31500	-	7.5000	7470A	PPB	17 May 2023 16:12:29	S	Std
10.0 - 1	510	43068	-	10.0000	7470A	PPB	17 May 2023 16:14:46	S	Std
ICV81 - 1	ICV81	17825	4.1842	-	7470A	PPB	17 May 2023 16:17:39	U	SMPL
ICB81 - 1	ICB81	-130	-0.0201	-	7470A	PPB	17 May 2023 16:19:54	U	SMPL
CCV76 - 1	CCV76	19431	4.5603	-	7470A	PPB	17 May 2023 16:22:11	U	SMPL
CCB76 - 1	CCB76	-93	-0.0115	-	7470A	PPB	17 May 2023 16:24:26	U	SMPL
CRA - 1	CRA	868	0.2136	-	7470A	PPB	17 May 2023 16:26:43	U	SMPL
HighStd - 1	HighStd	42134	9.8765	-	7470A	PPB	17 May 2023 16:28:58	U	SMPL
ChkStd - 1	ChkStd	29499	6.9178	-	7470A	PPB	17 May 2023 16:31:14	U	SMPL
PB152867BL - 1	PBW	-250	-0.0482	-	7470A	PPB	17 May 2023 16:33:32	U	SMPL
PB152867BS - 1	LCSW	17113	4.0175	-	7470A	PPB	17 May 2023 16:35:50	U	SMPL
O2794-02 - 1	AUD-1997	-99	-0.0129	-	7470A	PPB	17 May 2023 16:38:06	U	SMPL
O2794-03 - 1	A287	58	0.0239	-	7470A	PPB	17 May 2023 16:40:25	U	SMPL
O2806-01 - 1	33922	735	0.1824	-	7470A	PPB	17 May 2023 16:42:42	U	SMPL
O2812-19 - 1	WG-0516-B3-GW	24763	5.8088	-	7470A	PPB	17 May 2023 16:45:00	U	SMPL
O2812-19DUP - 1	WG-0516-B3-GWDUP	23410	5.4920	-	7470A	PPB	17 May 2023 16:47:15	U	SMPL
CCV77 - 1	CCV77	22066	5.1773	-	7470A	PPB	17 May 2023 16:49:33	U	SMPL
CCB77 - 1	CCB77	8	0.0122	-	7470A	PPB	17 May 2023 16:51:49	U	SMPL
O2812-19MS - 1	WG-0516-B3-GWMS	39828	9.3365	-	7470A	PPB	17 May 2023 16:54:05	U	SMPL
O2812-19MSD - 1	WG-0516-B3-GWMSD	39516	9.2634	-	7470A	PPB	17 May 2023 16:59:10	U	SMPL
O2812-20 - 1	FIELD-BLANK	-136	-0.0215	-	7470A	PPB	17 May 2023 17:01:25	U	SMPL
O2812-19LX5 - 1		3654	0.8659	-	7470A	PPB	17 May 2023 17:03:43	U	SMPL
O2812-19A - 1		42284	9.9116	-	7470A	PPB	17 May 2023 17:05:59	U	SMPL
CCV78 - 1	CCV78	22990	5.3937	-	7470A	PPB	17 May 2023 17:11:01	U	SMPL
CCB78 - 1	CCB78	-6	0.0089	-	7470A	PPB	17 May 2023 17:13:17	U	SMPL