



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB125525

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/17/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	05/17/2023	16:11
3	CAL3	1	Water	NA	NA	20.8	10.03	05/17/2023	16:15
4	ICV	1	Water	NA	NA	20.6	6.99	05/17/2023	16:20
5	CCV1	1	Water	NA	NA	20.7	2.01	05/17/2023	16:22
6	O2820-01	1	Water	NA	NA	22.8	5.34	05/17/2023	16:25
7	O2820-01DUP	1	Water	NA	NA	22.9	5.35	05/17/2023	16:26
8	O2820-02	1	Water	NA	NA	22.4	7.38	05/17/2023	16:30
9	CCV2	1	Water	NA	NA	20.8	12.02	05/17/2023	16:35

VB125525

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O2820 WorkList ID : 170093 Department : Wet-Chemistry Date : 05-17-2023 15:52:28

Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
Water	O2820-01	Corrosivity	Cool 4 deg C	PSEG03	O11	8321	05/17/2023	9040C
Water	O2820-02	Corrosivity	Cool 4 deg C	PSEG03	O11	8327	05/17/2023	9040C

Date/Time 05/17/23 16:00
Raw Sample Received by: 2060C1
Raw Sample Relinquished by: CP SM

Date/Time 05-17-23 17:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 2060C1