

LB125528


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

57042



A= 0.0000e+000

B= 1.7195e-004

C= -4.5166e-002

Rho= 0.9984211

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o)
0.0	0.000	-0.042	-0.042	20	0.000	20					
0.2	0.200	0.186	-0.014	1344	0.0 %	1344					-7
2.5	2.500	2.619	0.119	15492	0.0 %	15492					5
5.0	5.000	4.795	-0.205	28150	0.0 %	28150					-4
7.5	7.500	7.879	0.379	46085	0.0 %	46085					8
10.0	10.000	9.763	-0.237	57042	0.0 %	57042					-2

LB125528 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 May 2023 14:00:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7471B	PPB	18 May 2023 14:06:41	S	Std
0.2 - 1	50.2	1344	-	0.2000	7471B	PPB	18 May 2023 14:08:59	S	Std
2.5 - 1	52.5	15492	-	2.5000	7471B	PPB	18 May 2023 14:11:17	S	Std
5.0 - 1	55	28150	-	5.0000	7471B	PPB	18 May 2023 14:13:34	S	Std
7.5 - 1	57.5	46085	-	7.5000	7471B	PPB	18 May 2023 14:15:51	S	Std
10.0 - 1	510	57042	-	10.0000	7471B	PPB	18 May 2023 14:20:45	S	Std
ICV82 - 1	ICV82	22731	3.8634	-	7471B	PPB	18 May 2023 14:26:23	U	SMPL
ICB82 - 1	ICB82	-110	-0.0641	-	7471B	PPB	18 May 2023 14:28:37	U	SMPL
CCV79 - 1	CCV79	28401	4.8383	-	7471B	PPB	18 May 2023 14:30:55	U	SMPL
CCB79 - 1	CCB79	-144	-0.0699	-	7471B	PPB	18 May 2023 14:33:10	U	SMPL
CRA - 1	CRA	1319	0.1816	-	7471B	PPB	18 May 2023 14:35:28	U	SMPL
HighStd - 1	HighStd	57306	9.8084	-	7471B	PPB	18 May 2023 14:37:43	U	SMPL
ChkStd - 1	ChkStd	41929	7.1644	-	7471B	PPB	18 May 2023 14:39:59	U	SMPL
PB152877BL - 1	PBS	-125	-0.0667	-	7471B	PPB	18 May 2023 14:42:18	U	SMPL
PB152877BS - 1	LCSS	19530	3.3130	-	7471B	PPB	18 May 2023 14:44:37	U	SMPL
O2810-01 - 1	OR-02-051623	4168	0.6715	-	7471B	PPB	18 May 2023 14:46:53	U	SMPL
O2812-02 - 1	WG-0516-B5-0-10	32443	5.5333	-	7471B	PPB	18 May 2023 14:49:12	U	SMPL
O2812-04 - 1	WG-0516-B5-10-20	3713	0.5933	-	7471B	PPB	18 May 2023 14:51:29	U	SMPL
O2812-06 - 1	WG-0516-B3-0-10	47522	8.1261	-	7471B	PPB	18 May 2023 14:53:48	U	SMPL
O2812-08 - 1	WG-0516-B3-10-20	312	0.0085	-	7471B	PPB	18 May 2023 14:56:03	U	SMPL
CCV80 - 1	CCV80	28133	4.7922	-	7471B	PPB	18 May 2023 14:58:22	U	SMPL
CCB80 - 1	CCB80	-68	-0.0569	-	7471B	PPB	18 May 2023 15:00:38	U	SMPL
O2812-09 - 1	WG-0516-BD	662	0.0687	-	7471B	PPB	18 May 2023 15:02:55	U	SMPL
O2812-12 - 1	WG-0516-B4-0-10	25329	4.3101	-	7471B	PPB	18 May 2023 15:05:10	U	SMPL
O2812-14 - 1	WG-0516-B4-10-13	5951	0.9781	-	7471B	PPB	18 May 2023 15:07:26	U	SMPL
O2812-16 - 1	WG-0516-B2-0-10	6704	1.1076	-	7471B	PPB	18 May 2023 15:09:44	U	SMPL
O2812-18 - 1	WG-0516-B1-0-3	2438	0.3740	-	7471B	PPB	18 May 2023 15:12:00	U	SMPL
O2823-01 - 1	HD-01-051723	5845	0.9599	-	7471B	PPB	18 May 2023 15:14:17	U	SMPL
O2823-03 - 1	HD-02-051723	2206	0.3341	-	7471B	PPB	18 May 2023 15:16:34	U	SMPL
O2825-01 - 1	OK-01-051723	89786	15.3933	-	7471B	PPB	18 May 2023 15:18:51	U	SMPL
O2826-01 - 1	SU-3-051723	2351	0.3591	-	7471B	PPB	18 May 2023 15:26:06	U	SMPL
O2826-01DUP - 1	SU-3-051723DUP	2313	0.3525	-	7471B	PPB	18 May 2023 15:28:22	U	SMPL
CCV81 - 1	CCV81	28880	4.9207	-	7471B	PPB	18 May 2023 15:30:37	U	SMPL
CCB81 - 1	CCB81	-95	-0.0615	-	7471B	PPB	18 May 2023 15:32:53	U	SMPL
O2826-01MS - 1	SU-3-051723MS	25623	4.3606	-	7471B	PPB	18 May 2023 15:35:11	U	SMPL
O2826-01MSD - 1	SU-3-051723MSD	23849	4.0556	-	7471B	PPB	18 May 2023 15:37:26	U	SMPL
O2826-01LX5 - 1		404	0.0243	-	7471B	PPB	18 May 2023 15:39:44	U	SMPL
O2826-01A - 1		24588	4.1827	-	7471B	PPB	18 May 2023 15:42:02	U	SMPL
O2825-01DLX5 - 1		17549	2.9723	-	7471B	PPB	18 May 2023 15:44:18	U	SMPL
CCV82 - 1	CCV82	28745	4.8974	-	7471B	PPB	18 May 2023 15:46:39	U	SMPL
CCB82 - 1	CCB82	-131	-0.0677	-	7471B	PPB	18 May 2023 15:48:57	U	SMPL