



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125531

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/18/2023	12:00
2	O2806-01		1	14	100.00	>212.0	>212.0	05/18/2023	12:30
3	O2806-01DUP		1	12	100.00	>212.0	>212.0	05/18/2023	13:00
4	O2820-01		1	12	100.00	>212.0	>212.0	05/18/2023	13:30
5	O2820-02		1	13	100.00	>212.0	>212.0	05/18/2023	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB125531

WorkList Name : fp-5-18

WorkList ID : 170097

Department : Wet-Chemistry

Date : 05-18-2023 08:20:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2806-01	33922	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/16/2023	1010B
O2820-01	8321	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/17/2023	1010B
O2820-02	8327	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/17/2023	1010B

Date/Time 05/18/2023 11:50
 Raw Sample Received by: PM CWL
 Raw Sample Relinquished by: JH(CW)

Date/Time 05/18/2023 14:35
 Raw Sample Received by: JH(CW)
 Raw Sample Relinquished by: PM CWL