


LB125540

7470A INSTRUMENT ID: CUE

Linear

μ Abs.:

45062



A= 0.0000e+000

B= 2.2142e-004

C= 1.6271e-002

Rho= 0.9999798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	0.010	0.010	-28	0.000	-28	0				
0.2	0.200	0.220	0.020	919	0.0 %	919					
2.5	2.500	2.463	-0.037	11052	0.0 %	11052					
5.0	5.000	4.981	-0.019	22420	0.0 %	22420					
7.5	7.500	7.532	0.032	33945	0.0 %	33945					
10.0	10.000	9.994	-0.006	45062	0.0 %	45062					

LB125540 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 May 2023 16:17:39

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.000	7470A	PPB	18 May 2023 16:18:42	S	Std
0.2 - 1	50.2	918	-	0.200	7470A	PPB	18 May 2023 16:21:00	S	Std
2.5 - 1	52.5	11052	-	2.500	7470A	PPB	18 May 2023 16:23:17	S	Std
5.0 - 1	55	22420	-	5.000	7470A	PPB	18 May 2023 16:25:34	S	Std
7.5 - 1	57.5	33945	-	7.500	7470A	PPB	18 May 2023 16:27:51	S	Std
10.0 - 1	510	45062	-	10.000	7470A	PPB	18 May 2023 16:30:08	S	Std
ICV83 - 1	ICV83	17389	3.8665	-	7470A	PPB	18 May 2023 16:33:04	U	SMPL
ICB83 - 1	ICB83	-113	-0.0087	-	7470A	PPB	18 May 2023 16:35:19	U	SMPL
CCV83 - 1	CCV83	22272	4.9477	-	7470A	PPB	18 May 2023 16:37:34	U	SMPL
CCB83 - 1	CCB83	-145	-0.0158	-	7470A	PPB	18 May 2023 16:39:50	U	SMPL
CRA - 1	CRA	864	0.2076	-	7470A	PPB	18 May 2023 16:42:06	U	SMPL
HighStd - 1	HighStd	44548	9.8801	-	7470A	PPB	18 May 2023 16:44:21	U	SMPL
ChkStd - 1	ChkStd	31385	6.9655	-	7470A	PPB	18 May 2023 16:46:37	U	SMPL
PB152885BL - 1	PBW	-230	-0.0347	-	7470A	PPB	18 May 2023 16:48:56	U	SMPL
PB152885BS - 1	LCSW	15556	3.4607	-	7470A	PPB	18 May 2023 16:51:14	U	SMPL
O2779-04 - 1	API-1-BOTTOMS-2	-156	-0.0183	-	7470A	PPB	18 May 2023 16:53:31	U	SMPL
O2798-02 - 1	WATERWITCH-AVE-SAMPLE	202	0.0610	-	7470A	PPB	18 May 2023 16:55:48	U	SMPL
O2798-02DUP - 1	WATERWITCH-AVE-SAMPLEDUP	141	0.0475	-	7470A	PPB	18 May 2023 16:58:05	U	SMPL
O2798-02MS - 1	WATERWITCH-AVE-SAMPLEMS	16086	3.5780	-	7470A	PPB	18 May 2023 17:00:22	U	SMPL
O2798-02MSD - 1	WATERWITCH-AVE-SAMPLEMSD	15423	3.4312	-	7470A	PPB	18 May 2023 17:02:37	U	SMPL
CCV84 - 1	CCV84	22229	4.9382	-	7470A	PPB	18 May 2023 17:04:54	U	SMPL
CCB84 - 1	CCB84	-109	-0.0079	-	7470A	PPB	18 May 2023 17:07:10	U	SMPL
PB152839TB - 1	PB152839TB	-4	0.0154	-	7470A	PPB	18 May 2023 17:09:26	U	SMPL
O2798-02LX5 - 1		-8	0.0145	-	7470A	PPB	18 May 2023 17:11:41	U	SMPL
O2798-02A - 1		18867	4.1938	-	7470A	PPB	18 May 2023 17:13:57	U	SMPL
CCV85 - 1	CCV85	22031	4.8944	-	7470A	PPB	18 May 2023 17:16:13	U	SMPL
CCB85 - 1	CCB85	-131	-0.0127	-	7470A	PPB	18 May 2023 17:18:29	U	SMPL