



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 05/19/2023 16:00 TEMP1 OUT: 103 °C 05/20/2023 07:00
TEMP2 IN: 104 °C 05/20/2023 07:30 TEMP2 OUT: 104 °C 05/20/2023 09:00
TEMP3 IN: 540 °C 05/20/2023 09:30 TEMP3 OUT: 540 °C 05/20/2023 11:00
TEMP4 IN: 560 °C 05/20/2023 11:30 TEMP4 OUT: 560 °C 05/20/2023 13:00

Run Number: LB125548

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB125548BL	85.9671	85.9671	100	85.9671	85.9671	85.9671	85.9671	0.0000	0
2	O2840-06	89.9661	89.9661	100	91.8713	91.8713	91.6831	91.6831	0.1882	1882
3	O2840-06DUP	89.9456	89.9456	100	91.8827	91.8827	91.6952	91.6952	0.1875	1875

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB 125548

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs 02840

WorkList ID : 170139

Department : Wet-Chemistry

Date : 05-19-2023 08:30:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2840-06	GROUND-WATER	Water	TVS	Cool 4 deg C	PSEG03	O11	05/18/2023	SM2540 G

Date/Time 05-19-23 13:00
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: JH

Date/Time 05-19-23 16:10
Raw Sample Received by: JH
Raw Sample Relinquished by: JH (WC)