



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125549

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/19/2023	08:15
2	CAL2	1	Water	NA	NA	20.7	6.99	05/19/2023	08:16
3	CAL3	1	Water	NA	NA	20.8	10.02	05/19/2023	08:17
4	ICV	1	Water	NA	NA	20.7	7.01	05/19/2023	08:20
5	CCV1	1	Water	NA	NA	20.8	2.01	05/19/2023	08:22
6	O2839-01	1	Water	NA	NA	20.4	5.12	05/19/2023	08:25
7	O2840-06	1	Water	NA	NA	22.9	6.66	05/19/2023	08:35
8	O2842-01	1	Water	NA	NA	20.9	11.64	05/19/2023	08:45
9	O2842-01DUP	1	Water	NA	NA	20.8	11.65	05/19/2023	08:46
10	CCV2	1	Water	NA	NA	20.8	12.02	05/19/2023	08:50

13125549

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2840

WorkList ID : 170124

Department : Wet-Chemistry

Date : 05-19-2023 07:57:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2839-01	POOL-WATER	Water	pH	Cool 4 deg C	STER01	O11	05/17/2023	9040C
O2840-06	GROUND-WATER	Water	pH	Cool 4 deg C	PSEG03	O11	05/18/2023	9040C
O2842-01	TOTE-1	Water	pH	Cool 4 deg C	PSEG03	O11	05/18/2023	9040C

Date/Time 05-19-23 08:10

Raw Sample Received by: SD(WC)

Raw Sample Relinquished by: CD gm

Date/Time 05-19-23 10:00

Raw Sample Received by: CD gm

Raw Sample Relinquished by: SD(WC)