

LB125558


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58261



A= 0.0000e+000

B= 1.7239e-004

C= -3.3686e-002

Rho= 0.9997032

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Vol
0.0	0.000	-0.031	-0.031	13	0.000	13					0.0
0.2	0.200	0.174	-0.026	1204	0.0 %	1204					13
2.5	2.500	2.517	0.017	14793	0.0 %	14793					1
5.0	5.000	5.167	0.167	30168	0.0 %	30168					3
7.5	7.500	7.364	-0.136	42911	0.0 %	42911					2
10.0	10.000	10.010	0.010	58261	0.0 %	58261					0

LB125558 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 May 2023 15:31:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7471B	PPB	19 May 2023 15:51:05	S	Std
0.2 - 1	50.2	1204	-	0.2000	7471B	PPB	19 May 2023 15:53:23	S	Std
2.5 - 1	52.5	14793	-	2.5000	7471B	PPB	19 May 2023 15:55:41	S	Std
5.0 - 1	55	30168	-	5.0000	7471B	PPB	19 May 2023 15:57:58	S	Std
7.5 - 1	57.5	42911	-	7.5000	7471B	PPB	19 May 2023 16:00:15	S	Std
10.0 - 1	60	58261	-	10.0000	7471B	PPB	19 May 2023 16:09:31	S	Std
ICV85 - 1	ICV85	25025	4.2805	-	7471B	PPB	19 May 2023 16:12:16	U	SMPL
ICB85 - 1	ICB85	-250	-0.0768	-	7471B	PPB	19 May 2023 16:14:32	U	SMPL
CCV91 - 1	CCV91	30244	5.1802	-	7471B	PPB	19 May 2023 16:19:20	U	SMPL
CCB91 - 1	CCB91	-229	-0.0732	-	7471B	PPB	19 May 2023 16:21:35	U	SMPL
CRA - 1	CRA	1309	0.1920	-	7471B	PPB	19 May 2023 16:26:27	U	SMPL
HighStd - 1	HighStd	58867	10.1146	-	7471B	PPB	19 May 2023 16:28:42	U	SMPL
ChkStd - 1	ChkStd	39857	6.8374	-	7471B	PPB	19 May 2023 16:30:58	U	SMPL
PB152898BL - 1	PBS	85	-0.0190	-	7471B	PPB	19 May 2023 16:35:58	U	SMPL
PB152898BS - 1	LCSS	21699	3.7071	-	7471B	PPB	19 May 2023 16:38:14	U	SMPL
O2845-01 - 1	TP1	1617	0.2451	-	7471B	PPB	19 May 2023 16:40:30	U	SMPL
O2845-01DUP - 1	TP1DUP	1657	0.2520	-	7471B	PPB	19 May 2023 16:42:49	U	SMPL
O2845-01MS - 1	TP1MS	27157	4.6480	-	7471B	PPB	19 May 2023 16:45:06	U	SMPL
O2845-01MSD - 1	TP1MSD	25541	4.3694	-	7471B	PPB	19 May 2023 16:47:23	U	SMPL
O2845-11 - 1	TP2	3136	0.5069	-	7471B	PPB	19 May 2023 16:49:42	U	SMPL
CCV92 - 1	CCV92	29899	5.1207	-	7471B	PPB	19 May 2023 16:52:00	U	SMPL
CCB92 - 1	CCB92	-246	-0.0761	-	7471B	PPB	19 May 2023 16:54:16	U	SMPL
O2845-21 - 1	TP3	1773	0.2720	-	7471B	PPB	19 May 2023 16:56:34	U	SMPL
O2846-01 - 1	EO-01-051823	2944	0.4738	-	7471B	PPB	19 May 2023 16:58:50	U	SMPL
O2848-01 - 1	TP4	2725	0.4361	-	7471B	PPB	19 May 2023 17:01:06	U	SMPL
O2848-10 - 1	TP5	1707	0.2606	-	7471B	PPB	19 May 2023 17:03:22	U	SMPL
O2848-20 - 1	TP6	3436	0.5587	-	7471B	PPB	19 May 2023 17:05:38	U	SMPL
O2849-01 - 1	TP7	1968	0.3056	-	7471B	PPB	19 May 2023 17:07:55	U	SMPL
O2849-11 - 1	TP8	1000	0.1387	-	7471B	PPB	19 May 2023 17:10:11	U	SMPL
O2849-21 - 1	TP9	3325	0.5395	-	7471B	PPB	19 May 2023 17:12:29	U	SMPL
O2849-31 - 1	TP10	1876	0.2897	-	7471B	PPB	19 May 2023 17:14:46	U	SMPL
O2849-41 - 1	TP11	1733	0.2651	-	7471B	PPB	19 May 2023 17:17:02	U	SMPL
CCV93 - 1	CCV93	30274	5.1853	-	7471B	PPB	19 May 2023 17:19:18	U	SMPL
CCB93 - 1	CCB93	39	-0.0270	-	7471B	PPB	19 May 2023 17:24:15	U	SMPL
O2845-01LX5 - 1		402	0.0356	-	7471B	PPB	19 May 2023 17:26:31	U	SMPL
O2845-01A - 1		26024	4.4527	-	7471B	PPB	19 May 2023 17:33:29	U	SMPL
CCV94 - 1	CCV94	30311	5.1917	-	7471B	PPB	19 May 2023 17:35:46	U	SMPL
CCB94 - 1	CCB94	-82	-0.0478	-	7471B	PPB	19 May 2023 17:38:04	U	SMPL