



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB125569
Analysis Date: 05/22/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 05/22/2023
Extraction IN Time: 08:20
Extraction OUT Time: 08:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB125569BL	LB125569BL	WATER	1.3	1000	100	2.9854	2.9854	3.01	2.9855	2.9855	0.0001	0.1
2	LB125569BS	LB125569BS	WATER	1.3	1000	100	2.9901	2.9901	3.02	3.0069	3.0069	0.0168	16.8
3	LB125569BSD	LB125569BSD	WATER	1.3	1000	100	3.0148	3.0148	3.03	3.0318	3.0318	0.0170	17
4	O2872-01	OUTFALL-DSN-001	WATER	1.3	1000	100	2.9653	2.9653	3.02	2.9654	2.9654	0.0001	0.1
5	O2872-02	OUTFALL-DSN-002	WATER	1.6	1000	100	3.0177	3.0177	3.05	3.0191	3.0191	0.0014	1.4



QC Batch# LB125569 Test: Non-Polar Material Analysis Date: 05/22/2023

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3034
pH Paper 0-14	M4909
Sodium Sulfate	EP2340
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:01
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time1: 09:15
Bal Check Time: 08:25 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 10:35
Out Time1: 10:00

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:01
1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 11:10
Bal Check Time: 12:40 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 12:37
Out Time2: 12:00

W125569

WORKLIST(Hardcopy Internal Chain)

WorkList Name : non poplar o2872 WorkList ID : 170178 Department : Wet-Chemistry Date : 05-22-2023 08:01:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2872-01	OUTFALL-DSN-001	Water	Non-Polar Material	Conc H2SO4 to pH < 2	TRIS02	N11	05/18/2023	1664A
O2872-02	OUTFALL-DSN-002	Water	Non-Polar Material	Conc H2SO4 to pH < 2	TRIS02	N11	05/18/2023	1664A

Date/Time 05/22/23 08:13
Raw Sample Received by: JP (w/c)
Raw Sample Relinquished by: JP sm

Date/Time 05/22/23 12:00
Raw Sample Received by: JP sm
Raw Sample Relinquished by: JP (w/c)