

LB125570


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

59213



A= 0.0000e+000

B= 1.6722e-004

C= 4.4287e-003

Rho= 0.9995306

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/6/11
0.0	0.000	0.007	0.007	13	0.000	13					3
0.2	0.200	0.207	0.007	1211	0.0 %	1211					-6
2.5	2.500	2.345	-0.155	13996	0.0 %	13996					4
5.0	5.000	5.207	0.207	31112	0.0 %	31112					0
7.5	7.500	7.529	0.029	44998	0.0 %	44998					-1
10.0	10.000	9.906	-0.094	59213	0.0 %	59213					

LB125570 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 May 2023 10:53:07

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7471B	PPB	22 May 2023 11:11:41	S	Std
0.2 - 1	50.2	1211	-	0.2000	7471B	PPB	22 May 2023 11:13:58	S	Std
2.5 - 1	50.5	13996	-	2.5000	7471B	PPB	22 May 2023 11:16:15	S	Std
5.0 - 1	55	31112	-	5.0000	7471B	PPB	22 May 2023 11:18:32	S	Std
7.5 - 1	57.5	44998	-	7.5000	7471B	PPB	22 May 2023 11:20:51	S	Std
10.0 - 1	510	59213	-	10.0000	7471B	PPB	22 May 2023 11:27:12	S	Std
ICV86 - 1	ICV86	23645	3.9583	-	7471B	PPB	22 May 2023 11:30:12	U	SMPL
ICB86 - 1	ICB86	-190	-0.0273	-	7471B	PPB	22 May 2023 11:32:27	U	SMPL
CCV95 - 1	CCV95	28453	4.7623	-	7471B	PPB	22 May 2023 11:34:44	U	SMPL
CCB95 - 1	CCB95	-217	-0.0319	-	7471B	PPB	22 May 2023 11:36:59	U	SMPL
CRA - 1	CRA	1243	0.2123	-	7471B	PPB	22 May 2023 11:39:16	U	SMPL
HighStd - 1	HighStd	59277	9.9166	-	7471B	PPB	22 May 2023 11:41:32	U	SMPL
ChkStd - 1	ChkStd	43003	7.1953	-	7471B	PPB	22 May 2023 11:43:48	U	SMPL
PB152923BL - 1	PB152923BL	-134	-0.0180	-	7471B	PPB	22 May 2023 11:46:06	U	SMPL
PB152923BS - 1	PB152923BS	22975	3.8463	-	7471B	PPB	22 May 2023 11:48:25	U	SMPL
O2840-02 - 1	TP-1	2837	0.4788	-	7471B	PPB	22 May 2023 11:50:41	U	SMPL
O2840-04 - 1	TP-2	4909	0.8253	-	7471B	PPB	22 May 2023 11:53:00	U	SMPL
O2847-01 - 1	TR-04-051823	1145	0.1959	-	7471B	PPB	22 May 2023 11:55:17	U	SMPL
O2850-05 - 1	TRE-1729	1436	0.2446	-	7471B	PPB	22 May 2023 11:57:35	U	SMPL
O2864-01 - 1	ETGI-289	10291	1.7253	-	7471B	PPB	22 May 2023 11:59:50	U	SMPL
CCV96 - 1	CCV96	28377	4.7496	-	7471B	PPB	22 May 2023 12:02:06	U	SMPL
CCB96 - 1	CCB96	-122	-0.0160	-	7471B	PPB	22 May 2023 12:04:21	U	SMPL
O2864-03 - 1	ETGI-280	109449	18.3062	-	7471B	PPB	22 May 2023 12:06:39	U	SMPL
O2864-05 - 1	ETGI-359	33427	5.5940	-	7471B	PPB	22 May 2023 12:08:54	U	SMPL
O2864-07 - 1	SOIL-BIN	29552	4.9460	-	7471B	PPB	22 May 2023 12:11:14	U	SMPL
O2867-01 - 1	TP16	2439	0.4123	-	7471B	PPB	22 May 2023 12:13:32	U	SMPL
O2867-11 - 1	TP15	2079	0.3521	-	7471B	PPB	22 May 2023 12:15:51	U	SMPL
O2867-21 - 1	TP14	1610	0.2736	-	7471B	PPB	22 May 2023 12:18:07	U	SMPL
O2867-21DUP - 1	TP14DUP	1612	0.2740	-	7471B	PPB	22 May 2023 12:20:24	U	SMPL
O2867-21MS - 1	TP14MS	27170	4.5477	-	7471B	PPB	22 May 2023 12:22:41	U	SMPL
O2867-21MSD - 1	TP14MSD	23514	3.9364	-	7471B	PPB	22 May 2023 12:24:58	U	SMPL
PB152924BL - 1	PB152924BL	96	0.0205	-	7471B	PPB	22 May 2023 12:27:16	U	SMPL
CCV97 - 1	CCV97	28644	4.7942	-	7471B	PPB	22 May 2023 12:29:33	U	SMPL
CCB97 - 1	CCB97	-156	-0.0217	-	7471B	PPB	22 May 2023 12:31:50	U	SMPL
PB152924BS - 1	PB152924BS	23026	3.8548	-	7471B	PPB	22 May 2023 12:34:07	U	SMPL
O2868-01 - 1	TP13	1583	0.2691	-	7471B	PPB	22 May 2023 12:36:23	U	SMPL
O2868-11 - 1	TP12	1332	0.2272	-	7471B	PPB	22 May 2023 12:38:41	U	SMPL
O2868-21 - 1	TP17	2223	0.3762	-	7471B	PPB	22 May 2023 12:40:56	U	SMPL
O2869-01 - 1	TP18	2416	0.4084	-	7471B	PPB	22 May 2023 12:43:12	U	SMPL
O2869-11 - 1	TP19	804	0.1389	-	7471B	PPB	22 May 2023 12:45:29	U	SMPL
O2869-21 - 1	TP22	2283	0.3862	-	7471B	PPB	22 May 2023 12:47:46	U	SMPL
O2870-01 - 1	TP21	1638	0.2783	-	7471B	PPB	22 May 2023 12:50:03	U	SMPL
O2870-11 - 1	TP20	1720	0.2920	-	7471B	PPB	22 May 2023 12:52:20	U	SMPL
O2871-01 - 1	ETGI-367	4533	0.7624	-	7471B	PPB	22 May 2023 12:54:36	U	SMPL
CCV98 - 1	CCV98	28499	4.7700	-	7471B	PPB	22 May 2023 12:56:52	U	SMPL
CCB98 - 1	CCB98	-135	-0.0181	-	7471B	PPB	22 May 2023 12:59:07	U	SMPL
O2871-03 - 1	VNJ-226	8506	1.4268	-	7471B	PPB	22 May 2023 13:01:25	U	SMPL
O2873-01 - 1	PL-01-051923	820	0.1415	-	7471B	PPB	22 May 2023 13:03:41	U	SMPL
O2873-01DUP - 1	PL-01-051923DUP	1047	0.1795	-	7471B	PPB	22 May 2023 13:05:57	U	SMPL
O2873-01MS - 1	PL-01-051923MS	25179	4.2148	-	7471B	PPB	22 May 2023 13:08:12	U	SMPL
O2873-01MSD - 1	PL-01-051923MSD	24167	4.0456	-	7471B	PPB	22 May 2023 13:10:28	U	SMPL
O2867-21LX5 - 1		181	0.0347	-	7471B	PPB	22 May 2023 13:12:46	U	SMPL
O2867-21A - 1		25647	4.2931	-	7471B	PPB	22 May 2023 13:15:05	U	SMPL
O2873-01LX5 - 1		7	0.0056	-	7471B	PPB	22 May 2023 13:17:22	U	SMPL
O2873-01A - 1		23590	3.9491	-	7471B	PPB	22 May 2023 13:19:40	U	SMPL
O2864-02DLX5 - 1		20864	3.4933	-	7471B	PPB	22 May 2023 13:21:56	U	SMPL
CCV99 - 1	CCV99	28822	4.8240	-	7471B	PPB	22 May 2023 13:24:14	U	SMPL
CCB99 - 1	CCB99	-126	-0.0166	-	7471B	PPB	22 May 2023 13:26:31	U	SMPL