

LA125582


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

4556E



A= 0.0000e+000

B= 2.1624e-004

C= 2.2497e-002

Rho= 0.9995591

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/vd
0.0	0.000	0.034	0.034	51	0.000	51	0				
0.2	0.200	0.217	0.017	900	0.0 %	900					
2.5	2.500	2.344	-0.156	10737	0.0 %	10737					
5.0	5.000	5.066	0.066	23323	0.0 %	23323					
7.5	7.500	7.663	0.163	35335	0.0 %	35335					
10.0	10.000	9.876	-0.124	45568	0.0 %	45568					

o/vd
-8
-6
-1
-2
-1

LB125582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 May 2023 14:34:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7470A	PPB	22 May 2023 14:38:56	S	Std
0.2 - 1	50.2	900	-	0.2000	7470A	PPB	22 May 2023 14:41:14	S	Std
2.5 - 1	52.5	10737	-	2.5000	7470A	PPB	22 May 2023 14:43:32	S	Std
5.0 - 1	55	23323	-	5.0000	7470A	PPB	22 May 2023 14:45:49	S	Std
7.5 - 1	57.5	35335	-	7.5000	7470A	PPB	22 May 2023 14:48:05	S	Std
10.0 - 1	516	45568	-	10.0000	7470A	PPB	22 May 2023 14:59:30	S	Std
ICV87 - 1	ICV87	18912	4.1120	-	7470A	PPB	22 May 2023 15:04:53	U	SMPL
ICB87 - 1	ICB87	-151	-0.0102	-	7470A	PPB	22 May 2023 15:07:08	U	SMPL
CCV01 - 1	CCV01	21217	4.6104	-	7470A	PPB	22 May 2023 15:09:25	U	SMPL
CCB01 - 1	CCB01	-177	-0.0158	-	7470A	PPB	22 May 2023 15:11:41	U	SMPL
CRA - 1	CRA	811	0.1979	-	7470A	PPB	22 May 2023 15:13:58	U	SMPL
HighStd - 1	HighStd	44558	9.6576	-	7470A	PPB	22 May 2023 15:16:13	U	SMPL
ChkStd - 1	ChkStd	31454	6.8240	-	7470A	PPB	22 May 2023 15:18:29	U	SMPL
PB152925BL - 1	PBW	-257	-0.0331	-	7470A	PPB	22 May 2023 15:20:48	U	SMPL
PB152925BS - 1	LCSW	17604	3.8292	-	7470A	PPB	22 May 2023 15:23:06	U	SMPL
O2840-01 - 1	PAINT-CHIP-DRUM-COMP	764	0.1877	-	7470A	PPB	22 May 2023 15:25:23	U	SMPL
O2845-02 - 1	TP1	-38	0.0143	-	7470A	PPB	22 May 2023 15:27:40	U	SMPL
O2845-12 - 1	TP2	-5	0.0214	-	7470A	PPB	22 May 2023 15:29:58	U	SMPL
O2845-22 - 1	TP3	-3	0.0218	-	7470A	PPB	22 May 2023 15:32:15	U	SMPL
O2848-11 - 1	TP5	6	0.0238	-	7470A	PPB	22 May 2023 15:34:30	U	SMPL
CCV02 - 1	CCV02	21203	4.6074	-	7470A	PPB	22 May 2023 15:41:34	U	SMPL
CCB02 - 1	CCB02	-152	-0.0104	-	7470A	PPB	22 May 2023 15:43:50	U	SMPL
O2848-21 - 1	TP6	-12	0.0199	-	7470A	PPB	22 May 2023 15:46:07	U	SMPL
O2848-30 - 1	TP4	8	0.0242	-	7470A	PPB	22 May 2023 15:48:22	U	SMPL
O2849-02 - 1	TP7	0	0.0225	-	7470A	PPB	22 May 2023 15:50:38	U	SMPL
O2849-12 - 1	TP8	9	0.0244	-	7470A	PPB	22 May 2023 15:52:54	U	SMPL
O2849-22 - 1	TP9	37	0.0305	-	7470A	PPB	22 May 2023 15:55:10	U	SMPL
O2849-32 - 1	TP10	-7	0.0210	-	7470A	PPB	22 May 2023 15:57:27	U	SMPL
O2849-42 - 1	TP11	-20	0.0182	-	7470A	PPB	22 May 2023 15:59:43	U	SMPL
O2867-02 - 1	TP16	19	0.0266	-	7470A	PPB	22 May 2023 16:02:00	U	SMPL
O2867-12 - 1	TP15	-22	0.0177	-	7470A	PPB	22 May 2023 16:04:17	U	SMPL
O2867-22 - 1	TP14	-7	0.0210	-	7470A	PPB	22 May 2023 16:06:33	U	SMPL
CCV03 - 1	CCV03	20964	4.5557	-	7470A	PPB	22 May 2023 16:08:49	U	SMPL
CCB03 - 1	CCB03	-142	-0.0082	-	7470A	PPB	22 May 2023 16:11:05	U	SMPL
O2868-02 - 1	TP13	-18	0.0186	-	7470A	PPB	22 May 2023 16:13:22	U	SMPL
O2868-12 - 1	TP12	-16	0.0190	-	7470A	PPB	22 May 2023 16:15:38	U	SMPL
O2868-22 - 1	TP17	100	0.0441	-	7470A	PPB	22 May 2023 16:17:53	U	SMPL
O2869-02 - 1	TP18	-14	0.0195	-	7470A	PPB	22 May 2023 16:20:10	U	SMPL
PB152929BL - 1	PBW	-7	0.0210	-	7470A	PPB	22 May 2023 16:31:33	U	SMPL
PB152929BS - 1	LCSW	17429	3.7913	-	7470A	PPB	22 May 2023 16:33:50	U	SMPL
CCV04 - 1	CCV04	21927	4.7639	-	7470A	PPB	22 May 2023 16:38:50	U	SMPL
CCB04 - 1	CCB04	-189	-0.0184	-	7470A	PPB	22 May 2023 16:41:06	U	SMPL
O2864-02 - 1	ETGI-289	-49	0.0119	-	7470A	PPB	22 May 2023 16:43:23	U	SMPL
O2864-04 - 1	ETGI-280	0	0.0225	-	7470A	PPB	22 May 2023 16:45:39	U	SMPL
O2864-06 - 1	ETGI-359	-12	0.0199	-	7470A	PPB	22 May 2023 16:47:54	U	SMPL
O2864-08 - 1	SOIL-BIN	-22	0.0177	-	7470A	PPB	22 May 2023 16:50:10	U	SMPL
O2869-22 - 1	TP22	-12	0.0199	-	7470A	PPB	22 May 2023 16:52:26	U	SMPL
O2870-02 - 1	TP21	-30	0.0160	-	7470A	PPB	22 May 2023 16:54:43	U	SMPL
CCV05 - 1	CCV05	22131	4.8081	-	7470A	PPB	22 May 2023 17:06:08	U	SMPL
CCB05 - 1	CCB05	-229	-0.0270	-	7470A	PPB	22 May 2023 17:08:25	U	SMPL
PB152969BL - 1	PBW	-32	0.0156	-	7470A	PPB	22 May 2023 17:10:43	U	SMPL
PB152969BS - 1	LCSW	15863	3.4527	-	7470A	PPB	22 May 2023 17:16:09	U	SMPL
O2840-06 - 1	GROUND-WATER	5881	1.2942	-	7470A	PPB	22 May 2023 17:18:25	U	SMPL
O2850-06 - 1	TRE-1741	1027	0.2446	-	7470A	PPB	22 May 2023 17:39:09	U	SMPL
PB152906TB - 1		-28	0.0164	-	7470A	PPB	22 May 2023 17:41:26	U	SMPL
PB152921TB - 1		-2	0.0221	-	7470A	PPB	22 May 2023 17:44:48	U	SMPL
CCV06 - 1	CCV06	24180	5.2511	-	7470A	PPB	22 May 2023 17:57:21	U	SMPL
CCB06 - 1	CCB06	-182	-0.0169	-	7470A	PPB	22 May 2023 17:59:37	U	SMPL