

LB125593


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47721



A= 0.0000e+000

B= 2.0884e-004

C= 2.9967e-002

Rho= 0.9997935

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	0.036	0.036	27	0.000	27	0				16
0.2	0.200	0.232	0.032	968	0.0 %	968					-1
2.5	2.500	2.475	-0.025	11710	0.0 %	11710					-3
5.0	5.000	4.858	-0.142	23118	0.0 %	23118					1
7.5	7.500	7.603	0.103	36263	0.0 %	36263					0
10.0	10.000	9.996	-0.004	47721	0.0 %	47721					

LB125593 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2023 08:11:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	7470A	PPB	23 May 2023 10:10:34	S	Std
0.2 - 1	50.2	968	-	0.2000	7470A	PPB	23 May 2023 10:12:56	S	Std
2.5 - 1	52.5	11710	-	2.5000	7470A	PPB	23 May 2023 10:15:13	S	Std
5.0 - 1	55	23118	-	5.0000	7470A	PPB	23 May 2023 10:17:30	S	Std
7.5 - 1	57.5	36263	-	7.5000	7470A	PPB	23 May 2023 10:19:48	S	Std
10.0 - 1	510	47721	-	10.0000	7470A	PPB	23 May 2023 10:24:40	S	Std
ICV88 - 1	ICV88	19387	4.0787	-	7470A	PPB	23 May 2023 10:30:06	U	SMPL
ICB88 - 1	ICB88	-86	0.0120	-	7470A	PPB	23 May 2023 10:32:21	U	SMPL
CCV07 - 1	CCV07	22811	4.7938	-	7470A	PPB	23 May 2023 10:34:38	U	SMPL
CCB07 - 1	CCB07	-99	0.0093	-	7470A	PPB	23 May 2023 10:36:53	U	SMPL
CRA - 1	CRA	902	0.2183	-	7470A	PPB	23 May 2023 10:39:11	U	SMPL
HIGH STD - 1	HIGH STD	46890	9.8224	-	7470A	PPB	23 May 2023 10:41:26	U	SMPL
CHK STD - 1	CHK STD	31441	6.5960	-	7470A	PPB	23 May 2023 10:43:42	U	SMPL
O2869-12 - 1	TP19	-259	-0.0241	-	7470A	PPB	23 May 2023 10:46:00	U	SMPL
O2869-12DUP - 1	TP19DUP	-82	0.0128	-	7470A	PPB	23 May 2023 10:48:19	U	SMPL
O2869-12MS - 1	TP19MS	20765	4.3665	-	7470A	PPB	23 May 2023 10:50:35	U	SMPL
O2869-12MSD - 1	TP19MSD	18449	3.8828	-	7470A	PPB	23 May 2023 10:52:52	U	SMPL
O2869-12LX5 - 1		-146	-0.0005	-	7470A	PPB	23 May 2023 10:55:10	U	SMPL
O2869-12A - 1		19115	4.0219	-	7470A	PPB	23 May 2023 10:57:29	U	SMPL
CCV08 - 1	CCV08	23977	5.0373	-	7470A	PPB	23 May 2023 10:59:44	U	SMPL
CCB08 - 1	CCB08	-132	0.0024	-	7470A	PPB	23 May 2023 11:02:01	U	SMPL
O2870-12 - 1	TP20	-82	0.0128	-	7470A	PPB	23 May 2023 11:04:18	U	SMPL
O2870-12DUP - 1	TP20DUP	-46	0.0204	-	7470A	PPB	23 May 2023 11:06:34	U	SMPL
O2870-12MS - 1	TP20MS	19456	4.0931	-	7470A	PPB	23 May 2023 11:08:49	U	SMPL
O2870-12MSD - 1	TP20MSD	19071	4.0127	-	7470A	PPB	23 May 2023 11:11:05	U	SMPL
O2870-12LX5 - 1		-109	0.0072	-	7470A	PPB	23 May 2023 11:13:22	U	SMPL
O2870-12A - 1		19074	4.0133	-	7470A	PPB	23 May 2023 11:15:40	U	SMPL
O2842-01 - 1	TOTE-1	3992	0.8636	-	7470A	PPB	23 May 2023 11:23:28	U	SMPL
O2842-01DUP - 1	TOTE-1DUP	3712	0.8052	-	7470A	PPB	23 May 2023 11:25:45	U	SMPL
O2842-01MS - 1	TOTE-1MS	20029	4.2128	-	7470A	PPB	23 May 2023 11:28:03	U	SMPL
O2842-01MSD - 1	TOTE-1MSD	20165	4.2412	-	7470A	PPB	23 May 2023 11:30:20	U	SMPL
CCV09 - 1	CCV09	23695	4.9784	-	7470A	PPB	23 May 2023 11:32:38	U	SMPL
CCB09 - 1	CCB09	-106	0.0078	-	7470A	PPB	23 May 2023 11:34:55	U	SMPL
O2842-01LX5 - 1		615	0.1584	-	7470A	PPB	23 May 2023 11:37:13	U	SMPL
O2842-01A - 1		19934	4.1929	-	7470A	PPB	23 May 2023 11:39:29	U	SMPL
CCV10 - 1	CCV10	24041	5.0506	-	7470A	PPB	23 May 2023 11:41:45	U	SMPL
CCB10 - 1	CCB10	-130	0.0028	-	7470A	PPB	23 May 2023 11:44:02	U	SMPL