



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125600

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	05/22/2023	18:00
2	CAL2	1	Water	NA	NA	20.7	6.99	05/22/2023	18:01
3	CAL3	1	Water	NA	NA	20.7	10.02	05/22/2023	18:02
4	ICV	1	Water	NA	NA	20.8	7.02	05/22/2023	18:05
5	CCV1	1	Water	NA	NA	20.6	2.01	05/22/2023	18:10
6	O2886-02	1	Solid	20.02	20	21.2	6.83	05/22/2023	18:14
7	O2886-02DUP	1	Solid	20.03	20	21.3	6.84	05/22/2023	18:15
8	O2891-02	1	Solid	20.03	20	21.9	6.13	05/22/2023	18:22
9	O2892-02	1	Solid	20.05	20	22.1	6.85	05/22/2023	18:26
10	O2892-05	1	Solid	20.03	20	21.3	7.07	05/22/2023	18:30
11	O2892-08	1	Solid	20.02	20	21.5	7.88	05/22/2023	18:33
12	O2892-11	1	Solid	20.01	20	22.1	7.82	05/22/2023	18:35
13	O2892-14	1	Solid	20.02	20	22.2	8.15	05/22/2023	18:40
14	O2894-02	1	Solid	20.02	20	21.3	8.56	05/22/2023	18:44
15	O2894-05	1	Solid	20.03	20	22.0	8.07	05/22/2023	18:47
16	CCV2	1	Water	NA	NA	20.6	12.02	05/22/2023	18:50

125600

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O2886 WorkList ID : 170225 Department : Wet-Chemistry Date : 05-22-2023 16:54:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2886-02	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/21/2023	9045D
O2891-02	290-328	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2892-02	TP-30	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2892-05	TP-31	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2892-08	TP-33	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2892-11	TP-32	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2892-14	TP-29	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2894-02	TP-40A	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D
O2894-05	TP-50A	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/22/2023	9045D

Date/Time 05/22/23 17:37
Raw Sample Received by: JP/OC
Raw Sample Relinquished by: JP/OC

Date/Time 05/22/23 19:00
Raw Sample Received by: JP/OC
Raw Sample Relinquished by: JP/OC