



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125601

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/23/2023	12:15
2	CAL2	1	Water	NA	NA	20.7	7.00	05/23/2023	12:16
3	CAL3	1	Water	NA	NA	20.7	10.03	05/23/2023	12:16
4	ICV	1	Water	NA	NA	20.6	6.98	05/23/2023	12:20
5	CCV1	1	Water	NA	NA	20.6	2.01	05/23/2023	12:22
6	O2884-03	1	Water	NA	NA	22.8	5.15	05/23/2023	12:25
7	O2884-03DUP	1	Water	NA	NA	22.7	5.16	05/23/2023	12:26
8	O2884-04	1	Water	NA	NA	22.8	9.09	05/23/2023	12:35
9	O2884-05	1	Water	NA	NA	22.9	6.46	05/23/2023	12:40
10	O2884-06	1	Water	NA	NA	22.9	6.38	05/23/2023	12:42
11	O2891-03	1	Water	NA	NA	22.4	6.63	05/23/2023	12:45
12	CCV2	1	Water	NA	NA	20.7	12.02	05/23/2023	12:47