



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB125609

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	05/23/2023	16:45
2	CAL2	1	Water	NA	NA	20.8	7.00	05/23/2023	16:46
3	CAL3	1	Water	NA	NA	20.8	10.03	05/23/2023	16:47
4	ICV	1	Water	NA	NA	20.6	6.99	05/23/2023	16:50
5	CCV1	1	Water	NA	NA	20.7	2.01	05/23/2023	16:52
6	O2897-02	1	Water	NA	NA	21.4	5.32	05/23/2023	17:00
7	O2897-02DUP	1	Water	NA	NA	21.5	5.33	05/23/2023	17:01
8	CCV2	1	Water	NA	NA	20.9	12.02	05/23/2023	17:02

W3125609

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O2897

WorkList ID : 170260

Department : Wet-Chemistry

Date : 05-23-2023 16:29:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2897-02	OILY-WATER-0523	Water	Corrosivity	Cool 4 deg C	PSEG03	O11	05/23/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

05/23/23 16:32
TP (QC)
AU (5u)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

05/23/23 14:00
DJ (5u)
70191C