

LB125629


7470A

INSTRUMENT ID: CVI

Linear

 μ Abs.:

49796



A= 0.0000e+000

B= 2.0351e-004

C= 1.8153e-002

Rho= 0.9991219

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/0D
0.0	0.000	0.080	0.080	304	0.000	304	0				7
0.2	0.200	0.214	0.014	962	0.0 %	962					-4
2.5	2.500	2.391	-0.109	11661	0.0 %	11661					3
5.0	5.000	5.146	0.146	25198	0.0 %	25198					-4
7.5	7.500	7.216	-0.284	35370	0.0 %	35370					2
10.0	10.000	10.152	0.152	49796	0.0 %	49796					

LB125629 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 May 2023 14:39:14

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	304	-	0.0000	7470A	PPB	24 May 2023 14:40:48	S	Std
0.2 - 1	50.2	962	-	0.2000	7470A	PPB	24 May 2023 14:43:05	S	Std
2.5 - 1	52.5	11661	-	2.5000	7470A	PPB	24 May 2023 14:45:23	S	Std
5.0 - 1	55	25198	-	5.0000	7470A	PPB	24 May 2023 14:50:15	S	Std
7.5 - 1	57.5	35370	-	7.5000	7470A	PPB	24 May 2023 14:55:10	S	Std
10.0 - 1	510	49796	-	10.0000	7470A	PPB	24 May 2023 14:57:26	S	Std
ICV89 - 1	ICV89	19610	4.0090	-	7470A	PPB	24 May 2023 15:00:21	U	SMPL
ICB89 - 1	ICB89	102	0.0389	-	7470A	PPB	24 May 2023 15:02:36	U	SMPL
CCV13 - 1	CCV13	24034	4.9093	-	7470A	PPB	24 May 2023 15:04:53	U	SMPL
CCB13 - 1	CCB13	126	0.0438	-	7470A	PPB	24 May 2023 15:07:08	U	SMPL
CRA - 1	CRA	836	0.1883	-	7470A	PPB	24 May 2023 15:09:24	U	SMPL
HighStd - 1	HighStd	49488	10.0895	-	7470A	PPB	24 May 2023 15:11:40	U	SMPL
ChkStd - 1	ChkStd	33298	6.7947	-	7470A	PPB	24 May 2023 15:13:55	U	SMPL
PB153020BL - 1	PBW	-74	0.0031	-	7470A	PPB	24 May 2023 15:16:14	U	SMPL
PB153020BS - 1	LCSW	18081	3.6978	-	7470A	PPB	24 May 2023 15:18:32	U	SMPL
O2886-02 - 1	WC-1	-35	0.0110	-	7470A	PPB	24 May 2023 15:20:48	U	SMPL
O2891-02 - 1	290-328	134	0.0454	-	7470A	PPB	24 May 2023 15:23:06	U	SMPL
O2892-02 - 1	TP-30	94	0.0373	-	7470A	PPB	24 May 2023 15:25:23	U	SMPL
O2892-05 - 1	TP-31	137	0.0460	-	7470A	PPB	24 May 2023 15:27:40	U	SMPL
O2892-08 - 1	TP-33	94	0.0373	-	7470A	PPB	24 May 2023 15:29:55	U	SMPL
CCV14 - 1	CCV14	24047	4.9120	-	7470A	PPB	24 May 2023 15:32:11	U	SMPL
CCB14 - 1	CCB14	130	0.0446	-	7470A	PPB	24 May 2023 15:34:26	U	SMPL
O2892-11 - 1	TP-32	148	0.0483	-	7470A	PPB	24 May 2023 15:36:43	U	SMPL
O2892-14 - 1	TP-29	332	0.0857	-	7470A	PPB	24 May 2023 15:38:59	U	SMPL
O2894-02 - 1	TP-40A	166	0.0519	-	7470A	PPB	24 May 2023 15:41:14	U	SMPL
O2894-05 - 1	TP-50A	124	0.0434	-	7470A	PPB	24 May 2023 15:43:30	U	SMPL
O2894-05DUP - 1	TP-50ADUP	139	0.0464	-	7470A	PPB	24 May 2023 15:45:46	U	SMPL
O2894-05MS - 1	TP-50AMS	17853	3.6514	-	7470A	PPB	24 May 2023 15:48:03	U	SMPL
O2894-05MSD - 1	TP-50AMSD	18645	3.8126	-	7470A	PPB	24 May 2023 15:50:19	U	SMPL
PB153023BL - 1	PBW	-46	0.0088	-	7470A	PPB	24 May 2023 15:52:38	U	SMPL
PB153023BS - 1	LCSW	17592	3.5983	-	7470A	PPB	24 May 2023 15:54:56	U	SMPL
O2884-04 - 1	2003	281	0.0753	-	7470A	PPB	24 May 2023 15:57:11	U	SMPL
CCV15 - 1	CCV15	24558	5.0160	-	7470A	PPB	24 May 2023 15:59:28	U	SMPL
CCB15 - 1	CCB15	39	0.0261	-	7470A	PPB	24 May 2023 16:01:44	U	SMPL
O2891-03 - 1	A653	2549	0.5369	-	7470A	PPB	24 May 2023 16:09:01	U	SMPL
O2891-03DUP - 1	A653DUP	2454	0.5176	-	7470A	PPB	24 May 2023 16:11:16	U	SMPL
O2891-03MS - 1	A653MS	19571	4.0011	-	7470A	PPB	24 May 2023 16:13:32	U	SMPL
O2891-03MSD - 1	A653MSD	20819	4.2550	-	7470A	PPB	24 May 2023 16:15:48	U	SMPL
O2897-02 - 1	OILY-WATER-0523	349	0.0892	-	7470A	PPB	24 May 2023 16:18:04	U	SMPL
O2899-11 - 1	TWP01	1269	0.2764	-	7470A	PPB	24 May 2023 16:20:22	U	SMPL
O2894-05LX5 - 1		24	0.0230	-	7470A	PPB	24 May 2023 16:22:39	U	SMPL
O2894-05A - 1		18836	3.8515	-	7470A	PPB	24 May 2023 16:24:56	U	SMPL
O2891-03LX5 - 1		402	0.1000	-	7470A	PPB	24 May 2023 16:27:13	U	SMPL
O2891-03A - 1		20814	4.2540	-	7470A	PPB	24 May 2023 16:29:29	U	SMPL
CCV16 - 1	CCV16	24283	4.9600	-	7470A	PPB	24 May 2023 16:31:45	U	SMPL
CCB16 - 1	CCB16	-73	0.0033	-	7470A	PPB	24 May 2023 16:34:03	U	SMPL
PB152981TB - 1	PB152981TB	132	0.0450	-	7470A	PPB	24 May 2023 16:36:21	U	SMPL
CCV17 - 1	CCV17	24445	4.9930	-	7470A	PPB	24 May 2023 16:38:38	U	SMPL
CCB17 - 1	CCB17	-58	0.0063	-	7470A	PPB	24 May 2023 16:40:54	U	SMPL