



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125630

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/24/2023	13:30
2	O2884-03		1	14	100.00	>212.0	>212.0	05/24/2023	14:00
3	O2884-03DUP		1	13	100.00	>212.0	>212.0	05/24/2023	14:30
4	O2884-04		1	12	100.00	>212.0	>212.0	05/24/2023	15:00
5	O2884-05		1	12	100.00	>212.0	>212.0	05/24/2023	15:30
6	O2884-06		1	14	100.00	>212.0	>212.0	05/24/2023	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26125630

WorkList Name : fp-5-22 WorkList ID : 170182 Department : Wet-Chemistry Date : 05-22-2023 08:25:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2884-03	2002	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B
O2884-04	2003	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B
O2884-05	2004	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B
O2884-06	2005	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B

Date/Time 05/24/2023 12:58
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JLCW1

Date/Time 05/24/2023 16:40
 Raw Sample Received by: JLCW1
 Raw Sample Relinquished by: JLCW1