


Calibration for 061CCBE.d

Batch Folder: D:\Agilent\ICPMH\1\DATA\P8052423 MS.b\
 Analysis File: P8052423 MS.batch.bin
 DA Date-Time: 2023-05-25 00:31:59
 Calibration Title:
 Calibration Method: External Calibration
 VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	004CALB.d	S00	2023-05-24 13:32:57
2	005CAL.S.d	S02	2023-05-24 13:35:59
3	006CAL.S.d	S03	2023-05-24 13:39:04
4	007CAL.S.d	S04	2023-05-24 13:42:03
5	008CAL.S.d	S05	2023-05-24 13:45:04
6	009CAL.S.d	S06	2023-05-24 13:48:03
7	010CAL.S.d	S07	2023-05-24 13:51:05
8	011CAL.S.d	S08	2023-05-24 13:54:04



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	50.82	0.0000	P	11.5
2	<input type="checkbox"/>	1.000	1.073	1688.64	0.0006	P	1.3
3	<input type="checkbox"/>	50.000	53.736	80241.62	0.0289	P	6.3
4	<input type="checkbox"/>	125.000	126.889	195954.50	0.0682	P	2.8
5	<input type="checkbox"/>	250.000	257.756	388329.10	0.1384	P	1.9
6	<input type="checkbox"/>	500.000	507.689	761413.16	0.2727	P	0.9
7	<input type="checkbox"/>	1000.000	993.794	1484270.08	0.5337	A	2.8
8	<input type="checkbox"/>			926.58	0.0003	P	2.7

$y = 5.3706E-004 * x + 1.7757E-005$


R = 0.9999

DL = 0.01139


BEC = 0.03306

Weight: <None>


Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			3380549.29		A	1.0
2	<input type="checkbox"/>			3250932.26		A	0.2
3	<input type="checkbox"/>			3520632.37		A	0.6
4	<input type="checkbox"/>			3753397.44		A	0.8
5	<input type="checkbox"/>			4321460.25		A	4.3
6	<input type="checkbox"/>			3881976.04		A	2.1
7	<input type="checkbox"/>			4032687.94		A	3.6
8	<input type="checkbox"/>			3946071.40		A	4.4



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			39383.89		P	1.6
2	<input type="checkbox"/>			37141.98		P	1.6
3	<input type="checkbox"/>			41481.37		P	0.9
4	<input type="checkbox"/>			46811.17		P	2.1
5	<input type="checkbox"/>			53693.08		P	2.2
6	<input type="checkbox"/>			47730.89		P	1.7
7	<input type="checkbox"/>			51080.99		P	1.2
8	<input type="checkbox"/>			54660.42		P	2.4



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	16309.52	0.1162	P	1.2
2	<input type="checkbox"/>	500.000	502.391	109269.30	0.7767	P	1.2
3	<input type="checkbox"/>	5000.000	5410.221	1005078.43	7.2298	M	1.8
4	<input type="checkbox"/>	12500.000	13241.570	2454995.69	17.5269	A	1.4
5	<input type="checkbox"/>	25000.000	26050.627	4713342.25	34.3689	A	0.6
6	<input type="checkbox"/>	50000.000	53257.909	9495595.78	70.1426	A	0.7
7	<input type="checkbox"/>	100000.00	103940.67	19053183.93	136.7830	A	1.9
8	<input type="checkbox"/>	500000.00	498810.89	89475188.11	655.9801	A	1.7

$y = 0.0013 * x + 0.1162$


R = 1.0000

DL = 3.194

BEC = 88.34

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	466.69	0.0033	P	12.4
2	<input type="checkbox"/>	500.000	504.398	52603.83	0.3740	P	3.1
3	<input type="checkbox"/>	5000.000	5404.949	552570.68	3.9751	P	2.2
4	<input type="checkbox"/>	12500.000	13762.054	1417066.17	10.1161	A	0.8
5	<input type="checkbox"/>	25000.000	26566.850	2677654.27	19.5255	A	0.2
6	<input type="checkbox"/>	50000.000	53061.063	5278845.86	38.9943	A	0.3
7	<input type="checkbox"/>	100000.00	102324.75	10474392.68	75.1948	A	1.8
8	<input type="checkbox"/>	500000.00	499114.99	50026953.89	366.7688	A	2.1

$y = 7.3483E-004 * x + 0.0033$


R = 1.0000

DL = 1.683

BEC = 4.529

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	37.50	0.0003	P	22.9
2	<input type="checkbox"/>	20.000	20.200	827.42	0.0059	P	1.7
3	<input type="checkbox"/>	1000.000	1123.303	43428.47	0.3124	P	2.0
4	<input type="checkbox"/>	2500.000	2536.429	98762.08	0.7051	P	0.6
5	<input type="checkbox"/>	5000.000	5082.788	193724.03	1.4126	P	0.7
6	<input type="checkbox"/>	10000.000	10374.143	390272.09	2.8830	P	0.9
7	<input type="checkbox"/>	20000.000	20069.530	776879.47	5.5771	P	1.8
8	<input type="checkbox"/>	100000.00	99942.396	3788402.05	27.7718	A	0.4

$y = 2.7787E-004 * x + 2.6782E-004$


R = 1.0000

DL = 0.661

BEC = 0.9638

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	981204.03	0.2291	A	2.6
2	<input type="checkbox"/>	10.000	0.365	1001253.88	0.2330	A	2.6
3	<input type="checkbox"/>	50.000	62.810	3699845.44	0.8952	A	7.0
4	<input type="checkbox"/>	125.000	18.845	1882448.81	0.4290	A	6.3
5	<input type="checkbox"/>	250.000	39.003	2773723.22	0.6428	A	10.7
6	<input type="checkbox"/>	500.000	572.494	26246215.17	6.3003	A	2.3
7	<input type="checkbox"/>	1000.000	1029.227	48181222.95	11.1438	A	1.9
8	<input type="checkbox"/>			2687693.79	0.6507	A	10.8

$y = 0.0106 * x + 0.2291$


R = 0.9731

DL = 1.694

BEC = 21.61

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	-18.136	163.89	0.0012	P	4.0
2	<input type="checkbox"/>	0.000	18.136	241.67	0.0017	P	9.9
3	<input type="checkbox"/>	1000.000	1049.178	2416.93	0.0174	P	3.3
4	<input type="checkbox"/>	2500.000	2682.648	5909.68	0.0422	P	3.4
5	<input type="checkbox"/>	5000.000	5290.488	11218.50	0.0818	P	2.6
6	<input type="checkbox"/>	10000.000	10313.299	21399.63	0.1581	P	1.0
7	<input type="checkbox"/>	20000.000	19745.439	41978.07	0.3013	P	1.8
8	<input type="checkbox"/>			211.12	0.0015	P	36.0

$y = 1.5188E-005 * x + 0.0014$


R = 0.9997

DL = 21.41

BEC = 95.02

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	-147.776	468459.87	0.1094	P	1.7
2	<input type="checkbox"/>	0.000	147.776	477448.71	0.1111	P	1.9
3	<input type="checkbox"/>	1000.000	1795.217	498832.51	0.1207	P	6.1
4	<input type="checkbox"/>	2500.000	2050.528	535379.90	0.1221	P	3.1
5	<input type="checkbox"/>	5000.000	4791.234	596115.33	0.1380	P	1.5
6	<input type="checkbox"/>	10000.000	10632.742	715975.32	0.1719	P	2.7
7	<input type="checkbox"/>	20000.000	19752.244	971672.68	0.2247	P	1.2
8	<input type="checkbox"/>			445382.29	0.1078	P	0.4

$y = 5.7958E-006 * x + 0.1103$


R = 0.9979

DL = 1029


BEC = 1.902E+04

Weight: <None>


Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			149007.64		P	0.8
2	<input type="checkbox"/>			146620.23		P	0.8
3	<input type="checkbox"/>			165065.94		P	0.8
4	<input type="checkbox"/>			165041.45		P	1.2
5	<input type="checkbox"/>			170168.75		P	0.9
6	<input type="checkbox"/>			167307.95		P	0.9
7	<input type="checkbox"/>			168699.23		P	1.2
8	<input type="checkbox"/>			176482.22		P	1.6



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			847.27		P	9.1
2	<input type="checkbox"/>			877.83		P	17.6
3	<input type="checkbox"/>			1111.18		P	3.8
4	<input type="checkbox"/>			1094.52		P	10.3
5	<input type="checkbox"/>			1036.18		P	9.3
6	<input type="checkbox"/>			1072.30		P	6.2
7	<input type="checkbox"/>			919.50		P	3.2
8	<input type="checkbox"/>			1086.18		P	9.3



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	29708.99	0.2117	P	3.1
2	<input type="checkbox"/>	500.000	491.935	65415.52	0.4650	P	1.7
3	<input type="checkbox"/>	2500.000	2583.318	214380.09	1.5420	P	1.7
4	<input type="checkbox"/>	6250.000	6286.862	483104.52	3.4491	P	1.6
5	<input type="checkbox"/>	12500.000	12640.832	921701.84	6.7212	P	0.4
6	<input type="checkbox"/>	25000.000	25946.369	1837398.13	13.5729	A	1.4
7	<input type="checkbox"/>	50000.000	49082.773	3550271.26	25.4871	A	2.1
8	<input type="checkbox"/>	250000.00	250080.02	17595061.33	128.9919	A	1.2

$y = 5.1496E-004 * x + 0.2117$


R = 1.0000

DL = 38.73

BEC = 411.1

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	127.07	0.0000	P	19.5
2	<input type="checkbox"/>	500.000	511.387	7060.04	0.0016	P	1.0
3	<input type="checkbox"/>	5000.000	6187.140	80760.97	0.0195	P	7.3
4	<input type="checkbox"/>	12500.000	14244.542	197097.69	0.0450	P	2.4
5	<input type="checkbox"/>	25000.000	28851.500	393146.64	0.0910	P	1.8
6	<input type="checkbox"/>	50000.000	58563.731	769698.15	0.1847	P	1.7
7	<input type="checkbox"/>	100000.00	118848.65	1620821.98	0.3749	A	1.6
8	<input type="checkbox"/>	500000.00	495125.82	6450612.38	1.5617	A	2.2

$y = 3.1540E-006 * x + 2.9703E-005$


R = 0.9992

DL = 5.499

BEC = 9.418

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	17995.66	0.0042	P	0.8
2	<input type="checkbox"/>	500.000	507.969	128520.05	0.0299	P	1.3
3	<input type="checkbox"/>	5000.000	5993.085	1270737.28	0.3074	A	6.8
4	<input type="checkbox"/>	12500.000	13859.898	3093074.73	0.7055	A	2.9
5	<input type="checkbox"/>	25000.000	27914.617	6118820.52	1.4166	A	1.3
6	<input type="checkbox"/>	50000.000	56452.826	11917876.22	2.8606	A	2.0
7	<input type="checkbox"/>	100000.00	109686.54	24012223.93	5.5541	A	2.0
8	<input type="checkbox"/>	500000.00	497227.74	103938056.2	25.1628	A	1.2

$y = 5.0598E-005 * x + 0.0042$


R = 0.9998

DL = 2.062

BEC = 83.05

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	83.34	0.0000	P	25.3
2	<input type="checkbox"/>	5.000	5.123	3414.38	0.0008	P	3.2
3	<input type="checkbox"/>	500.000	545.463	341097.35	0.0825	P	7.1
4	<input type="checkbox"/>	1250.000	1267.324	840574.45	0.1917	P	2.6
5	<input type="checkbox"/>	2500.000	2543.436	1661722.12	0.3848	A	3.0
6	<input type="checkbox"/>	5000.000	5187.653	3269173.01	0.7848	A	2.8
7	<input type="checkbox"/>	10000.000	9890.876	6469484.48	1.4963	A	1.5
8	<input type="checkbox"/>			1152.85	0.0003	P	5.4

$y = 1.5128E-004 * x + 1.9425E-005$


R = 0.9997

DL = 0.09761

BEC = 0.1284

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	4.44	0.0000	P	115.0
2	<input type="checkbox"/>	5.000	4.867	4398.50	0.0313	P	1.9
3	<input type="checkbox"/>	50.000	52.092	46477.54	0.3343	P	1.6
4	<input type="checkbox"/>	125.000	127.308	114437.21	0.8170	P	1.3
5	<input type="checkbox"/>	250.000	255.611	224947.81	1.6403	P	1.3
6	<input type="checkbox"/>	500.000	509.638	442720.01	3.2704	P	1.2
7	<input type="checkbox"/>	1000.000	993.386	887908.98	6.3746	P	2.3
8	<input type="checkbox"/>			334.45	0.0025	P	9.5

$y = 0.0064 * x + 3.1370E-005$


R = 0.9999

DL = 0.01687

BEC = 0.004889

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	523.35	0.0037	P	10.5
2	<input type="checkbox"/>	2.000	2.140	2958.11	0.0210	P	3.8
3	<input type="checkbox"/>	50.000	52.751	59772.05	0.4300	P	2.7
4	<input type="checkbox"/>	125.000	127.654	145016.36	1.0353	P	1.0
5	<input type="checkbox"/>	250.000	256.121	284346.10	2.0734	P	0.2
6	<input type="checkbox"/>	500.000	510.312	558752.25	4.1276	P	1.1
7	<input type="checkbox"/>	1000.000	992.844	1118202.98	8.0269	P	1.4
8	<input type="checkbox"/>			8871.72	0.0650	P	1.5

$y = 0.0081 * x + 0.0037$


R = 0.9999

DL = 0.146

BEC = 0.4619

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	585.21	0.0042	P	16.5
2	<input type="checkbox"/>	50.000	52.999	3376.38	0.0240	P	2.9
3	<input type="checkbox"/>	2500.000	2711.557	141579.25	1.0185	P	2.8
4	<input type="checkbox"/>	6250.000	6644.266	348725.91	2.4897	P	0.7
5	<input type="checkbox"/>	12500.000	13283.782	681995.03	4.9734	P	1.5
6	<input type="checkbox"/>	25000.000	26860.181	1360755.55	10.0520	A	0.6
7	<input type="checkbox"/>	50000.000	52453.074	2734426.54	19.6258	A	0.6
8	<input type="checkbox"/>	250000.00	249272.20	12720629.56	93.2519	A	0.6

$y = 3.7408E-004 * x + 0.0042$


R = 0.9999

DL = 5.529

BEC = 11.14

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	78.89	0.0006	P	36.6
2	<input type="checkbox"/>	1.000	0.960	693.36	0.0049	P	7.4
3	<input type="checkbox"/>	500.000	516.233	326170.69	2.3465	P	2.7
4	<input type="checkbox"/>	1250.000	1273.320	810585.52	5.7870	P	0.8
5	<input type="checkbox"/>	2500.000	2612.504	1628144.96	11.8728	A	0.6
6	<input type="checkbox"/>	5000.000	5154.727	3171205.92	23.4256	A	0.6
7	<input type="checkbox"/>	10000.000	9890.784	6262364.77	44.9480	A	0.4
8	<input type="checkbox"/>			3217.06	0.0236	P	1.5

$y = 0.0045 * x + 5.6456E-004$


R = 0.9997

DL = 0.1365

BEC = 0.1242

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	3648.69	0.0260	P	2.8
2	<input type="checkbox"/>	50.000	50.790	52303.38	0.3718	P	1.0
3	<input type="checkbox"/>	2500.000	2756.886	2612811.17	18.7950	A	2.2
4	<input type="checkbox"/>	6250.000	6648.197	6343421.26	45.2873	A	0.5
5	<input type="checkbox"/>	12500.000	13306.737	12426556.99	90.6189	A	1.1
6	<input type="checkbox"/>	25000.000	26150.852	24105876.47	178.0624	A	1.1
7	<input type="checkbox"/>	50000.000	51170.857	48542343.68	348.3999	A	0.3
8	<input type="checkbox"/>	250000.00	249597.88	231805371.8	1,699.301	A	1.0

$y = 0.0068 * x + 0.0260$


R = 1.0000

DL = 0.3182

BEC = 3.816

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	181.49	0.0013	P	19.5
2	<input type="checkbox"/>	50.000	44.789	1274.15	0.0091	P	4.9
3	<input type="checkbox"/>	2500.000	2621.934	63362.80	0.4558	P	2.4
4	<input type="checkbox"/>	6250.000	6446.882	156713.69	1.1188	P	1.0
5	<input type="checkbox"/>	12500.000	12891.468	306638.65	2.2360	P	0.2
6	<input type="checkbox"/>	25000.000	25702.315	603305.73	4.4567	P	0.7
7	<input type="checkbox"/>	50000.000	49490.956	1195448.14	8.5804	P	0.6
8	<input type="checkbox"/>	250000.00	250005.86	5911260.32	43.3388	A	1.8

$y = 1.7335E-004 * x + 0.0013$


R = 1.0000

DL = 4.361

BEC = 7.467

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	57.78	0.0004	P	7.8
2	<input type="checkbox"/>	1.000	0.961	2041.28	0.0145	P	3.9
3	<input type="checkbox"/>	50.000	51.288	104631.18	0.7526	P	1.9
4	<input type="checkbox"/>	125.000	124.215	255254.40	1.8223	P	0.5
5	<input type="checkbox"/>	250.000	249.707	502326.42	3.6628	P	1.6
6	<input type="checkbox"/>	500.000	496.797	986434.46	7.2869	P	0.7
7	<input type="checkbox"/>	1000.000	1001.708	2047107.95	14.6924	A	1.2
8	<input type="checkbox"/>			1802.35	0.0132	P	3.8

$y = 0.0147 * x + 4.1134E-004$


R = 1.0000

DL = 0.006603

BEC = 0.02805

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	392.23	0.0028	P	3.3
2	<input type="checkbox"/>	1.000	0.980	918.93	0.0065	P	4.4
3	<input type="checkbox"/>	50.000	53.819	28939.51	0.2082	P	1.7
4	<input type="checkbox"/>	125.000	131.305	70578.40	0.5039	P	0.7
5	<input type="checkbox"/>	250.000	260.258	136586.99	0.9960	P	0.3
6	<input type="checkbox"/>	500.000	517.459	267695.93	1.9775	P	0.8
7	<input type="checkbox"/>	1000.000	987.727	525506.70	3.7721	P	0.9
8	<input type="checkbox"/>			1904.59	0.0140	P	6.6

$y = 0.0038 * x + 0.0028$


R = 0.9997

DL = 0.07139

BEC = 0.7318

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	895.59	0.0064	P	9.0
2	<input type="checkbox"/>	2.000	2.076	4229.55	0.0301	P	5.9
3	<input type="checkbox"/>	500.000	523.035	830415.30	5.9738	P	2.2
4	<input type="checkbox"/>	1250.000	1315.210	2102774.83	15.0119	A	0.2
5	<input type="checkbox"/>	2500.000	2613.184	4089410.29	29.8208	A	2.4
6	<input type="checkbox"/>	5000.000	5104.644	7884756.48	58.2465	A	1.4
7	<input type="checkbox"/>	10000.000	9910.079	15751036.99	113.0729	A	2.1
8	<input type="checkbox"/>			3841.67	0.0282	P	1.6

$y = 0.0114 * x + 0.0064$


R = 0.9998

DL = 0.1517

BEC = 0.5596

Weight: <None>


Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	79.77	0.0006	P	14.8
2	<input type="checkbox"/>	2.000	2.463	560.74	0.0040	P	4.0
3	<input type="checkbox"/>	500.000	533.776	103095.78	0.7416	P	1.6
4	<input type="checkbox"/>	1250.000	1294.478	251800.62	1.7977	P	0.8
5	<input type="checkbox"/>	2500.000	2603.928	495849.71	3.6155	P	1.5
6	<input type="checkbox"/>	5000.000	5111.215	960640.03	7.0963	P	0.7
7	<input type="checkbox"/>	10000.000	9911.162	1917019.41	13.7600	A	0.8
8	<input type="checkbox"/>			10392.25	0.0762	P	4.8

$y = 0.0014 * x + 5.6703E-004$
 R = 0.9998
 DL = 0.1809
 BEC = 0.4084


Weight: <None>
 Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	4.00	0.0000	P	90.5
2	<input type="checkbox"/>	1.000	1.153	128.68	0.0001	P	3.6
3	<input type="checkbox"/>	50.000	53.194	5740.75	0.0046	P	1.0
4	<input type="checkbox"/>	125.000	129.281	14078.42	0.0112	P	0.8
5	<input type="checkbox"/>	250.000	259.740	27725.18	0.0225	P	1.2
6	<input type="checkbox"/>	500.000	512.528	54188.57	0.0444	P	0.5
7	<input type="checkbox"/>	1000.000	990.606	107272.04	0.0858	P	1.1
8	<input type="checkbox"/>			25.67	0.0000	P	13.9

$y = 8.6641E-005 * x + 3.2012E-006$
 R = 0.9998
 DL = 0.1003
 BEC = 0.03695

Weight: <None>
 Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	1.11	0.0000	P	173.2
2	<input type="checkbox"/>	5.000	5.345	15.56	0.0000	P	11.2
3	<input type="checkbox"/>	50.000	52.316	142.23	0.0001	P	12.8
4	<input type="checkbox"/>	125.000	129.834	354.45	0.0003	P	7.7
5	<input type="checkbox"/>	250.000	272.392	727.80	0.0006	P	8.0
6	<input type="checkbox"/>	500.000	509.546	1347.85	0.0011	P	10.3
7	<input type="checkbox"/>	1000.000	988.907	2676.95	0.0021	P	5.2
8	<input type="checkbox"/>			2.22	0.0000	P	86.6

$y = 2.1656E-006 * x + 8.8095E-007$


R = 0.9996

DL = 2.114

BEC = 0.4068

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	33.33	0.0000	P	2.1
2	<input type="checkbox"/>	5.000	6.554	92.01	0.0001	P	16.3
3	<input type="checkbox"/>	50.000	56.797	541.73	0.0004	P	6.3
4	<input type="checkbox"/>	125.000	134.200	1246.52	0.0010	P	2.3
5	<input type="checkbox"/>	250.000	264.448	2376.45	0.0019	P	1.1
6	<input type="checkbox"/>	500.000	522.267	4617.23	0.0038	P	0.8
7	<input type="checkbox"/>	1000.000	983.757	8877.70	0.0071	P	2.3
8	<input type="checkbox"/>			45.34	0.0000	P	23.2

$y = 7.1943E-006 * x + 2.6652E-005$


R = 0.9995

DL = 0.2368


BEC = 3.705

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			38312.00		P	2.0
2	<input type="checkbox"/>			39682.21		P	1.5
3	<input type="checkbox"/>			39456.77		P	1.0
4	<input type="checkbox"/>			39746.24		P	1.3
5	<input type="checkbox"/>			39038.73		P	0.4
6	<input type="checkbox"/>			37857.79		P	1.7
7	<input type="checkbox"/>			38671.14		P	3.5
8	<input type="checkbox"/>			37994.55		P	2.5



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	-4.58	0.0000	P	-1026.
2	<input type="checkbox"/>	5.000	5.711	460.79	0.0000	P	5.4
3	<input type="checkbox"/>	50.000	55.702	4347.71	0.0004	P	6.4
4	<input type="checkbox"/>	125.000	128.670	10730.85	0.0010	P	3.0
5	<input type="checkbox"/>	250.000	252.331	20946.64	0.0019	P	3.0
6	<input type="checkbox"/>	500.000	517.015	41148.74	0.0039	P	3.1
7	<input type="checkbox"/>	1000.000	990.163	80666.14	0.0075	P	1.0
8	<input type="checkbox"/>			55.00	0.0000	P	22.0

$y = 7.5513E-006 * x - 3.8289E-007$


R = 0.9998

DL = 1.562


BEC = -0.05071

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			231.64		P	14.0
2	<input type="checkbox"/>			219.56		P	4.6
3	<input type="checkbox"/>			212.48		P	17.4
4	<input type="checkbox"/>			223.31		P	6.5
5	<input type="checkbox"/>			230.39		P	6.5
6	<input type="checkbox"/>			231.64		P	12.5
7	<input type="checkbox"/>			235.81		P	7.8
8	<input type="checkbox"/>			247.05		P	7.7



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	363.91	0.0000	P	29.2
2	<input type="checkbox"/>	1.000	1.075	1752.94	0.0002	P	8.6
3	<input type="checkbox"/>	625.000	1353.680	1675705.66	0.1620	A	6.7
4	<input type="checkbox"/>	3125.000	3095.649	4091968.80	0.3704	A	3.4
5	<input type="checkbox"/>	6250.000	6190.545	8144944.63	0.7406	A	1.5
6	<input type="checkbox"/>	12500.000	12696.919	16011181.78	1.5190	A	3.2
7	<input type="checkbox"/>	25000.000	24901.856	32141812.52	2.9792	A	1.6
8	<input type="checkbox"/>			9275.55	0.0009	P	4.7

$y = 1.1963E-004 * x + 3.3982E-005$


R = 0.9995

DL = 0.2486

BEC = 0.284

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	172.23	0.0000	P	13.2
2	<input type="checkbox"/>	1.000	0.995	11299.34	0.0010	P	0.6
3	<input type="checkbox"/>	625.000	1320.966	14179639.72	1.3706	A	6.4
4	<input type="checkbox"/>	3125.000	3077.106	35275578.22	3.1927	A	3.2
5	<input type="checkbox"/>	6250.000	6246.977	71277743.96	6.4817	A	1.6
6	<input type="checkbox"/>	12500.000	12634.405	138194012.8	13.1092	A	2.7
7	<input type="checkbox"/>	25000.000	24922.141	278974984.0	25.8586	A	0.4
8	<input type="checkbox"/>			77115.69	0.0073	P	2.4

$y = 0.0010 * x + 1.6040E-005$


R = 0.9996

DL = 0.006108

BEC = 0.01546

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	422.24	0.0000	P	17.1
2	<input type="checkbox"/>	1.000	0.924	7115.89	0.0007	P	4.3
3	<input type="checkbox"/>	50.000	52.601	366186.71	0.0354	P	5.5
4	<input type="checkbox"/>	125.000	118.630	881313.69	0.0798	P	1.8
5	<input type="checkbox"/>	250.000	244.803	1809310.24	0.1645	A	2.3
6	<input type="checkbox"/>	500.000	506.382	3587846.30	0.3403	A	2.1
7	<input type="checkbox"/>	1000.000	998.775	7240573.42	0.6712	A	2.3
8	<input type="checkbox"/>			1789.06	0.0002	P	4.5

$y = 6.7195E-004 * x + 3.9420E-005$


R = 0.9999

DL = 0.0301

BEC = 0.05867

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	69.45	0.0000	P	55.2
2	<input type="checkbox"/>	1.000	0.905	1536.24	0.0001	P	6.6
3	<input type="checkbox"/>	50.000	52.008	80901.79	0.0078	P	4.6
4	<input type="checkbox"/>	125.000	117.171	194517.05	0.0176	P	1.5
5	<input type="checkbox"/>	250.000	238.050	393136.21	0.0358	P	2.2
6	<input type="checkbox"/>	500.000	500.385	792187.41	0.0751	P	2.6
7	<input type="checkbox"/>	1000.000	1003.673	1626186.81	0.1507	A	1.4
8	<input type="checkbox"/>			380.57	0.0000	P	9.4

$y = 1.5016E-004 * x + 6.5135E-006$


R = 0.9999

DL = 0.07184

BEC = 0.04338

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	136.12	0.0000	P	35.0
2	<input type="checkbox"/>	5.000	5.663	8427.82	0.0008	P	2.8
3	<input type="checkbox"/>	500.000	530.726	746077.80	0.0721	P	4.4
4	<input type="checkbox"/>	1250.000	1235.340	1853549.76	0.1678	A	2.7
5	<input type="checkbox"/>	2500.000	2433.751	3634221.13	0.3305	A	1.0
6	<input type="checkbox"/>	5000.000	4985.401	7136045.21	0.6770	A	2.9
7	<input type="checkbox"/>	10000.000	10024.157	14684750.16	1.3612	A	1.5
8	<input type="checkbox"/>			3389.40	0.0003	P	7.7

$y = 1.3579E-004 * x + 1.2740E-005$


R = 1.0000

DL = 0.09865

BEC = 0.09383

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	38.89	0.0000	P	32.6
2	<input type="checkbox"/>	5.000	4.645	9912.12	0.0009	P	1.0
3	<input type="checkbox"/>	500.000	532.071	1085728.73	0.1049	P	4.3
4	<input type="checkbox"/>	1250.000	1251.819	2726333.11	0.2468	A	3.1
5	<input type="checkbox"/>	2500.000	2484.669	5385984.49	0.4898	A	1.5
6	<input type="checkbox"/>	5000.000	5046.417	10487047.46	0.9948	A	2.5
7	<input type="checkbox"/>	10000.000	9978.794	21220979.12	1.9670	A	1.5
8	<input type="checkbox"/>			4687.09	0.0004	P	9.7

$y = 1.9712E-004 * x + 3.6232E-006$


R = 1.0000

DL = 0.01797

BEC = 0.01838

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	66.67	0.0000	P	23.4
2	<input type="checkbox"/>	5.000	4.768	11229.87	0.0010	P	2.9
3	<input type="checkbox"/>	500.000	525.750	1181618.63	0.1142	P	5.3
4	<input type="checkbox"/>	1250.000	1237.234	2968582.18	0.2687	A	3.5
5	<input type="checkbox"/>	2500.000	2473.064	5905649.53	0.5371	A	2.1
6	<input type="checkbox"/>	5000.000	5013.316	11480144.44	1.0887	A	1.6
7	<input type="checkbox"/>	10000.000	10000.385	23431073.44	2.1718	A	1.1
8	<input type="checkbox"/>			7457.78	0.0007	P	7.4

$y = 2.1717E-004 * x + 6.2023E-006$


R = 1.0000

DL = 0.02006

BEC = 0.02856

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	36.11	0.0000	P	25.9
2	<input type="checkbox"/>	5.000	4.678	6276.60	0.0006	P	3.8
3	<input type="checkbox"/>	500.000	525.275	672372.50	0.0650	P	6.3
4	<input type="checkbox"/>	1250.000	1224.699	1674166.11	0.1515	A	2.8
5	<input type="checkbox"/>	2500.000	2467.245	3356527.55	0.3053	A	2.1
6	<input type="checkbox"/>	5000.000	4969.012	6480657.05	0.6148	A	2.8
7	<input type="checkbox"/>	10000.000	10025.582	13382345.26	1.2404	A	0.8
8	<input type="checkbox"/>			2778.16	0.0003	P	11.9

$y = 1.2372E-004 * x + 3.3623E-006$


R = 1.0000

DL = 0.02109

BEC = 0.02718

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	47.22	0.0000	P	45.6
2	<input type="checkbox"/>	5.000	4.699	16254.44	0.0015	P	1.0
3	<input type="checkbox"/>	500.000	524.133	1734914.02	0.1676	A	5.2
4	<input type="checkbox"/>	1250.000	1212.651	4285144.85	0.3879	A	3.2
5	<input type="checkbox"/>	2500.000	2431.588	8552280.32	0.7777	A	1.0
6	<input type="checkbox"/>	5000.000	5038.495	16988729.55	1.6115	A	3.0
7	<input type="checkbox"/>	10000.000	10001.317	34511083.41	3.1988	A	0.7
8	<input type="checkbox"/>			7193.84	0.0007	P	16.2

$y = 3.1984E-004 * x + 4.4262E-006$


R = 1.0000

DL = 0.01892

BEC = 0.01384

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	6173.77	0.0009	P	3.1
2	<input type="checkbox"/>	1.000	0.371	6115.43	0.0009	P	6.3
3	<input type="checkbox"/>	50.000	51.732	11844.28	0.0017	P	2.0
4	<input type="checkbox"/>	125.000	123.781	21687.04	0.0029	P	3.8
5	<input type="checkbox"/>	250.000	255.761	37198.27	0.0050	P	1.2
6	<input type="checkbox"/>	500.000	499.012	66159.40	0.0089	P	5.1
7	<input type="checkbox"/>	1000.000	999.120	126704.05	0.0170	P	2.9
8	<input type="checkbox"/>			5898.67	0.0008	P	4.5

$y = 1.6171E-005 * x + 8.5604E-004$


R = 1.0000

DL = 4.928

BEC = 52.94

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	77.78	0.0000	P	73.4
2	<input type="checkbox"/>	1.000	1.040	5081.64	0.0007	P	1.7
3	<input type="checkbox"/>	50.000	55.766	264133.00	0.0378	P	4.0
4	<input type="checkbox"/>	125.000	125.452	644974.16	0.0850	P	1.8
5	<input type="checkbox"/>	250.000	251.040	1267105.93	0.1700	P	1.1
6	<input type="checkbox"/>	500.000	514.687	2584862.84	0.3486	A	3.3
7	<input type="checkbox"/>	1000.000	992.052	5005010.35	0.6720	A	0.4
8	<input type="checkbox"/>			2397.49	0.0003	P	6.4

$y = 6.7733E-004 * x + 1.0888E-005$


R = 0.9998

DL = 0.03541

BEC = 0.01607

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	19.44	0.0000	P	62.1
2	<input type="checkbox"/>	1.000	1.004	100.00	0.0000	P	6.5
3	<input type="checkbox"/>	50.000	56.094	4468.83	0.0006	P	5.4
4	<input type="checkbox"/>	125.000	128.116	11054.53	0.0015	P	2.9
5	<input type="checkbox"/>	250.000	256.437	21706.13	0.0029	P	1.9
6	<input type="checkbox"/>	500.000	501.389	42216.00	0.0057	P	3.0
7	<input type="checkbox"/>	1000.000	997.002	84300.37	0.0113	P	0.1
8	<input type="checkbox"/>			86.11	0.0000	P	15.6

$y = 1.1350E-005 * x + 2.6999E-006$


R = 1.0000

DL = 0.4429

BEC = 0.2379

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	16.67	0.0000	P	86.6
2	<input type="checkbox"/>	1.000	1.068	4898.22	0.0007	P	2.9
3	<input type="checkbox"/>	50.000	55.869	251351.32	0.0360	P	4.6
4	<input type="checkbox"/>	125.000	125.422	612703.92	0.0807	P	1.4
5	<input type="checkbox"/>	250.000	255.532	1225548.16	0.1645	P	1.0
6	<input type="checkbox"/>	500.000	505.906	2415297.49	0.3256	A	2.0
7	<input type="checkbox"/>	1000.000	995.318	4770866.21	0.6406	A	1.0
8	<input type="checkbox"/>			2294.70	0.0003	P	7.2

$y = 6.4363E-004 * x + 2.2866E-006$


R = 0.9999

DL = 0.009233

BEC = 0.003553

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	1223.92	0.0002	P	2.4
2	<input type="checkbox"/>	1.000	0.976	2098.97	0.0003	P	2.2
3	<input type="checkbox"/>	50.000	56.188	51857.59	0.0074	P	5.2
4	<input type="checkbox"/>	125.000	126.867	125503.89	0.0165	P	3.5
5	<input type="checkbox"/>	250.000	254.730	246194.88	0.0330	P	1.1
6	<input type="checkbox"/>	500.000	505.596	485138.60	0.0654	P	1.8
7	<input type="checkbox"/>	1000.000	995.477	957888.93	0.1286	P	1.0
8	<input type="checkbox"/>			1487.52	0.0002	P	6.4

$y = 1.2904E-004 * x + 1.6971E-004$


R = 1.0000

DL = 0.09485

BEC = 1.315

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	544.47	0.0001	P	12.5
2	<input type="checkbox"/>	5.000	5.050	16271.26	0.0023	P	3.8
3	<input type="checkbox"/>	50.000	54.005	166122.50	0.0238	P	4.7
4	<input type="checkbox"/>	125.000	123.950	413223.08	0.0545	P	2.5
5	<input type="checkbox"/>	250.000	250.005	817882.54	0.1098	P	0.1
6	<input type="checkbox"/>	500.000	500.859	1630329.40	0.2198	A	1.6
7	<input type="checkbox"/>	1000.000	999.500	3266790.94	0.4386	A	0.6
8	<input type="checkbox"/>			1939.08	0.0003	P	5.3

$y = 4.3870E-004 * x + 7.5417E-005$


R = 1.0000

DL = 0.06458

BEC = 0.1719

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	36.11	0.0000	P	69.3
2	<input type="checkbox"/>	2.000	2.066	7068.67	0.0010	P	5.0
3	<input type="checkbox"/>	50.000	53.481	179376.88	0.0257	P	4.6
4	<input type="checkbox"/>	125.000	121.985	444198.61	0.0585	P	2.3
5	<input type="checkbox"/>	250.000	249.766	892891.37	0.1198	P	1.9
6	<input type="checkbox"/>	500.000	509.366	1812203.60	0.2444	A	2.6
7	<input type="checkbox"/>	1000.000	995.578	3557003.99	0.4777	A	1.6
8	<input type="checkbox"/>			1855.74	0.0003	P	7.0

$y = 4.7977E-004 * x + 4.9575E-006$


R = 0.9999

DL = 0.02148

BEC = 0.01033

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	11.11	0.0000	P	112.8
2	<input type="checkbox"/>	10.000	9.524	8619.65	0.0012	P	5.6
3	<input type="checkbox"/>	250.000	263.071	234205.07	0.0335	P	5.5
4	<input type="checkbox"/>	625.000	598.391	578639.59	0.0762	P	1.5
5	<input type="checkbox"/>	1250.000	1224.331	1162247.47	0.1560	P	1.5
6	<input type="checkbox"/>	2500.000	2506.148	2368136.18	0.3193	A	1.5
7	<input type="checkbox"/>	5000.000	5006.017	4750155.23	0.6378	A	2.9
8	<input type="checkbox"/>			2933.74	0.0004	P	7.9

$y = 1.2740E-004 * x + 1.5220E-006$


R = 1.0000

DL = 0.04044

BEC = 0.01195

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	19.45	0.0000	P	22.4
2	<input type="checkbox"/>	10.000	9.718	14872.44	0.0021	P	2.5
3	<input type="checkbox"/>	250.000	271.181	408041.58	0.0584	P	5.3
4	<input type="checkbox"/>	625.000	621.454	1015481.57	0.1338	P	2.1
5	<input type="checkbox"/>	1250.000	1261.813	2024532.68	0.2717	A	2.2
6	<input type="checkbox"/>	2500.000	2519.002	4022188.68	0.5424	A	2.4
7	<input type="checkbox"/>	5000.000	4986.930	7997103.91	1.0737	A	1.6
8	<input type="checkbox"/>			5087.20	0.0007	P	0.6

$y = 2.1531E-004 * x + 2.6890E-006$


R = 1.0000

DL = 0.008394


BEC = 0.01249

Weight: <None>


Min Conc: 0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			11.11		P	69.3
2	<input type="checkbox"/>			11.11		P	69.3
3	<input type="checkbox"/>			42.22		P	24.1
4	<input type="checkbox"/>			124.45		P	20.3
5	<input type="checkbox"/>			217.78		P	22.1
6	<input type="checkbox"/>			502.24		P	14.1
7	<input type="checkbox"/>			995.62		P	10.3
8	<input type="checkbox"/>			28.89		P	48.0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			1.11		P	173.2
2	<input type="checkbox"/>			2.22		P	173.2
3	<input type="checkbox"/>			4.44		P	43.4
4	<input type="checkbox"/>			4.44		P	114.6
5	<input type="checkbox"/>			15.56		P	65.5
6	<input type="checkbox"/>			37.78		P	5.1
7	<input type="checkbox"/>			70.00		P	26.5
8	<input type="checkbox"/>			13.33		P	75.0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			11.11		P	86.6
2	<input type="checkbox"/>			8.33		P	0.0
3	<input type="checkbox"/>			36.11		P	48.0
4	<input type="checkbox"/>			55.56		P	17.3
5	<input type="checkbox"/>			80.56		P	21.5
6	<input type="checkbox"/>			219.45		P	17.9
7	<input type="checkbox"/>			322.24		P	20.1
8	<input type="checkbox"/>			716.71		P	9.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			0.00		P	
2	<input type="checkbox"/>			4.44		P	114.6
3	<input type="checkbox"/>			13.33		P	66.1
4	<input type="checkbox"/>			24.44		P	20.8
5	<input type="checkbox"/>			48.89		P	10.4
6	<input type="checkbox"/>			76.67		P	8.7
7	<input type="checkbox"/>			140.00		P	9.5
8	<input type="checkbox"/>			463.35		P	5.8




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			66.67		P	37.5
2	<input type="checkbox"/>			80.56		P	26.0
3	<input type="checkbox"/>			94.45		P	10.2
4	<input type="checkbox"/>			113.89		P	22.4
5	<input type="checkbox"/>			127.78		P	13.6
6	<input type="checkbox"/>			163.89		P	25.1
7	<input type="checkbox"/>			294.46		P	8.2
8	<input type="checkbox"/>			1075.08		P	8.6




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			103.33		P	11.6
2	<input type="checkbox"/>			106.67		P	19.5
3	<input type="checkbox"/>			120.00		P	19.4
4	<input type="checkbox"/>			117.78		P	11.4
5	<input type="checkbox"/>			130.00		P	24.5
6	<input type="checkbox"/>			183.34		P	7.9
7	<input type="checkbox"/>			236.67		P	24.9
8	<input type="checkbox"/>			574.46		P	6.4




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			44.45		P	57.3
2	<input type="checkbox"/>			52.78		P	39.7
3	<input type="checkbox"/>			77.78		P	16.4
4	<input type="checkbox"/>			52.78		P	50.8
5	<input type="checkbox"/>			50.00		P	33.3
6	<input type="checkbox"/>			69.45		P	18.3
7	<input type="checkbox"/>			105.56		P	35.6
8	<input type="checkbox"/>			66.67		P	12.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			36.67		P	15.7
2	<input type="checkbox"/>			20.00		P	44.1
3	<input type="checkbox"/>			21.11		P	55.4
4	<input type="checkbox"/>			25.56		P	64.3
5	<input type="checkbox"/>			30.00		P	22.2
6	<input type="checkbox"/>			32.22		P	15.8
7	<input type="checkbox"/>			52.22		P	9.7
8	<input type="checkbox"/>			42.22		P	48.2




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			50.00		P	33.3
2	<input type="checkbox"/>			41.67		P	72.1
3	<input type="checkbox"/>			61.11		P	41.7
4	<input type="checkbox"/>			55.56		P	31.2
5	<input type="checkbox"/>			41.67		P	20.0
6	<input type="checkbox"/>			100.00		P	14.4
7	<input type="checkbox"/>			177.78		P	19.5
8	<input type="checkbox"/>			350.02		P	9.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			33.34		P	57.7
2	<input type="checkbox"/>			20.37		P	63.0
3	<input type="checkbox"/>			33.34		P	28.9
4	<input type="checkbox"/>			27.78		P	34.6
5	<input type="checkbox"/>			35.18		P	39.7
6	<input type="checkbox"/>			66.67		P	0.0
7	<input type="checkbox"/>			79.63		P	17.6
8	<input type="checkbox"/>			177.78		P	20.5



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			1727.94		P	3.2
2	<input type="checkbox"/>			1739.06		P	5.6
3	<input type="checkbox"/>			2564.21		P	7.3
4	<input type="checkbox"/>			3806.22		P	2.8
5	<input type="checkbox"/>			5565.20		P	3.7
6	<input type="checkbox"/>			9712.19		P	7.3
7	<input type="checkbox"/>			18085.04		P	0.7
8	<input type="checkbox"/>			1783.50		P	6.6



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>			7869.17		P	2.2
2	<input type="checkbox"/>			7770.96		P	1.1
3	<input type="checkbox"/>			7828.41		P	1.0
4	<input type="checkbox"/>			8721.51		P	0.7
5	<input type="checkbox"/>			9383.11		P	1.7
6	<input type="checkbox"/>			11855.54		P	2.0
7	<input type="checkbox"/>			16304.80		P	2.3
8	<input type="checkbox"/>			7984.02		P	2.3



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	97.23	0.0000	P	16.3
2	<input type="checkbox"/>	1.000	0.918	3695.07	0.0008	P	5.8
3	<input type="checkbox"/>	50.000	49.802	187399.29	0.0427	P	6.2
4	<input type="checkbox"/>	125.000	113.649	469641.10	0.0973	P	1.7
5	<input type="checkbox"/>	250.000	234.855	950454.46	0.2011	P	3.9
6	<input type="checkbox"/>	500.000	497.173	1969668.08	0.4257	A	3.6
7	<input type="checkbox"/>	1000.000	1006.629	3958608.16	0.8620	A	1.2
8	<input type="checkbox"/>			847.27	0.0002	P	5.0

$y = 8.5629E-004 * x + 2.0765E-005$


R = 0.9998

DL = 0.01185

BEC = 0.02425

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	200.01	0.0000	P	14.5
2	<input type="checkbox"/>	1.000	0.943	8967.24	0.0020	P	1.8
3	<input type="checkbox"/>	50.000	52.018	463930.67	0.1056	P	6.2
4	<input type="checkbox"/>	125.000	115.754	1133870.22	0.2350	P	2.6
5	<input type="checkbox"/>	250.000	244.035	2341513.98	0.4955	A	4.3
6	<input type="checkbox"/>	500.000	501.772	4714134.96	1.0187	A	2.7
7	<input type="checkbox"/>	1000.000	1001.660	9338132.85	2.0335	A	1.2
8	<input type="checkbox"/>			1855.75	0.0004	P	3.2

$y = 0.0020 * x + 4.2786E-005$


R = 0.9999

DL = 0.009156

BEC = 0.02108

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	440.76	0.0001	P	7.8
2	<input type="checkbox"/>	1.000	0.882	3330.11	0.0007	P	6.5
3	<input type="checkbox"/>	250.000	247.759	781580.28	0.1780	P	6.7
4	<input type="checkbox"/>	625.000	589.142	2041466.71	0.4232	A	2.8
5	<input type="checkbox"/>	1250.000	1223.170	4152142.46	0.8785	A	3.2
6	<input type="checkbox"/>	2500.000	2493.507	8287374.26	1.7908	A	2.3
7	<input type="checkbox"/>	5000.000	5014.548	16539635.67	3.6012	A	1.3
8	<input type="checkbox"/>			3537.57	0.0008	P	3.6

$y = 7.1813E-004 * x + 9.4225E-005$


R = 1.0000

DL = 0.0306

BEC = 0.1312

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	359.27	0.0001	P	17.4
2	<input type="checkbox"/>	1.000	0.905	2928.17	0.0006	P	2.2
3	<input type="checkbox"/>	250.000	252.886	690869.29	0.1573	P	5.7
4	<input type="checkbox"/>	625.000	593.185	1779731.17	0.3688	A	1.9
5	<input type="checkbox"/>	1250.000	1229.024	3611536.45	0.7641	A	3.2
6	<input type="checkbox"/>	2500.000	2505.245	7207498.37	1.5575	A	2.3
7	<input type="checkbox"/>	5000.000	5006.454	14293296.50	3.1125	A	1.5
8	<input type="checkbox"/>			3157.85	0.0008	P	3.6

$y = 6.2167E-004 * x + 7.6710E-005$


R = 1.0000

DL = 0.06456

BEC = 0.1234

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	1650.07	0.0004	P	3.5
2	<input type="checkbox"/>	1.000	0.912	13554.91	0.0030	P	2.5
3	<input type="checkbox"/>	250.000	256.517	3221839.67	0.7336	A	6.0
4	<input type="checkbox"/>	625.000	596.313	8225324.20	1.7049	A	2.2
5	<input type="checkbox"/>	1250.000	1219.880	16482626.84	3.4873	A	2.7
6	<input type="checkbox"/>	2500.000	2488.839	32924385.87	7.1144	A	2.0
7	<input type="checkbox"/>	5000.000	5016.371	65855030.65	14.3391	A	1.1
8	<input type="checkbox"/>			14162.50	0.0034	P	3.4

$y = 0.0029 * x + 3.5275E-004$


R = 1.0000

DL = 0.01314

BEC = 0.1234

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	0.000	0.000	41.67	0.0000	P	53.5
2	<input type="checkbox"/>	1.000	0.920	12150.63	0.0027	P	3.3
3	<input type="checkbox"/>	50.000	49.762	627765.32	0.1430	P	7.9
4	<input type="checkbox"/>	125.000	113.016	1567291.11	0.3249	P	2.5
5	<input type="checkbox"/>	250.000	237.930	3232091.26	0.6839	A	3.5
6	<input type="checkbox"/>	500.000	492.312	6548326.92	1.4152	A	2.6
7	<input type="checkbox"/>	1000.000	1008.371	13312122.43	2.8986	A	1.1
8	<input type="checkbox"/>			1269.58	0.0003	P	18.9

$y = 0.0029 * x + 8.9621E-006$


R = 0.9998

DL = 0.005007


BEC = 0.003118

Weight: <None>


Min Conc: 0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		2863350.39		A	1.2
2	<input type="checkbox"/>	1.000		2842292.19		A	2.5
3	<input type="checkbox"/>	1.000		2786003.71		A	6.4
4	<input type="checkbox"/>	1.000		2876465.19		A	3.3
5	<input type="checkbox"/>	1.000		2805514.29		A	1.8
6	<input type="checkbox"/>	1.000		2792660.22		A	2.1
7	<input type="checkbox"/>	1.000		2782237.22		A	2.7
8	<input type="checkbox"/>	1.000		2671886.07		A	0.3




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		4282827.65		A	1.1
2	<input type="checkbox"/>	1.000		4298219.14		A	2.0
3	<input type="checkbox"/>	1.000		4145193.93		A	6.4
4	<input type="checkbox"/>	1.000		4386118.48		A	2.9
5	<input type="checkbox"/>	1.000		4319827.85		A	1.8
6	<input type="checkbox"/>	1.000		4167593.34		A	2.7
7	<input type="checkbox"/>	1.000		4324127.16		A	1.5
8	<input type="checkbox"/>	1.000		4130663.41		A	0.2




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		140403.93		P	2.1
2	<input type="checkbox"/>	1.000		140688.81		P	1.0
3	<input type="checkbox"/>	1.000		139043.70		P	1.7
4	<input type="checkbox"/>	1.000		140074.61		P	0.7
5	<input type="checkbox"/>	1.000		137136.87		P	0.7
6	<input type="checkbox"/>	1.000		135373.98		P	0.5
7	<input type="checkbox"/>	1.000		139328.20		P	1.9
8	<input type="checkbox"/>	1.000		136411.74		P	1.0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		10723125.60		A	1.7
2	<input type="checkbox"/>	1.000		10782137.34		A	0.4
3	<input type="checkbox"/>	1.000		10363648.60		A	4.1
4	<input type="checkbox"/>	1.000		11053562.47		A	2.3
5	<input type="checkbox"/>	1.000		10997242.61		A	0.9
6	<input type="checkbox"/>	1.000		10548578.80		A	3.6
7	<input type="checkbox"/>	1.000		10788661.30		A	0.8
8	<input type="checkbox"/>	1.000		10542992.55		A	1.0




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		1250757.56		P	0.6
2	<input type="checkbox"/>	1.000		1247747.81		P	1.5
3	<input type="checkbox"/>	1.000		1244824.05		P	0.9
4	<input type="checkbox"/>	1.000		1256600.92		P	1.1
5	<input type="checkbox"/>	1.000		1231925.07		P	1.7
6	<input type="checkbox"/>	1.000		1220214.17		P	0.4
7	<input type="checkbox"/>	1.000		1249946.15		M	1.6
8	<input type="checkbox"/>	1.000		1213756.56		P	1.1




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		5538647.56		A	1.4
2	<input type="checkbox"/>	1.000		5532565.75		A	0.6
3	<input type="checkbox"/>	1.000		5306238.62		A	4.7
4	<input type="checkbox"/>	1.000		5622617.27		A	2.0
5	<input type="checkbox"/>	1.000		5515889.27		A	1.2
6	<input type="checkbox"/>	1.000		5322829.46		A	3.2
7	<input type="checkbox"/>	1.000		5344902.98		A	0.8
8	<input type="checkbox"/>	1.000		4967061.62		A	2.6




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		2627351.92		A	1.4
2	<input type="checkbox"/>	1.000		2609038.14		A	1.6
3	<input type="checkbox"/>	1.000		2608845.57		A	2.1
4	<input type="checkbox"/>	1.000		2630134.81		A	1.7
5	<input type="checkbox"/>	1.000		2606412.65		A	1.2
6	<input type="checkbox"/>	1.000		2554260.17		A	0.6
7	<input type="checkbox"/>	1.000		2523801.96		A	1.7
8	<input type="checkbox"/>	1.000		2333962.47		A	0.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		5759935.93		A	0.9
2	<input type="checkbox"/>	1.000		5751191.57		A	0.7
3	<input type="checkbox"/>	1.000		5483249.67		A	5.2
4	<input type="checkbox"/>	1.000		5952275.16		A	3.4
5	<input type="checkbox"/>	1.000		5762126.32		A	1.6
6	<input type="checkbox"/>	1.000		5660400.11		A	1.6
7	<input type="checkbox"/>	1.000		5600057.99		A	1.3
8	<input type="checkbox"/>	1.000		5422590.24		A	0.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		951608.52		P	0.8
2	<input type="checkbox"/>	1.000		948482.90		P	0.2
3	<input type="checkbox"/>	1.000		945999.32		P	1.3
4	<input type="checkbox"/>	1.000		955326.59		P	1.1
5	<input type="checkbox"/>	1.000		941076.99		P	0.3
6	<input type="checkbox"/>	1.000		916415.81		P	0.3
7	<input type="checkbox"/>	1.000		929454.46		P	2.7
8	<input type="checkbox"/>	1.000		895302.80		P	1.5




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		7209311.63		A	2.4
2	<input type="checkbox"/>	1.000		7101054.69		A	2.3
3	<input type="checkbox"/>	1.000		6999373.86		A	4.6
4	<input type="checkbox"/>	1.000		7590933.50		A	1.8
5	<input type="checkbox"/>	1.000		7452087.11		A	1.3
6	<input type="checkbox"/>	1.000		7419122.39		A	2.8
7	<input type="checkbox"/>	1.000		7448285.44		A	2.2
8	<input type="checkbox"/>	1.000		7196884.90		A	1.1




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		3474477.69		A	1.5
2	<input type="checkbox"/>	1.000		3470803.70		A	0.6
3	<input type="checkbox"/>	1.000		3475785.57		A	1.4
4	<input type="checkbox"/>	1.000		3510500.36		A	0.2
5	<input type="checkbox"/>	1.000		3500597.45		A	1.7
6	<input type="checkbox"/>	1.000		3521644.18		A	1.8
7	<input type="checkbox"/>	1.000		3623951.68		A	4.1
8	<input type="checkbox"/>	1.000		3369408.56		A	0.9




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		6938014.39		A	0.6
2	<input type="checkbox"/>	1.000		7041470.47		A	1.7
3	<input type="checkbox"/>	1.000		6786818.29		A	5.9
4	<input type="checkbox"/>	1.000		7316700.77		A	1.2
5	<input type="checkbox"/>	1.000		7254159.19		A	1.3
6	<input type="checkbox"/>	1.000		7198545.13		A	2.3
7	<input type="checkbox"/>	1.000		7291992.79		A	0.4
8	<input type="checkbox"/>	1.000		7106592.78		A	1.3



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		3545699.15		A	0.8
2	<input type="checkbox"/>	1.000		3551924.11		A	1.8
3	<input type="checkbox"/>	1.000		3517339.08		A	1.5
4	<input type="checkbox"/>	1.000		3550538.73		A	1.5
5	<input type="checkbox"/>	1.000		3599260.64		A	1.6
6	<input type="checkbox"/>	1.000		3524980.19		A	1.4
7	<input type="checkbox"/>	1.000		3671558.59		A	1.7
8	<input type="checkbox"/>	1.000		3503300.68		A	2.5



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		4675552.92		A	2.0
2	<input type="checkbox"/>	1.000		4580188.27		A	1.0
3	<input type="checkbox"/>	1.000		4402573.37		A	6.1
4	<input type="checkbox"/>	1.000		4826404.75		A	2.7
5	<input type="checkbox"/>	1.000		4728746.88		A	2.9
6	<input type="checkbox"/>	1.000		4629824.96		A	3.3
7	<input type="checkbox"/>	1.000		4592377.64		A	1.4
8	<input type="checkbox"/>	1.000		4189719.49		A	2.5



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	De t	RSD
1	<input type="checkbox"/>	1.000		2649158.36		A	0.9
2	<input type="checkbox"/>	1.000		2664644.53		A	0.3
3	<input type="checkbox"/>	1.000		2630944.60		A	1.0
4	<input type="checkbox"/>	1.000		2707136.70		A	0.4
5	<input type="checkbox"/>	1.000		2641339.07		A	0.9
6	<input type="checkbox"/>	1.000		2628822.34		A	0.5
7	<input type="checkbox"/>	1.000		2651030.50		A	2.5
8	<input type="checkbox"/>	1.000		2369655.05		A	1.3