



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125635

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/25/2023	09:25
2	CAL2	1	Water	NA	NA	20.6	6.99	05/25/2023	09:26
3	CAL3	1	Water	NA	NA	20.6	10.02	05/25/2023	09:30
4	ICV	1	Water	NA	NA	20.7	7.00	05/25/2023	09:33
5	CCV1	1	Water	NA	NA	20.6	2.02	05/25/2023	09:35
6	O2901-01	1	Water	NA	NA	20.5	6.34	05/25/2023	09:40
7	O2901-01DUP	1	Water	NA	NA	20.6	6.35	05/25/2023	09:41
8	CCV2	1	Water	NA	NA	20.6	12.02	05/25/2023	09:45

U39125635

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 2901

WorkList ID : 170324

Department : Wet-Chemistry

Date : 05-25-2023 09:14:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2901-01	50038 C	Water	pH	Cool 4 deg C	PSEG03	O11	05/23/2023	9040C

Date/Time 05/25/23 09:12:0

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05/25/23 11:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]