

Clear table

Instrument ID: IC-2											
Method: 300.0 / 9056A											
Analyst: IZ											
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-05-24_	5/24/23 11:45		IZ/MA
STD2	0.4650	0.6450	0.6890	2.2470	0.5640	1.1610	3.4730	2023-05-24_	5/24/23 12:14		IZ/MA
STD3	0.8420	1.1870	1.1910	3.9750	0.9980	1.9540	5.9810	2023-05-24_	5/24/23 12:44		IZ/MA
STD4	1.0050	1.5490	1.5190	5.0350	1.2650	2.5300	7.5390	2023-05-24_	5/24/23 13:14		IZ/MA
STD5	1.8220	2.9260	2.9020	9.7750	2.4290	4.8170	14.5120	2023-05-24_	5/24/23 13:43		IZ/MA
STD6	3.9990	5.8960	5.8810	19.5890	4.8900	9.9100	29.3750	2023-05-24_	5/24/23 14:13		IZ/MA
STD7	5.0490	7.5980	7.6190	25.3790	6.3530	12.6260	38.1190	2023-05-24_	5/24/23 14:43		IZ/MA
ICV	1.8290	2.9850	3.0620	9.9450	2.5060	4.6940	14.7450	2023-05-24_	5/24/23 15:12		IZ/MA
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-05-24_	5/24/23 15:45		IZ/MA
CCV	1.9050	3.0260	2.9820	9.9440	2.4970	4.6510	14.7930	2023-05-24_	5/24/23 16:15		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-05-24_	5/24/23 16:46		IZ/MA
LB125637BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-05-24_	5/24/23 17:16		IZ/MA
LB125637BSW	1.9590	3.0080	2.9510	9.8510	2.4940	4.9980	14.8240	2023-05-24_	5/24/23 17:45		IZ/MA
O2908-13	0.8880	182.6410	0.0000	0.0000	0.0000	0.7400	9.3600	2023-05-24_	5/24/23 18:15		IZ/MA
O2908-13MS	3.6940	178.9770	2.7470	10.0170	2.4880	7.4460	23.6920	2023-05-24_	5/24/23 18:45		IZ/MA
O2908-13MSD	3.6150	177.8060	2.9280	10.4380	2.6310	8.8140	24.3500	2023-05-24_	5/24/23 19:14		IZ/MA
CCV	1.9880	3.0410	3.1500	10.0360	2.5320	6.5360	14.9380	2023-05-24_	5/24/23 19:44		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-05-24_	5/24/23 20:14		IZ/MA

Report date: 5/25/2023 10:13:17 AM
Printed by: wet


Ident: CCV
Analysis from: 5/24/2023 4:15:32 PM
File: _2023-05-24_

Last save: 5/24/2023 4:42:32 PM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34404

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.11	0.943	6.01	2.24	391.356	6.07
2	8.97	0.189	33.26	12.37	450.397	6.98
3	10.82	0.236	54.36	20.22	952.494	14.76
4	14.02	0.257	37.88	14.09	665.025	10.31
5	16.03	0.313	43.79	16.29	954.107	14.79
6	21.51	0.516	31.63	11.77	1320.114	20.46
7	25.04	0.414	61.92	23.03	1717.738	26.63
7	27.00	0.410	268.85	100.00	6451.231	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:13:25 AM
Printed by: wet


Ident: CCB
Analysis from: 5/24/2023 4:46:34 PM
File: _2023-05-24_

Last save: 5/24/2023 5:13:34 PM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34405

Last save: 5/24/2023 3:10:

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:13:32 AM
Printed by: wet


Ident: LB125637BLW
Analysis from: 5/24/2023 5:16:12 PM
File: _2023-05-24_

Last save: 5/25/2023 10:12:37 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34406

Last save: 5/24/2023 3:10:

SAMPLE:
:
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:13:39 AM
Printed by: wet


Ident: LB125637BSW
Analysis from: 5/24/2023 5:45:50 PM
File: _2023-05-24_

Last save: 5/25/2023 10:12:37 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34407

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.08	0.908	6.41	2.32	404.141	6.16
2	8.85	0.187	33.53	12.14	447.467	6.82
3	10.65	0.232	54.65	19.78	941.904	14.36
4	13.75	0.247	39.45	14.28	658.393	10.04
5	15.67	0.300	45.86	16.60	952.863	14.53
6	21.44	0.525	33.79	12.23	1432.593	21.84
7	24.95	0.411	62.56	22.65	1721.577	26.25
7	27.00	0.401	276.25	100.00	6558.938	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:13:49 AM
Printed by: wet


Ident: O2908-13
Analysis from: 5/24/2023 6:15:28 PM
File: _2023-05-24_

Last save: 5/24/2023 6:42:28 PM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34408

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.18	1.160	2.01	0.13	151.267	0.51
2	8.97	0.292	1457.35	97.37	28557.979	95.80
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	21.71	0.870	0.84	0.06	55.208	0.19
7	25.13	0.427	36.45	2.44	1044.393	3.50
7	27.00	0.393	1496.65	100.00	29808.848	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:14:15 AM
Printed by: wet


Ident: O2908-13MS
Analysis from: 5/24/2023 6:45:06 PM
File: _2023-05-24_

Last save: 5/25/2023 10:14:03 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34409

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.22	1.012	12.02	0.69	813.657	2.24
2	9.00	0.287	1445.96	83.38	27984.531	77.02
3	10.69	0.279	34.26	1.98	871.928	2.40
4	13.88	0.229	43.15	2.49	670.221	1.84
5	15.75	0.289	47.71	2.75	950.229	2.62
6	21.63	0.580	47.50	2.74	2224.393	6.12
7	25.12	0.408	103.45	5.97	2820.661	7.76
7	27.00	0.440	1734.06	100.00	36335.619	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:14:34 AM
Printed by: wet


Ident: O2908-13MSD
Analysis from: 5/24/2023 7:14:44 PM
File: _2023-05-24_

Last save: 5/25/2023 10:14:31 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34410

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.15	1.016	11.73	0.67	795.012	2.16
2	8.92	0.284	1447.95	82.47	27801.349	75.53
3	10.59	0.275	36.74	2.09	934.054	2.54
4	13.75	0.227	45.46	2.59	700.152	1.90
5	15.60	0.287	50.37	2.87	1008.461	2.74
6	21.49	0.576	57.07	3.25	2666.855	7.25
7	24.97	0.407	106.38	6.06	2902.219	7.88
7	27.00	0.439	1755.70	100.00	36808.102	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:14:46 AM
Printed by: wet


Ident: O2908-13MSD
Analysis from: 5/24/2023 7:14:44 PM
File: _2023-05-24_

Last save: 5/25/2023 10:14:31 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34410

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.15	1.016	11.73	0.67	795.012	2.16
2	8.92	0.284	1447.95	82.47	27801.349	75.53
3	10.59	0.275	36.74	2.09	934.054	2.54
4	13.75	0.227	45.46	2.59	700.152	1.90
5	15.60	0.287	50.37	2.87	1008.461	2.74
6	21.49	0.576	57.07	3.25	2666.855	7.25
7	24.97	0.407	106.38	6.06	2902.219	7.88
7	27.00	0.439	1755.70	100.00	36808.102	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:14:53 AM
Printed by: wet


Ident: CCV
Analysis from: 5/24/2023 7:44:25 PM
File: _2023-05-24_

Last save: 5/25/2023 9:20:33 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34411

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.05	0.894	6.63	2.32	410.977	5.72
2	8.82	0.189	33.07	11.56	452.647	6.30
3	10.59	0.237	56.48	19.74	1010.318	14.07
4	13.62	0.251	39.01	13.63	671.562	9.35
5	15.52	0.302	45.63	15.95	968.093	13.48
6	21.45	0.551	43.10	15.07	1929.941	26.88
7	24.96	0.416	62.19	21.74	1735.796	24.18
7	27.00	0.406	286.11	100.00	7179.332	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2023 10:15:09 AM
Printed by: wet


Ident: CCB
Analysis from: 5/24/2023 8:14:02 PM
File: _2023-05-24_

Last save: 5/25/2023 10:15:06 AM

Method: AnionsIC2-052423.mtw
Run operator: wet
Analysis number: 34412

Last save: 5/24/2023 3:10:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD