

LB125638


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57722



A= 0.0000e+000

B= 1.7145e-004

C= 8.9334e-003

Rho= 0.9995290

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4.0
0.0	0.000	0.033	0.033	138	0.000	138					19
0.2	0.200	0.238	0.038	1337	0.0 %	1337					8
2.5	2.500	2.300	-0.200	13363	0.0 %	13363					3
5.0	5.000	5.127	0.127	29851	0.0 %	29851					1
7.5	7.500	7.597	0.097	44257	0.0 %	44257					1
10.0	10.000	9.905	-0.095	57722	0.0 %	57722					

LB125638 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 May 2023 10:30:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	138	-	0.0000	7471B	PPB	25 May 2023 10:49:47	S	Std
0.2 - 1	50.2	1337	-	0.2000	7471B	PPB	25 May 2023 10:52:05	S	Std
2.5 - 1	52.5	13363	-	2.5000	7471B	PPB	25 May 2023 10:54:23	S	Std
5.0 - 1	55	29851	-	5.0000	7471B	PPB	25 May 2023 10:56:40	S	Std
7.5 - 1	57.5	44257	-	7.5000	7471B	PPB	25 May 2023 11:01:37	S	Std
10.0 - 1	510	57722	-	10.0000	7471B	PPB	25 May 2023 11:03:54	S	Std
ICV90 - 1	ICV90	23547	4.0461	-	7471B	PPB	25 May 2023 11:06:47	U	SMPL
ICB90 - 1	ICB90	15	0.0115	-	7471B	PPB	25 May 2023 11:09:02	U	SMPL
CCV18 - 1	CCV18	30080	5.1662	-	7471B	PPB	25 May 2023 11:11:20	U	SMPL
CCB18 - 1	CCB18	-35	0.0029	-	7471B	PPB	25 May 2023 11:13:35	U	SMPL
CRA - 1	CRA	1263	0.2255	-	7471B	PPB	25 May 2023 11:15:53	U	SMPL
HighStd - 1	HighStd	57824	9.9229	-	7471B	PPB	25 May 2023 11:18:08	U	SMPL
ChkStd - 1	ChkStd	40024	6.8711	-	7471B	PPB	25 May 2023 11:20:24	U	SMPL
PB153032BL - 1	PBS	-252	-0.0343	-	7471B	PPB	25 May 2023 11:22:42	U	SMPL
PB153032BS - 1	LCSS	22551	3.8753	-	7471B	PPB	25 May 2023 11:25:00	U	SMPL
O2908-01 - 1	SB10A	6944	1.1995	-	7471B	PPB	25 May 2023 11:27:17	U	SMPL
O2908-02 - 1	SB10B	1233	0.2203	-	7471B	PPB	25 May 2023 11:29:36	U	SMPL
O2908-04 - 1	SB09	2189	0.3842	-	7471B	PPB	25 May 2023 11:31:53	U	SMPL
O2908-06 - 1	DUP01	1683	0.2975	-	7471B	PPB	25 May 2023 11:34:10	U	SMPL
O2908-07 - 1	SB13	671	0.1240	-	7471B	PPB	25 May 2023 11:36:26	U	SMPL
CCV19 - 1	CCV19	29747	5.1091	-	7471B	PPB	25 May 2023 11:38:41	U	SMPL
CCB19 - 1	CCB19	-15	0.0064	-	7471B	PPB	25 May 2023 11:40:57	U	SMPL
O2913-01 - 1	TP-70A	1254	0.2239	-	7471B	PPB	25 May 2023 11:43:14	U	SMPL
O2920-01 - 1	280608	997	0.1799	-	7471B	PPB	25 May 2023 11:45:30	U	SMPL
O2920-03 - 1	RBR-400027	19677	3.3826	-	7471B	PPB	25 May 2023 11:47:46	U	SMPL
O2925-01 - 1	CONCRETE-052423	1751	0.3091	-	7471B	PPB	25 May 2023 11:50:02	U	SMPL
O2935-01 - 1	ETGI-346	9418	1.6237	-	7471B	PPB	25 May 2023 11:52:19	U	SMPL
O2938-01 - 1	SU-04-052423	2014	0.3542	-	7471B	PPB	25 May 2023 11:54:36	U	SMPL
O2940-01 - 1	RBR-251671	9856	1.6988	-	7471B	PPB	25 May 2023 11:56:52	U	SMPL
O2940-01DUP - 1	RBR-251671DUP	8069	1.3924	-	7471B	PPB	25 May 2023 11:59:09	U	SMPL
O2940-01MS - 1	RBR-251671MS	36788	6.3163	-	7471B	PPB	25 May 2023 12:01:26	U	SMPL
O2940-01MSD - 1	RBR-251671MSD	32790	5.6308	-	7471B	PPB	25 May 2023 12:09:45	U	SMPL
CCV20 - 1	CCV20	30011	5.1544	-	7471B	PPB	25 May 2023 12:12:00	U	SMPL
CCB20 - 1	CCB20	-71	-0.0032	-	7471B	PPB	25 May 2023 12:14:18	U	SMPL
O2941-01 - 1	EME-TOPSOIL-001	2129	0.3740	-	7471B	PPB	25 May 2023 12:16:37	U	SMPL
O2941-02 - 1	TILCON-CLEANFILL-001	377	0.0736	-	7471B	PPB	25 May 2023 12:18:55	U	SMPL
O2941-03 - 1	GHOST-CLEANFILL-001	4601	0.7978	-	7471B	PPB	25 May 2023 12:21:12	U	SMPL
O2940-01LX5 - 1		1856	0.3271	-	7471B	PPB	25 May 2023 12:23:28	U	SMPL
O2940-01A - 1		36369	6.2444	-	7471B	PPB	25 May 2023 12:28:31	U	SMPL
CCV21 - 1	CCV21	29882	5.1322	-	7471B	PPB	25 May 2023 12:30:47	U	SMPL
CCB21 - 1	CCB21	-61	-0.0015	-	7471B	PPB	25 May 2023 12:33:06	U	SMPL