



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB125640

BalanceID: WC SC-4

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/24/2023	17:20
2	CAL2	1	Water	NA	NA	20.6	7.00	05/24/2023	17:21
3	CAL3	1	Water	NA	NA	20.6	10.03	05/24/2023	17:22
4	ICV	1	Water	NA	NA	20.7	6.98	05/24/2023	17:25
5	CCV1	1	Water	NA	NA	20.7	2.02	05/24/2023	17:30
6	O2897-01	1	Solid	20.02	20	22.7	6.24	05/24/2023	17:33
7	O2899-03	1	Solid	20.01	20	21.9	8.51	05/24/2023	17:36
8	O2899-05	1	Solid	20.02	20	21.7	8.64	05/24/2023	17:40
9	O2899-07	1	Solid	20.02	20	21.2	8.19	05/24/2023	17:42
10	O2899-09	1	Solid	20.03	20	21.4	7.90	05/24/2023	17:45
11	O2908-03	1	Solid	20.03	20	22.2	6.74	05/24/2023	17:50
12	O2908-05	1	Solid	20.02	20	22.6	9.47	05/24/2023	18:00
13	O2908-08	1	Solid	20.04	20	22.5	8.81	05/24/2023	18:05
14	O2913-02	1	Solid	20.02	20	22.0	7.45	05/24/2023	18:10
15	O2916-02	1	Solid	20.01	20	21.7	5.91	05/24/2023	18:11
16	CCV2	1	Water	NA	NA	20.6	12.02	05/24/2023	18:15
17	O2916-02DUP	1	Solid	20.02	20	21.8	5.92	05/24/2023	18:16
18	CCV3	1	Water	NA	NA	20.6	2.02	05/24/2023	18:20

VF 125640

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRROSIVITY O2913 WorkList ID : 170286 Department : Wet-Chemistry Date : 05-24-2023 11:12:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2897-01	OILY-SOLIDS-0523	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/23/2023	9045D
O2899-03	WC06	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/22/2023	9045D
O2899-05	WC01	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/22/2023	9045D
O2899-07	WC08	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/22/2023	9045D
O2899-09	WC14	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/22/2023	9045D
O2908-03	WC10	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/23/2023	9045D
O2908-05	WC09	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/23/2023	9045D
O2908-08	WC13	Solid	Corrosivity	Cool 4 deg C	loui01	O11	05/23/2023	9045D
O2913-02	TP-70A	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/23/2023	9045D
O2916-02	SU-621-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	05/23/2023	9045D

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

05/24/23
17:15
J.C (COC)
J.C (SUN)

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

05/24/23
18:40
J.C (SUN)
J.C (SUN)