



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/26/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 05/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 05/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125643

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2892-13	TP-29	12	1.15	8.81	9.96	8.53	83.8	
O2916-01	SU-621-COMP-01	1	1.18	8.58	9.76	8.44	84.6	
O2916-03	SU-621-VOC-01	2	1.18	8.58	9.76	8.44	84.6	
O2916-05	SU-621-01	3	1.13	8.71	9.84	8.62	86.0	
O2916-06	SU-621-02	4	1.19	8.52	9.71	8.55	86.4	
O2919-01	IMPACTED-STONE	5	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O2919-02	IMPACTED-STONE-EPH-2	6	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O2922-01	ORA-1618	7	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O2922-02	ORA-1618-EPH-2	8	1.44	8.45	9.89	8.98	89.2	
O2923-01	OR-03-052423	9	1.11	8.58	9.69	8.96	91.5	
O2923-02	OR-03-052423-EPH-2	10	1.11	8.58	9.69	8.96	91.5	
O2937-01	AU-06-052423	11	1.19	8.52	9.71	9.12	93.1	
O2944-16	BP-VPB-RW11-GW-398-400	13	1.16	8.83	9.99	2.55	15.7	SLUDGE SAMPLE
O2944-17	BP-VPB-RW11-GW-418-420	14	1.18	8.78	9.96	2.2	11.6	SLUDGE SAMPLE
O2945-01	SB20	15	1.18	8.45	9.63	8.00	80.7	
O2945-02	WC20	16	1.19	8.53	9.72	8.53	86.0	
O2945-03	SB17A	17	1.19	8.55	9.74	8.6	86.7	
O2945-04	SB17B	18	1.16	8.81	9.97	8.77	86.4	
O2945-05	WC17	19	1.13	8.55	9.68	8.4	85.0	
O2945-06	SB23	20	1.15	8.64	9.79	8.17	81.3	
O2945-07	WC23	21	1.16	8.83	9.99	8.45	82.6	
O2948-01	01-A-01-B-01-C	22	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$