



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125649

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/25/2023	12:30
2	O2887-03		1	14	100.00	>212.0	>212.0	05/25/2023	13:00
3	O2887-03DUP		1	12	100.00	>212.0	>212.0	05/25/2023	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66125649

WorkList Name : fp s5-25 WorkList ID : 170308 Department : Wet-Chemistry Date : 05-25-2023 08:23:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2887-03	SCRUBBER-DRUM	Solid	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B

Date/Time 05/25/2023 12:20
 Raw Sample Received by: RFA CAD
 Raw Sample Relinquished by: JALBCL

Date/Time 05/25/2023 14:10
 Raw Sample Received by: JALBCL
 Raw Sample Relinquished by: RFA CAD