



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125650

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	05/25/2023	08:30
2	O2891-03		1	14	100.00	>212.0	>212.0	05/25/2023	09:00
3	O2891-03DUP		1	12	100.00	>212.0	>212.0	05/25/2023	09:30
4	O2897-02		1	14	100.00	>212.0	>212.0	05/25/2023	10:00
5	O2899-10		1	12	100.00	>212.0	>212.0	05/25/2023	10:30
6	O2901-01		1	12	100.00	>212.0	>212.0	05/25/2023	11:00
7	O2908-13		1	13	100.00	>212.0	>212.0	05/25/2023	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB125650

WorkList Name : fp-5-25 WorkList ID : 170304 Department : Wet-Chemistry Date : 05-25-2023 08:19:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2891-Q3	A653	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/22/2023	1010B
O2897-Q2	OILY-WATER-0523	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/23/2023	1010B
O2899-10	TWP01	Water	Flash Point	Cool 4 deg C	loui01	O11	05/22/2023	1010B
O2901-Q1	50038	Water	Flash Point	Cool 4 deg C	PSEG03	O11	05/23/2023	1010B
O2908-13	TWP02	Water	Flash Point	Cool 4 deg C	loui01	O11	05/23/2023	1010B

Date/Time 05/25/2023 08:10
 Raw Sample Received by: RM (we)
 Raw Sample Relinquished by: 48 (we)

Date/Time 05/25/2023 12:10
 Raw Sample Received by: 48 (we)
 Raw Sample Relinquished by: RM (we)