



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125653

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/25/2023	13:22
2	CAL2	1	Water	NA	NA	20.6	7.00	05/25/2023	13:24
3	CAL3	1	Water	NA	NA	20.8	10.02	05/25/2023	13:25
4	ICV	1	Water	NA	NA	20.8	6.99	05/25/2023	13:30
5	CCV1	1	Water	NA	NA	20.6	2.01	05/25/2023	13:32
6	O2949-01	1	Water	NA	NA	20.2	5.47	05/25/2023	13:35
7	O2949-01DUP	1	Water	NA	NA	20.3	5.48	05/25/2023	13:36
8	O2949-02	1	Water	NA	NA	20.2	5.32	05/25/2023	13:40
9	O2949-03	1	Water	NA	NA	20.3	5.10	05/25/2023	13:41
10	CCV2	1	Water	NA	NA	20.7	12.02	05/25/2023	13:45

WB 125653

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2949 WorkList ID : 170340 Department : Wet-Chemistry Date : 05-25-2023 13:13:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2949-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER13	D12	05/24/2023	9040C
O2949-02	CELL-B-POOL-WATER	Water	pH	Cool 4 deg C	STER13	D12	05/24/2023	9040C
O2949-03	CELL-C-POOL-WATER	Water	pH	Cool 4 deg C	STER13	D12	05/24/2023	9040C

Date/Time 05/25/23 13:20
 Raw Sample Received by: JH (WPC)
 Raw Sample Relinquished by: JT (SM)

Date/Time 05/25/23 15:00
 Raw Sample Received by: JH (WPC)
 Raw Sample Relinquished by: JH (WPC)