


LB125654

245.1 INSTRUMENT ID: CV1

Linear

 μ Abs.:

43825



A= 0.0000e+000

B= 2.2961e-004

C= -1.9836e-002

Rho= 0.9997279

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	-0.020	-0.020	1	0.000	1					1
0.05	0.050	0.056	0.006	331	0.0 %	331	0				12
0.2	0.200	0.215	0.015	1023	0.0 %	1023					8
2.5	2.500	2.452	-0.048	10765	0.0 %	10765					12
5.0	5.000	5.156	0.156	22544	0.0 %	22544					3
7.5	7.500	7.348	-0.152	32087	0.0 %	32087					2
10.0	10.000	10.043	0.043	43825	0.0 %	43825					0

LB125654 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 May 2023 15:39:58

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50/ 50	1	-	0.0000245.1	PPB	25 May 2023 16:12:35	S	Std
0.05 - 1	50/2 50.05	331	-	0.0500245.1	PPB	25 May 2023 16:14:53	S	Std
0.2 - 1	50.5 51.2	1023	-	0.2000245.1	PPB	25 May 2023 16:17:10	S	Std
2.5 - 1	52.5 52.5	10765	-	2.5000245.1	PPB	25 May 2023 16:19:27	S	Std
5.0 - 1	55 55	22544	-	5.0000245.1	PPB	25 May 2023 16:24:18	S	Std
7.5 - 1	57.5 57.5	32087	-	7.5000245.1	PPB	25 May 2023 16:29:14	S	Std
10.0 - 1	510 510	43825	-	10.0000245.1	PPB	25 May 2023 16:31:31	S	Std
ICV91 - 1	ICV91	17698	4.0437	-245.1	PPB	25 May 2023 16:36:55	U	SMPL
ICB91 - 1	ICB91	-70	-0.0359	-245.1	PPB	25 May 2023 16:39:10	U	SMPL
CCV22 - 1	CCV22	22220	5.0820	-245.1	PPB	25 May 2023 16:41:26	U	SMPL
CCB22 - 1	CCB22	-116	-0.0465	-245.1	PPB	25 May 2023 16:43:42	U	SMPL
CRA - 1	CRA	344	0.0591	-245.1	PPB	25 May 2023 16:45:58	U	SMPL
HighStd - 1	HighStd	45086	10.3322	-245.1	PPB	25 May 2023 16:48:15	U	SMPL
ChkStd - 1	ChkStd	30354	6.9496	-245.1	PPB	25 May 2023 16:50:31	U	SMPL
PB153048BL - 1	PBW	-138	-0.0515	-245.1	PPB	25 May 2023 16:52:50	U	SMPL
PB153048BS - 1	LCSW	17924	4.0956	-245.1	PPB	25 May 2023 16:55:08	U	SMPL
O2830-01 - 1	SEMI-ANNUAL-CONDENSATE	39488	9.0468	-245.1	PPB	25 May 2023 16:57:25	U	SMPL
O2830-01DUP - 1	SEMI-ANNUAL-CONDENSATEDUP	42041	9.6330	-245.1	PPB	25 May 2023 16:59:42	U	SMPL
O2830-01MS - 1	SEMI-ANNUAL-CONDENSATEMS	55029	12.6151	-245.1	PPB	25 May 2023 17:02:02	U	SMPL
O2830-01MSD - 1	SEMI-ANNUAL-CONDENSATEMSD	57090	13.0883	-245.1	PPB	25 May 2023 17:04:22	U	SMPL
O2899-10 - 1	TWP01	23187	5.3040	-245.1	PPB	25 May 2023 17:06:40	U	SMPL
CCV23 - 1	CCV23	21621	4.9445	-245.1	PPB	25 May 2023 17:08:58	U	SMPL
CCB23 - 1	CCB23	-141	-0.0522	-245.1	PPB	25 May 2023 17:11:16	U	SMPL
O2908-13 - 1	TWP02	1170	0.2488	-245.1	PPB	25 May 2023 17:13:33	U	SMPL
O2945-09 - 1	TWP03	2249	0.4965	-245.1	PPB	25 May 2023 17:15:49	U	SMPL
O2945-11 - 1	TWP04	429	0.0787	-245.1	PPB	25 May 2023 17:18:05	U	SMPL
O2830-01LX5 - 1		8590	1.9525	-245.1	PPB	25 May 2023 17:20:21	U	SMPL
O2830-01A - 1		56629	12.9825	-245.1	PPB	25 May 2023 17:22:37	U	SMPL
CCV24 - 1	CCV24	21549	4.9279	-245.1	PPB	25 May 2023 17:24:54	U	SMPL
CCB24 - 1	CCB24	-131	-0.0499	-245.1	PPB	25 May 2023 17:27:14	U	SMPL
O2830-01DLX2 - 1		21877	5.0032	-245.1	PPB	25 May 2023 17:29:33	U	SMPL
O2830-01DUPDLX2 - 1		20146	4.6058	-245.1	PPB	25 May 2023 17:31:50	U	SMPL
O2830-01MSDLX2 - 1		28069	6.4250	-245.1	PPB	25 May 2023 17:34:08	U	SMPL
O2830-01MSDDLX2 - 1		30434	6.9680	-245.1	PPB	25 May 2023 17:36:25	U	SMPL
O2830-01LDLX10 - 1		3584	0.8031	-245.1	PPB	25 May 2023 17:38:42	U	SMPL
O2830-01ADLX2 - 1		28003	6.4098	-245.1	PPB	25 May 2023 17:41:00	U	SMPL
CCV25 - 1	CCV25	22402	5.1236	-245.1	PPB	25 May 2023 17:43:17	U	SMPL
CCB25 - 1	CCB25	-148	-0.0538	-245.1	PPB	25 May 2023 17:45:33	U	SMPL