

Sample Name: S0 Acquired: 5/25/2023 12:22:27 Type: Cal
Method: SFAM01.1 INSTRUMENT P4 WITH 6 POINT ICAL(v292)(v3) Mode: IR Corr. Factor: 1.000000
User: Bin Custom ID1: S0 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	-.00015	.00002	-.00029	.00021	.00052	-.00013	.21900	.00365
Stddev	.00002	.00020	.00011	.00011	.00002	.00034	.00263	.00016
%RSD	12.562	856.89	39.276	53.499	3.8708	256.50	1.1999	4.3081

#1	-.00017	-.00021	-.00037	.00033	.00053	-.00002	.21770	.00349
#2	-.00013	.00017	-.00016	.00016	.00050	-.00051	.21728	.00380
#3	-.00015	.00011	-.00035	.00012	.00053	.00014	.22203	.00366

Elem	Cd2144	Ca3736	Cr2677	Co2286	Cu3247	Fe2598	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.0002	.00056	.00001	.00058	.0003	.0002	-.00035	-.00033
Stddev	.0001	.00162	.00007	.00016	.0003	.0003	.00085	.00053
%RSD	77.73	290.87	515.38	28.076	118.2	140.5	240.82	160.56

#1	.0003	-.00126	.00002	.00049	-.0001	-.0001	-.00092	-.00028
#2	.0001	.00185	.00008	.00049	.0005	.0005	-.00076	-.00088
#3	.0001	.00107	-.00006	.00077	.0005	.0003	.00062	.00017

Elem	Ni2316	Ag3280	Na8183	V_2924	Zn2138	K_7698	P_1774	B_2496
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	-.00048	-.00045	-.0052	-.00004	.00208	.0277	-.00122	.00910
Stddev	.00020	.00003	.0015	.00011	.00033	.0039	.00001	.00050
%RSD	40.643	7.6995	27.97	252.42	15.660	14.11	.73999	5.4912

#1	-.00071	-.00048	-.0069	.00005	.00172	.0266	-.00122	.00855
#2	-.00040	-.00041	-.0041	-.00002	.00234	.0244	-.00123	.00924
#3	-.00034	-.00046	-.0047	-.00017	.00219	.0320	-.00122	.00951

Elem	Mo2020	S_1820	Si2516	Sn1899	Ti3361	Li6707	Sr4077	
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	
Avg	.00003	.00038	.0014	.00027	.00485	.01958	-.00305	
Stddev	.00016	.00007	.0007	.00009	.00024	.00190	.00048	
%RSD	495.56	17.699	48.67	32.380	4.8817	9.6883	15.694	

#1	.00021	.00045	.0007	.00029	.00469	.02033	-.00250	
#2	-.00010	.00037	.0020	.00034	.00474	.01742	-.00325	
#3	-.00002	.00032	.0016	.00017	.00512	.02098	-.00339	

Sample Name: S0 Acquired: 5/25/2023 12:22:27 Type: Cal
 Method: SFAM01.1 INSTRUMENT P4 WITH 6 POINT ICAL(v292)(v3) Mode: IR Corr. Factor: 1.000000
 User: Bin Custom ID1: S0 Custom ID2: Custom ID3:
 Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4186.5	61852.	5001.3	4013.9	3917.3
Stddev	2.0	211.	70.2	29.1	1.8
%RSD	.04808	.34191	1.4032	.72422	.04679
#1	4188.7	61688.	4961.6	3992.0	3916.4
#2	4184.7	61776.	5082.4	4002.9	3919.4
#3	4186.2	62090.	4960.0	4046.9	3916.1